INCH-POUND

A-A-54622
31 March 1992

COMMERCIAL ITEM DESCRIPTION

ARTHROGRAM KIT, DISPOSABLE, 10s

The General Services Administrations has authorized the use of this commercial item description.

This commercial item description covers a sterile disposable, pyrogen-free arthrogram set that includes all components, except drugs, needed for a contrast study.

Salient Characteristics:

Material. The components of the set shall be fabricated of suitable plastics or other materials. The needles shall be fabricated from austenitic corrosion-resistant steel in accordance with ASTM Designation F899.

Style, design and dimensions. The set shall consist of as a minimum, the following components:

An outer drape that unfolds to provide a sterile field. Shall be fabricated of a nonwoven inpervious material.

Two non-woven fabric absorbent towels, approximately 19 by 17-1/2 in. (48.3 by 44.5 cm).

One non-woven fabric absorbent fenestrated drape, approximately 18 by 27-1/4 in. (45.7 by 69.2 cm).

One hospital wrap approximately 27-1/4 by 27-1/4 in. (69.2 by 69.2 cm). May also be used as outer drape if hospital wrap meets above requirements for outer drape.

One adhesive bandage approximately 1.3 cm by 6.4 cm.

Three graduated plastic syringes with tip protectors: one 5 ml, one 10 ml, and one 30 ml capacity.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:36Z
Check the source to verify that this is the current version before use.

Three needles: one 25 gauge by 5/8 in. (1.5 cm) needle, one 18 gauge by 1-1/2 in. (3.8 cm) transfer needle, and one 20 gauge by 1-1/2 in. (3.8 cm) short bevel needle.

One 10 in. (25.4 cm) transparent, plastic connecting tube with a female connector adapter on one end and a male connector adapter on the opposite end. Inside diameter of the tubing shall be approximately 0.083 in.

Two plastic 200 ml basins (approximate capacity).

One 50 ml plastic medicine cup.

Two 3 by 3 in. (7.6 by 7.6 cm) gauze dressing pads.

Two sponge applicators with plastic handles.

The needle adapter shall be tapered to fit a standard Luer hub in accordance with ANSI/HIMA Standard MD70.1 The set components shall be packaged in a suitable tear-open package that will maintain the sterility of the product until opened or damaged. The outer wrapper shall clearly state the contents of the enclosed tray and provide instructions for use.

Pyrogen tests. The set components shall meet the USP requirements for pyrogens.

Sterility. The set components shall be sterile with a Sterility Assurance Level of 1×10^{-6} in accordance with AAMI guidelines.

Heavy metals extractability. Components coming into contact with intravenous solution or patients blood shall be tested. The quantity of lead and cadmium shall not exceed 5 parts per million (PPM) and the amount of barium shall not exceed 10 ppm when tested in accordance with the USP. A certificate of compliance, stipulating that offered items meet the heavy metals extractable testing, shall be acceptable, providing that the contractor has supplied acceptable items to the Government within the past three years and that there have been no changes to material formulation and method of processing since actual testing was performed.

Workmanship. The set shall be free from defects which detract from its appearance or impair its serviceability. Set components shall be free from cuts, tears, flash, weak areas or foreign material.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:36Z
Check the source to verify that this is the currentversion before use.

Unit. Package (PG). One package containing ten individually packaged sterile arthrogram kits as specified, constitutes one unit. The individual package shall be a sealed peel-open type container capable of maintaining sterility of contents unless the package is damaged or opened.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one arthrogram set. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one arthrogram set. Sampling shall be inspection level S-2 with an AQL of 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-153-5432	ARTHROGRAM KIT, DISPOSABLE, 10s

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5303
Location: anth/S11

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54622	31 March 1992

3. DOCUMENT TITLE
ANTHROGRAM KIT, DISPOSABLE, 10s

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: 215-737-2870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete. — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.