METRIC

A-A-54628
17 April 1992

COMMERCIAL ITEM DESCRIPTION

SLIDE, MICROSCOPE (Plain)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a microscope slide for use with differential counters requiring 1 millimeter (mm) thick slides.

Salient characteristics:

Shall be a 75 x 25 millimeter (mm) plain microscope slide suitable for use in laboratory procedures and differential counters.

Shall be furnished with the following styles for the working ends (see ordering data):

Style I. Shall have all four edges flat and perpendicular to the microscope slide faces without any beveling. The edges of the short dimension shall be free of chips, and shall be finished to ensure smoothness and uniformity for use in blood smear work. The edges of the long dimension shall be straight and shall be free of roughness or sharpness to facilitate handling. The corners of the microscope slide shall be square.

Style II. Shall have the short dimensioned ends with double sided beveled edges (i.e. two bevels per end). The beveling shall be at an approximate 45 degree angle with each tip in the form of a knife edge, flat surface, or rounded surface. The short ends may also be partially beveled and need not be further ground.

Shall have microscope slide faces free of scratches which will affect their use.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Shall have dimensions of 75 mm length X 25 mm width, with a ± 0.8 mm tolerance for each dimension. The allowable thickness range shall be 0.93 to 1.07 mm.

Shall be fabricated from clear transparent glass, free from imbedded foreign material, bubbles, seeds, blisters, striae and internal cloudiness.

Shall be supplied precleaned, free of residue, and ready for use. The faces of the microscope slide shall be such that, after being dipped into a beaker of distilled water, both surfaces immediately after removal from the water exhibit a uniform dispersion of liquid and absence of smears.

Workmanship. The microscope slide shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing seventy-two microscope slides, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Sampling for inspections and tests shall also be conducted in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Style of slide required, expiration dating period required, intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6640-01-007-9950	SLIDE, MICROSCOPE, Plain 75 by 25 mm, 72s.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6640-1744

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.