INCH-POUND

A-A-54632
30 April 1992

COMMERCIAL ITEM DESCRIPTION

HOLSTER KIT, Physician, Compartmented

The General Services Administrations has authorized the use of this commercial item description.

This commercial item description covers a compartmented physician holster, suitable for use with and care of emergency medical system equipment.

Salient characteristics:

General. The holster kit shall consist of a nylon compartmented holster containing medical devices for use with and care of medical system equipment. The holster shall consist of the following equipment as a minimum:

Quantity	Component
1	Holster, compartmented nylon case
1	Forceps, Kelly, 5-1/2 in. (13.97 cm) long
1	Knife, duty
1	Penlight, disposable
1	Punch, window
1	Scissors, bandage
1	Shears, general purpose, heavy duty
1	Tourniquet, hooks and pile type

Holster case. The holster case shall be constructed of heavy duty nylon material and shall be 10 inches long opened by 5 inches wide. Shall be three layers of nylon material to form two rear compartments for the forceps, shears and scissors and three front compartments for the penlight, punch and knife. The center layer of nylon material shall have a 4 inch flap to cover the front compartments with a velcro hook and pile closure. The back layer of nylon material shall fold back 3 inches to form an opening for belt attachment and shall be secured with hook and pile closure material. The nylon material edges shall be covered with double stitched nylon edging material.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Forceps Kelly. The forceps shall be a hemostatic Kelly type, fabricated of corrosion resistant steel. Forceps shall be 5-1/2 inches long with grasping teeth and have a three postion handle locking mechanism.

Knife, EMS duty. The knife shall be an emergency medical service (EMS) duty knife for use in cutting cloth or other material. The knife shall be a "jack knife" type, 5 inches long folded and 8-1/2 inches long opened. The blade and handle shall each be 3/4 inches wide. The blade shall lock into the open position when opened and shall be manually closed when the locking mechanism is depressed. The handle shall be brass with plastic side inserts and the blade shall be corrosion resisting steel.

Penlight. The penlight shall be disposable and shall be 4-5/8 inches long by 1/2 inch in diameter. The penlight shall be provided with a battery and shall be activated by depressing a pocket clip. The side of the penlight shall contain a pupil guage with eight pupil sizes from 2mm to 9mm in diameter.

Punch window. The punch shall be suitable for use in window breakage in emergency situations. The punch shall be 4-7/8 inches long by 1/2 inch in diameter and shall consist of a housing containing a spring loaded 1/8 inch diameter pointed hardened steel shaft. The housing shall be fabricated of steel knurled on the outside to provide a secure gripping surface. The pointed shaft shall protrude 1 inch from the tapered end of the housing. The housing shall provide features for disassembly to replace the internal spring or steel point.

Scissors, bandage. The bandage scissors shall be 5-1/2 inches long Lister type. The scissors shall be fabricated of corrosion resisting steel.

Shears. The shears shall be general purpose heavy duty shears suitable for use in cutting heavy duty multi-ply cloth or other material. The shears shall be 7-1/2 inches long with the cutting blades set at an angle of 25 degrees to the handle. The blades shall have serrations for cutting and the handle finger loops shall be plastic covered for ease of use. A lifting and guiding tab shall be provided on the upper blade. The metal parts of the shears shall be fabricated of corrosion resistant steel.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Tourniquet. The tourniquet shall be rubber material with hook and pile closure material on the ends to secure the tournquet. The tourniquet shall be 14 inches long by 1 inch wide by 1/8 inch thick. The hook material shall be 5 inches long at one end with 3-1/2 inches of pile on the opposite side at the opposite end.

The holster shall be provided with all items in their respective compartments with the closed tourniquet around the holster and through the rear closed flap.

Material. The corrosion resistant steel shall be in accordance with ASTM Designation F899. Plastic or other material shall be suitable for component medical use.

Workmanship. The holster kit shall be free from defects which detract from its appearance or impair its serviceability. Components shall be free from burrs, flash, cracks, corrosion or foreign material.

Unit. (EA). One physicians holster, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one holster kit and accessories with all components as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for test shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one holster kit. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one holster kit. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL of 2.5, and sampling for tests shall be conducted at inspection level S-1, AQL 1.0.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activitiy and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number and its Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-251-3240	Holster Kit, Physician, Compartmented Nylon 7 by 4-3/4 by 1-1/4 inch

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5344
Location: HOLSTER/S11

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54632	30 April 1992

3. DOCUMENT TITLE
HOLSTER KIT, PHYSICIAN, COMPARTMENTED

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215)7372870	444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://dodsist.dla.mil — Downloaded: 2020416-201609.362
Check the source to verify that this is the current version before use.