INCH-POUND

A-A-54642
17 May 1992

COMMERCIAL ITEM DESCRIPTION

SLIPPERS, CONVALESCENT PATIENT'S

(Polyurethane Foam, Disposable)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for MIL-S-37080A which is cancelled.

This Commercial Item Description covers polyurethane foam, disposable, convalescent patient's, slippers.

Salient Characteristics:

The slipper shall be furnished in the following sizes:

Adults	Children
Small - 4 to 6	Toddler - 4 to 7
Medium - 6 to 8	Child - 8 to 12
Large - 8 to 10	Youth - 13 to 3
Extra large - 10 to 12	

The slipper shall be fabricated from a flexible, porous, hand washable, re-usable, urethane foam. The urethane foam shall not have an objectionable odor, and shall be non-allergenic, non-irritating and non-toxic. The toxicity shall be determined by any standard method. The urethane foam shall be fire retardant when tested in accordance with Motor Vehicle Safety Standard No. 302.

AMSC N/A FSC 6532

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

The foam shall conform to the following requirements:

Thickness (inches)	0.275 plus or minus 10 percent
Pounds per cubic foot	1.65 plus or minus 0.10
Tensile strength (PSI minimum)	25
Elongation (percent minimum)	300
Compression set at 90 percent deflection (percent maximum)	30
Indentation load deflection at 25 percent deflection (pounds-force per 50 square inches)	42 plus or minus 5

The above foam requirements shall be determined in accordance with ASTM Standard D1564, Slab, Flexible Urethane Foam, Testing.


The slipper shall be an elongated oval in shape with a rounded opening in the upper at each end for the insertion of the foot, and a sole on the bottom. The rounded opening may be formed using the same manufacturing process used to heat seal the slipper edge (see figure 1, section AA). The toe and heel shall be enclosed. The upper and sole shall each be fabricated from one piece of urethane foam. Each slipper shall be interchangeable, not shaped or contoured, for the right or left foot, so that each individual slipper shall be capable of being worn on either foot. The uppers and soles shall be heat sealed together as close as possible to the edge of the slipper. The heat sealed seam shall be straight, continuous, and parallel to the edge of the slipper. The slipper shall conform in appearance to figure 1.


The overall dimensions of the slipper shall conform, in inches, to those shown below (see figure 1):

Size	"A"	"B"	"C"
Small	8-13/16	3-1/4	3-13/16
Medium	9-9/16	3-3/8	4
Large	10-9/16	3-5/8	4-7/16
Extra large	11-7/16	3-7/8	4-5/8
Toddler	5-1/2	2-1/2	2-13/16
Child	6-13/16	2-3/4	3-3/16
Youth	7-7/8	3	3-7/16

The tolerance on all dimensions shall be plus or minus 3/16 inch. The variation in dimensions between the dimensions of each slipper in a pair of slippers shall not exceed 1/4 inch.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.





SEC. /A-/A

FIGURE 1. Slippers, convalescent patient's, polyurethane foam.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

The slipper opening in the upper shall be sized such that the slipper is held securely on the foot without tearing of the slipper. There shall be no straps, snaps, or other fastening or adjusting devices.

The color of the slipper sole shall be charcoal gray. The color of the upper for the various sizes shall be as follows:

Small	Gold	Toddler	Rose
Medium	Blue	Child	Green
Large	Avocado	Youth	Lilac
Extra large	Pumpkin		

The slipper shall not be stiff nor contain any spot, stain, hole, void, abraded area, tear, cut, mend, patch, or scorch. No part of any slipper shall be deformed, distorted, bent, or pleated as a result of the heat sealing, or by any other process of manufacturing.

One pair of slippers shall be packaged in a polyethylene bag of suitable size and shape, labeled "FIRE RETARDANT".

Workmanship. The convalescent patient's slippers shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Pair (PR). One pair of slippers of the size specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:36Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one pair of convalescent patient's slippers.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be one pair of convalescent patient's slippers.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document.

6532-00-079-7889 (Small)
6532-00-079-7899 (Medium)
6532-00-079-7902 (Large)
6532-00-079-7904 (Extra large)
6532-01-011-5055 (Toddler)
6532-01-011-5056 (Child)
6532-01-011-5057 (Youth)

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6532-1203

Location: SLPRC13B.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54642	2. DOCUMENT DATE (YYMMDD) 17 May 1992

3. DOCUMENT TITLE
SLIPPERS, CONVALESCENT PATIENT'S (Polyurethane Foam, Disposable)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/290

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:36Z
Check the source to verify that this is the current version before use.