METRIC

A-A-54646A
19 February 1993
SUPERSEDING
A-A-54646
6 May 1992

COMMERCIAL ITEM DESCRIPTION

SUCTION SET, TRACHEAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers sterile disposable suction kits for aseptic suctioning of the tracheobronchial tree.

Salient Characteristics:

Suction set, tracheal, shall be sterile, disposable, and specifically designed for aseptic suctioning of the tracheobronchial tree. The sets shall be provided in the following sizes and styles:

Size 1 - 14 French (4.7 mm O.D.) suction catheter
Size 2 - 18 French (6.0 mm O.D.) suction catheter

Style A - Consists of a suction catheter and two gloves packaged in a nonwoven inner wrap and a pop-up solution basin. Set is individually packaged in a peel open pouch.

Style B - Consists of a suction catheter, two gloves, a solution cup, and an inner sterile wrap. Set is individually wrapped in a suitable tear resistant package which will allow aseptic handling and maintain the sterility of the contents until opened or damaged.

Style C - Consists of a suction catheter, two gloves packaged in a nonwoven inner wrap, a waterproof towel, and a minimum of 100 ml sterile saline solution (Sodium Chloride For Irrigation, USP). Set is individually packaged in a rigid peel open basin.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Style D - Consists of a suction catheter, two gloves, a solution cup, an inner sterile wrap, and a minimum of 120 ml sterile saline solution (Sodium Chloride For Irrigation, USP) in a foil packet or nonbreakable jar. Set is individually wrapped in a suitable tear resistant package which will allow aseptic handling and maintain the sterility of the contents until opened or damaged.

Suction catheter. The suction catheter shall be 51 to 56 cm long. Shall have finger control valve to adjust and control the amount of suction. Shall be molded in the whistle tip design with two eyes at the distal end. The finger control valve shall fit snugly into the catheter lumen and shall not become disconnected during handling or use. The proximal portion of the finger control valve shall be ridged for secure attachment to suction tubing.

Gloves. The kit shall contain two disposable vinyl or latex gloves large enough to accommodate at least size 8 glove sized hands. The gloves shall be folded back to form a cuff facilitating aseptic donning. Shall be strong and not tear or rip during use. Shall be constructed to provide the finger sensitivity needed for suction control and catheter manipulation. Gloves for Style A and C sets shall be packaged in a nonwoven inner wrap.

Solution cup/basin. The kit shall contain a plastic coated paper solution cup or paper solution cup with a plastic liner with at least a 120 ml capacity. Solution basin for Style A set shall be cardboard pop-up design with clean plastic bag insert.

Inner Wrap. The kit shall contain a functionally folded and marked inner wrap which becomes the sterile field for the suctional procedure.

Waterproof towel. A waterproof towel shall be provided for Style C sets.

Sterile saline irrigation solution. Sodium Chloride For Irrigation, USP shall be provided for Style C and D sets. Style D sets will have minimum of 120 ml solution in a foil packet or nonbreakable jar. Style C sets shall have a minimum of 100 ml solution.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Sterility. Components and the inside surfaces of the package shall be sterile in accordance with USP or AAMI guidelines. Sterility assurance level shall be 0.0001. Saline shall be sterile and nonpyrogenic in accordance with USP.

Assembly. Components for Style B and D sets shall be arranged in the package in order of use.

Workmanship. The suction kits shall be free from defects which detract from their appearance or impair their serviceability. Components shall be free from flash, tears, holes, burrs, rough areas or foreign material.

Unit. Package (PG). One package containing forty Style D or fifty Style B individually packaged sterile suction sets, as specified constitutes one unit. The individual package shall be a sealed tear resistant type container which will allow aseptic handling and capable of maintaining sterility of contents unless the package is damaged or opened. The kit components shall be arranged in the package in the order of use.

Set (SE) (for VA). A single individually packaged suction set, as specified, constitutes one unit. National Stock Numbers (NSN)s 6515-01-005-7329 and 6515-01-207-3773 shall be provided in multiples of 50 or 100 sets. NSN 6515-01-206-9626 shall be provided in multiples of 12 or 36 sets. Number of units in a shipping container shall be as stated in the solicitation or order.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one suction set as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for test shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one suction set. Sampling shall be inspection level S-2 with an AQL of 1.0.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:36Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers, Item Identifications and their respective Sizes and Types are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION	SIZE	STYLE
6515-01-164-6862	SUCTION SET, TRACHEAL, Disposable, Sterile, 40s	1	D
6515-01-342-4870	SUCTION SET, TRACHEAL, Disposable, Sterile, 50s	1	B
6515-01-005-7329	SUCTION SET, TRACHEAL, Disposable, Sterile	1	A
6515-01-207-3773	SUCTION SET, TRACHEAL, Disposable, Sterile	2	A
6515-01-206-9626	SUCTION SET, TRACHEAL, Disposable, Sterile	1	C

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5478
Location: c:\word\nish\suction2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-54646A	19 February 1993

3 DOCUMENT TITLE

SUCTION SET, TRACHEAL

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)

7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME	b TELEPHONE (Include Area Code)
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial: (215) 737-2870 (2) AUTOVON: 444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403, Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 *Previous editions are obsolete* 198/29

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.