METRIC

A-A-54649
15 May 1992

COMMERCIAL ITEM DESCRIPTION

BONE PLATE, Narrow

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers narrow bone plates with self-compressing holes suitable for surgical procedures of the tibia or ulna.

Salient Characteristics.

The bone plates shall have a cross section of 12 x 3.8 mm. Plates shall have slightly curved profiles. Plates with the indicated number of holes shall have the corresponding lengths:

Number of holes	Length of plate (mm)
4	71
5	87
6	103
7	119
8	135
9	155
10	167
12	199

The bone plate holes shall be self-compressing. The holes shall be capable of accepting large, spherical headed bone screws. The end holes shall be capable of accepting cancellous bone screws. The end holes shall be capable of being used with a tension device. Holes shall have same centerline. Centerline shall longitudinally bisect the plate. Distance between holes shall be 16 mm. Distance between two centermost holes shall be 25 mm nominally.

Tolerances shall be as follows: width ± .25 mm, length ± 1.5 mm, thickness ± .1 mm, distance between holes ± .25 mm.

Material. The bone plates shall be cold-rolled conforming to the composition specified in ASTM F139, Grade 2. Grain size of the raw material shall be five or finer.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Surface preparation. Description of surface characteristics and methods of surface preparation shall be in accordance with ASTM F86.

Finish. The completed bone plate shall have mirror finish, properly passivated.

Corrosion properties. Bone plates shall show no evidence of corrosion when subjected to the following tests:

Intergranular corrosion. The intergranular corrosion susceptibility test specified in ASTM F139.

Saline. The bone plates shall be completely immersed into Normal Saline Solution (USP), and while immersed, placed into an oven at a temperature of $37^{\circ} \pm 1^{\circ}C$ for a period of seven days, removed from the oven and examined for corrosion while still immersed.

Boil. The bone plates shall be scrubbed with soap and water, rinsed thoroughly in distilled water, and dried. After cleaning, the instruments shall be boiled in distilled water in a glass beaker for 30 minutes, and allowed to remain submerged for an additional 24 hours.

Workmanship. The bone plates shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One bone plate, with holes specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for test shall be conducted in accordance with MIL-STD-105. Unit of product for test purposes shall be one bone plate. The acceptable quality level shall be 1.0 for an inspection level of S-2.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-165-7178	BONE PLATE, Narrow, 4 Holes, 71 mm Long, 12 mm Wide, 3.8 mm Thick
6515-01-166-0157	BONE PLATE, Narrow, 5 Holes, 87 mm Long, 12 mm Wide, 3.8 mm Thick
6515-01-166-0158	BONE PLATE, Narrow, 6 Holes, 103 mm Long, 12 mm Wide, 3.8 mm Thick

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-166-3529	BONE PLATE, Narrow, 7 Holes, 119 mm Long, 12 mm Wide, 3.8 mm Thick
6515-01-166-3530	BONE PLATE, Narrow, 8 Holes, 135 mm Long, 12 mm Wide, 3.8 mm Thick
6515-01-166-3531	BONE PLATE, Narrow, 9 Holes, 155 mm Long, 12 mm Wide, 3.8 mm Thick
6515-01-166-3532	BONE PLATE, Narrow, 10 Holes, 167 mm Long, 12 mm Wide, 3.8 mm Thick
6515-01-166-3533	BONE PLATE, Narrow, 12 Holes, 199 mm Long, 12 mm Wide, 3.8 mm Thick

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5356
Location: PLATE/S18

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER: A-A-54649	2 DOCUMENT DATE (YYMMDD): 15 May 1992

3 DOCUMENT TITLE: BONE PLATE, Narrow

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME: Defense Personnel Support Center ATTN: DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial: (215)737-2870 (2) AUTOVON: 444-2870
ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703)756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.