INCH-POUND

A-A-54664
29 May 1992

COMMERCIAL ITEM DESCRIPTION

NEEDLE, HYPODERMIC, CARTRIDGE TYPE, Disposable

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-N-36319A(1) which is cancelled.

This commercial item description covers a disposable, sterile, cartridge type hypodermic needle suitable for use in anesthetic procedures.

Salient Characteristics:

General. The needles shall be of the following types and sizes:

Type I - plastic hub, self-threading or threaded
Type II - aluminum hub, threaded

Size 1 - 25 Gage, 13/16 Inch
Size 2 - 25 Gage, 1-3/8 Inch
Size 3 - 27 Gage, 13/16 Inch
Size 4 - 27 Gage, 1-3/8 Inch

The needle shall be compatible with a thumb ring-handled operating cartridge syringe specifically designed to accept needles of the type and size described (See Note 2 and 3).

Material.

Cannula. The cannula shall be fabricated from either seamless drawn or welded and drawn Class 3, Type 304 stainless steel in accordance with ASTM Designation F899.

Hub. The Type I hub shall be fabricated from polypropylene plastic. The Type II hub shall be fabricated from aluminum alloy.

Insert. For Type I hub, with insert, the insert shall be fabricated from aluminum alloy.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Needle container. The needle container, comprised of a needle sheath and cap, shall be fabricated from polypropylene or other plastic. The container shall be capable of maintaining sterility of contents unless it is damaged or opened.

Style, design and dimensions. Each cartridge-type needle shall be packaged in an individual sealed container. Each needle container shall consist of a sheath and cap. The needle shall be suitable for use with a thumb ring aspirating cartridge syringe specifically designed to accept needles of the type and size described. The needle shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 20467.

Corrosion resistance. The cannula shall show no signs of corrosion when boil tested in accordance with ASTM Designation F1089. The Type II hub shall show no signs of discoloration or reaction when immersed in 3% sodium chloride solution at 20-25 degrees C after one hour.

Instructions. Each unit of issue shall contain instructions for use which include proper disposal.

Stiffness. When tested as specified, the maximum cannula deflection shall be as follows:

for	27 gage	0.022 inch
	25 gage	0.026 inch

Security of staking. Upon applying an axial tensile pull of 5 pounds for one minute, the cannula or aluminum insert (for Type I needle) shall not loosen or pull out of the hub.

Needle bending resistance test. The needle cannula shall not crack or separate from the needle hub when bent 25 degrees from its rest position and then bent 25 degrees in the opposite direction.

Sterility, pyrogencity, toxicity. The needle shall be sterile in accordance with the USP or AAMI guidelines. The sterility assurance level shall be 0.0001. The needle shall be nonpyrogenic and nontoxic when tested in accordance with the USP.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the currentversion before use.

Identification marking and labeling. Each needle container shall be marked with the name or registered trademark of the manufacturer, gage and needle length (i.e. short or long), and "Disposable after one use" or similar legend.

Workmanship. Needles shall be free from defects which detract from their appearance or may impair their serviceability. The cannula shall show a smooth finished outer surface free from pits and tool marks, burrs, scale or corrosion, and inner surface free from scale, deep pits, deep furrows, or die marks. The lumen of the cannula and the hub cavity shall be free from burrs, chips, dirt, slivers or other foreign matter. Needles shall be sharp, not blunt or bent, when examined under 7 power magnification.

PRODUCT QUALITY LEVEL

Visual and dimensional examination. The needle shall be examined to verify that the design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements as specified.

Unit. Package (PG). One package containing one hundred individual needles of the size specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one needle. Sampling shall be inspection level II with an acceptable quality level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one needle. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Stiffness test. The cannula shall be placed across two test edges (0.002 inch radius, maximum) separated by 0.375 inch $\pm$ 0.003 inch for the 27 gage, and 0.500 inch $\pm$ 0.003 inch for the 25 gage. A load of 20 ounces $\pm$ $^1/_4$ ounce for the 27 gage, or 25 ounces $\pm$ $^1/_4$ ounce for the 25 gage, shall be applied at the center of the span. The resulting maximum deflection of the cannula shall be measured using a gage with graduations of at least 0.001 inch. The force shall be applied though an edge with a maximum radius of 0.036 inch and centered to within $\pm$ 0.003 inch.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order.

NOTE 1: The following National Stock Numbers, their respective Item Identifications, Types and Sizes are covered by this document:

NSN	ITEM IDENTIFICATION	TYPE	SIZE
6515-00-181-7403	Needle, Hypodermic, Cartridge Type, Disposable, 25 gage, 13/16 inch long, 100s	I	1
6515-00-181-7404	Needle, Hypodermic, Cartridge Type, Disposable, 25 gage, 1-3/8 inch long, 100s	I, II	2
6515-00-181-7425	Needle, Hypodermic, Cartridge Type, Disposable, 27 gage, $^{13}/_{16}$ inch long, 100s	I, II	3
6515-00-181-7412	Needle, Hypodermic, Cartridge Type, Disposable, 27 gage, $1-^{3}/_{8}$ inch long, 100s	I, II	4

NOTE 2: A National Stock Number for such a syringe is 6515-01-010-8761, Syringe, Cartridge, Aspirating, Thumb Ring Handle.

NOTE 3: A National Stock Number for a cartridge to be used with the syringe in Note 2 and the needle described in the specification is 6505-01-146-1139, Lidocaine Hydrochloride and Epinephrine Injection, USP.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-5067
Location: PL Disk #2
(NEED-HYP)

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54664	5-29-92

3. DOCUMENT TITLE

Needle, Hypodermic, Cartridge Type, Disposable

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/290

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the current version before use.