A-A-53268
7 December 1987

COMMERCIAL ITEM DESCRIPTION

PRESS, DENTURE FLASK

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a denture flask press suitable for holding resin cases under pressure during the curing process.

Salient Characteristics:

Materials. Materials shall conform to the applicable requirements as specified in Table I. Materials not specifically designated in Table I shall be of high commercial quality throughout and shall be suitable for the purpose intended.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

TABLE I. Materials.

Components	Materials	Specification
Frame	Bronze casting or Aluminum Bronze	ASTM B 584, Alloy 865 ASTM B 148, Alloy 953
Lifting Loop	Bronze wire	------
Pressure screw	Copper silicon alloy	QQ-C-591, Alloy 661 or 655
Upper pressure plate	Bronze casting or Aluminum Bronze	ASTM B 584, Alloy 865 ASTM B 148, Alloy 953
Lower pressure plate	Bronze casting or Aluminum Bronze	ASTM B 584, Alloy 865 ASTM B 148, Alloy 953
Coil Spring	Corrosion-resistant steel, 18-8	------
Wrench	Cold rolled steel, B1112, C1022 or 1215 Nickel plated	FED-STD-66
Holding Handle Rod	Cold rolled steel, Nickel plated	------
Ferrule	Steel, Nickel plated	------
Handle	Hardwood	------

Design and dimensions. The flask press shall satisfactorily accommodate two upper or two lower ejector-type denture flasks, in accordance with Defense Personnel Support Center (DPSC) Drawing Nos. 20485 (for upper flask) and 20125 (for lower flask). The flask press shall consist of a frame, lifting loop, pressure screw, upper pressure plate, lower pressure plate, two coil springs, wrench, and holding handle. The frame base, top and sides shall be of one-piece cast construction and the upper and lower pressure plates shall be a one-piece casting. The pressure screw shall be permanently attached to the upper pressure plate. The press shall conform to Defense Personnel Support Center (DPSC) Drawing No. 20638 for design, dimensions and construction.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Spring compression. Each spring shall be capable of sustaining a compression load such that a free length of 1-5/16 ± 1/16 inches shall be compressed to a height of 27/32 ± 1/32 inch, when tested as specified herein.

Finish. Casting surfaces shall be substantially smooth. Castings shall be free of blowholes, porosity, hard spots, shrinkage defects, cracks, sand, dirt, fins, scales and other defects or extraneous materials. The hex wrench hole and other components shall be smoothly finished. All sharp edges shall be smoothly rounded.

Identification marking. Each unit shall be permanently and legibly marked with the name or registered trademark of the manufacturer.

Workmanship. The press shall be free from defects which detract from its appearance or may impair its serviceability.

Unit. Each (EA). One press, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one press. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

TABLE II. Sampling for Examination.

Type Defect	Inspection Level	AQL (Percent Defective)
Major A	II	1.0
Major B	S-2	4.0
Minor A	II	2.5
Minor B	S-2	6.5

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table III for the press. The acceptance number shall be zero for all sample sizes. The unit of product for test purposes shall be one press. Where feasible, the same sample unit may be used for the determination of two or more test characteristics.

TABLE III. Sample criteria for Testing.

Component	Characteristic	Inspection Level
All parts in Table I	Chemical Analysis	*
Coil springs	Compression	S-1

* A sufficient amount from each lot of raw materials used in the construction of the item shall be selected for chemical analysis.

Examination. The presses shall be examined to determine compliance with the requirements of this specification. Examination shall be conducted in accordance with the following classification of defects. Examination shall include, but shall not be limited to, the defects listed in Table IV.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

TABLE IV. Classification of Defects.

Categories	Defects
Critical	None defined.
Major A	
101	Style and design of press not as specified.
102	Component missing.
103	Component damaged, affecting serviceability.
104	Upper pressure plate does not lower and raise easily.
Major B	
151	Minimum distance between contact surfaces of plates before exerting pressure not as specified.
152	Minimum distance between sides of frame not as specified.
Minor A	
201	Finish not as specified.
202	Identification marking not as specified.
203	Packaging and packing not as specified.
204	Marking on packaging and pack not as specified.
Minor B	
251	Dimensions (other than 151 and 152) not as specified.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Test procedures.

Chemical analysis. In the absence of certificates of quality, chemical analysis of the individual components of the press shall be determined in accordance with Method 111 or 112 of FED-STD-151. In the event of a conflict, Method 111 shall govern.

Spring compression test. Remove the two stainless steel coil springs by disengaging the two screws from the lower pressure plate assembly. Each spring shall be tested individually by mounting in a suitable apparatus such that a compressive force of 200 pounds is applied. The height of the compressed spring shall be measured.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6520-00-515-5150	PRESS, DENTURE FLASK

Ordering data. Intermediate/exterior package quantities and labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

47900

Project No. 6520-2021
Location: DISKETTE - CID #4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.