A-A-54366
15 April 1991

# COMMERCIAL ITEM DESCRIPTION

## CANISTER ASSEMBLY, SUCTION, SURGICAL

(One Canister with Fifty Liners)

The General Services Administration has authorized the use of this Commercial Item Description.

Salient Characteristics:

The canister assembly shall consist of one (1) reusable, 3000 ml capacity canister and fifty (50) 3000 ml capacity, disposable liners The assembly shall be designed for collection and aseptic disposal of fluids during surgical procedures

The canister shall be fabricated of rigid, clear, crack-resistant plastic and shall be autoclavable for repeated reuse It shall be antistatic and shall be marked to indicate level of aspirated contents at least every 100 ml for easy monitoring Each canister shall be tapered to allow for nesting during storage

Each liner shall be made of tough, soft plastic which conforms to the canister during use It shall have a snap-seal lid which locks onto the canister to prevent spills and leaks; each liner shall include a built-in filter to minimize bacterial spray into the environment The lid shall include a right angle swivel connector to eliminate tube kinking The connections shall be labeled to prevent incorrect assembly The liner shall have an accessory port/pour spout to allow the contents to be emptied in an isolated area and the liner unit shall be disposable via incineration

Packaging and labeling shall meet the requirements of the Federal Food, Drug and Cosmetic Act and shall be that normally furnished by the supplier as commercial practice

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Workmanship The canister assembly shall be free from defects which detract from its appearance or impair its serviceability

Unit. Package (PG) One canister assembly, consisting of 1 canister and 50 liners, as specified, constitutes one unit

Quality Assurance Provisions.

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of II The unit of product for examination purposes shall be one canister assembly, as specified in this description

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of II The unit of product for test purposes shall be one canister assembly



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics As a minimum, the tests recommended by the manufacture shall be performed

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

NOTE: The following National Stock Number is covered by this document

6515-01-133-7430

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA - OSS
USPHS
FDA - MPQAS

Project Number: 6515-4992

Location: CANISCID.12B



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:33Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

## INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

| I. RECOMMEND A CHANGE: | 1. DOCUMENT NUMBER | 2. DOCUMENT DATE (YYMMDD) |
|---|---|---|
| | A-A-54366 | 15 April 1991 |

3. DOCUMENT TITLE
CANISTER ASSEMBLY, SUCTION, SURGICAL (One Canister with Fifty Liners)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

| a. NAME (Last, First, Middle Initial) | b. ORGANIZATION |
|---|---|
| c. ADDRESS (Include Zip Code) | d. TELEPHONE (Include Area Code)<br>(1) Commercial<br>(2) AUTOVON |
| | 7. DATE SUBMITTED (YYMMDD) |

8. PREPARING ACTIVITY

| a. NAME | b. TELEPHONE (Include Area Code) |
|---|---|
| DEFENSE PERSONNEL SUPPORT CENTER<br>ATTN: DPSC-MSTE | (1) Commercial: AC 215-737-2870<br>(2) AUTOVON: AV 444-2870 |
| c. ADDRESS (Include Zip Code)<br>2800 SO. 20TH STREET<br>PHILADELPHIA, PA 19101-8419 | IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT:<br>Defense Quality and Standardization Office<br>5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466<br>Telephone (703) 756-2340 AUTOVON 289-2340 |

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:33Z
Check the source to verify that this is the current version before use.