A-A-54371
15 April 1991

COMMERCIAL ITEM DESCRIPTION

HEATER, HEAT TREATMENT PAD

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers a mobile, lidded stainless steel apparatus, with water heating system for preparing moist heat treatment pads used in physical therapy

Salient Characteristics:

The apparatus shall be suitable for preparation of moist heat treatment pads for use in physical therapy It shall be a completely integrated, mobile, lidded, stainless steel container, temperature controlled electrical water heating system.

The apparatus shall be equipped with at least the following:

a Container

b Hinged container lid with outer joint, handle, and opening stop

c Electrical heating circuit with:

* Heater
* Thermostat
* Temperature control set-point adjustment mechanism
* Low water and high temperature safety cut-off mechanism
* Safety cut-off indicator light
* ON/OFF switch with pilot light
* Power cord with plug

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

d Removable internal rack

e External side rack

f Drain valve with handle and a male hose fitting

g Hose with male and female fitting at opposite ends

h Casters

i Twenty-four heating pads gel-filled

The apparatus container and both racks shall be constructed of corrosion-resisting metal (stainless steel 300 series). The approximate dimensions of the container shall be 20 inches wide by 32 inches long by 33 inches high The external side rack shall be approximately 20 inches wide by 13-3/4 inches deep by 28-1/4 inches high.

The working water capacity of the container without pads shall be 50 gallons minimum, with heater rated at 1500 watts minimum Thermostatically controlled water temperature shall be 160°F Capability to change control temperature set-point shall be provided by a temperature adjustment mechanism . A secondary safety feature shall interrupt electrical heater power if water level becomes low or if water temperature exceeds 180°F

The removable internal rack shall be of proper size and strength to support at least 24 fully soaked, silica gel-filled heating pads of the construction, size, and weight described below This rack shall be designed to allow for easy pad placement or removal, one at a time if desirable, without the need to move other pads or to remove the internal rack from the cabinet

The external side rack shall contain 3 shelves, with the top shelf of sheet metal The other two shelves shall be sheet metal or wire rack type Shelves shall be suitable for storage of towels and heating pads Total weight of shelf unit shall be 20 pounds maximum and shall withstand a load of at least 25 pounds Mounting of this external rack shall be accomplished by hanging from the built-in, slip-on handle

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

The heating pads shall be fabricated from 100 percent cotton material, with construction of 70 ends per inch minimum and 58 picks per inch minimum, weight of 10 0 oz. minimum The size of the pads shall be approximately 10 inches by 12 inches and shall be sewn, quilt-like, into approximately 12 rectangular compartments (2 rows of 6) Each compartment shall contain approximately equal amounts of dry gel totaling 17 oz minimum per heating pad The gel shall be a commercial hydrous aluminum silica, evenly sized pellets The pellet size shall be such that it shall be retained within the pad described above

Loops shall be sewn on each corner of all pads The loop material shall be 100 percent cotton and shall be color fast to 200°F The two loops sewn onto one edge of the pad (the 12 inch pad dimension) shall be a dark color; the two loops sewn on the opposite edge shall be white These loops shall enable the pads to hang from hooks on the internal rack and shall support the full weight of the completely soaked pad (approximately 5 pounds)

The entire apparatus shall not weigh more than 140 pounds without heating pads

The apparatus shall operate on 115 volts, 50-60 Hz electric power A three conductor power cord at least six feet long, with hospital grade plug shall be provided

All items shall conform to applicable requirements of a nationally recognized standard for the safety of medical equipment of the issue in effect on the date of solicitation All items shall be approved as conforming to such a standard by an independent nationally recognized testing laboratory

The apparatus shall be furnished with two (2) complete sets of instructions for operation, maintenance, and repair

Finish The apparatus shall be fabricated from materials which are inherently resistant to corrosion, or which are treated to render them corrosion-resistant Where painting or plating is used, appropriate cleaning and priming procedures shall be performed prior to finishing Finish shall be equivalent or superior to that of comparable commercial equipment, and shall be free of gaps, scratches chips, rust spots, pitting, and other imperfections.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Workmanship. The apparatus shall be free from defects which detract from its appearance or impair its serviceability

Unit Each (EA) One heater as specified, constitutes one unit

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved be the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of II The unit of product for examination purpose shall be one heater as specified in this description

Sampling for tests Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of II The unit of product for test purposes shall be one heater

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics Tests shall include but shall not be limited to the following·

Operation test The apparatus shall be filled with 50 gallons of water (65°F to 75°F) Room temperature shall be between 65°F to 75°F Water shall reach 160°F within 10 hours and maintain controlled temperature for 8 hours

Dielectric test Immediately following operation test, the unit shall be subjected to a dielectric withstanding test of 1000 volts between current carrying parts and dead metal parts The potential shall be slowly raised from 0 to 1000 volts and maintained at that value for one minute The power switch shall be in the 'on' position when test is conducted.

Risk current test The risk current test shall be conducted after performing the dielectric withstanding test. The test procedures and test equipment used shall comply with all applicable requirements of AAMI ESI, Safe Current Limits for Electromedical Apparatus Risk current shall not exceed a limit of 100 microamperes as frequency weighted by the AAMI Standard previously mentioned

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

NOTE The following National Stock Number is covered by this document

NSN 6530-01-204-8690

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6530-2239

Location: HEATCID 1DB

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54371	2. DOCUMENT DATE (YYMMDD): 15 April 1991

3. DOCUMENT TITLE: HEATER, HEAT TREATMENT PAD

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	e. DATE SUBMITTED (YYMMDD)

7. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSTE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19101-8419	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.