INCH-POUND

A-A-54372A
15 January 1993
SUPERSEDING
A-A-54372
15 April 1991

COMMERCIAL ITEM DESCRIPTION

MASK, SURGICAL

(Pleated)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a disposable, pleated, nonwoven fabric, surgical mask suitable for use by medical personnel in operating rooms to provide aseptic protection.

Salient characteristics:

The mask shall consist of a filter mat encased and bonded in two soft layers of nonwoven fabric. The filter media shall consist of a filter mat composed of polypropylene fibers. The outside facing material shall be a nonwoven fabric of suitable fibers. The mask shall have a minimum of three 1/2 inch pleats along its horizontal width to allow for a snug fit at the edges and without binding at the nose or mouth. The peripheral margins of the mask shall be bound with nonwoven tape by means of stitching, welding or gluing. The tape along the horizontal margins shall be extended to serve as tie tapes. The four tie tapes shall be a minimum of 15 inches long from the edge. A nosepiece shall be encased within the facing material and centered along the top horizontal edge of the mask. The nosepiece shall be fabricated from a single strip of malleable aluminum rounded on both ends. The strip shall be flexible enough to be readily bent to the contour of the nose and hold the mask snugly against the face to prevent fogging of spectacles. The color of the mask shall be blue, green or other pastel shades.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

The finished mask shall have the following minimum dimensions:

Horizontal (length)	-	6-1/2 inches
Vertical (width)	-	3-1/2 inches

(Measurements shall be taken when the mask is flat and pleats are fully folded).

The air exchange pressure, Delta P, of the mask material shall be no more than 3.0 mm of water pressure when tested in accordance with any commercially acceptable method.

The filtering efficiency of the mask shall not be less than 95 percent when tested by the method of Greene and Vesley (J. Bacteriol, Vol. 83. No. 3, p. 663, March 1982) or a comparable method to permit free breathing and unrestrained communication.

The flammability requirements of the mask shall be in accordance with class 1 of Code of Federal Regulations (CFR) 1610 Part 16 and the time of the flame spread shall be 4 seconds or more.

The mask shall have no objectionable odor and shall not cause irritation to the skin or eyes when tested using human panel and rabbit subjects in compliance with Federal Regulation CFR 16 1500.41 and 1500.42.

Workmanship. The masks shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing six containers of fifty masks, as specified, constitutes one unit. Each container shall permit one-at-a-time removal of each mask.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one package of surgical masks.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.5 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one package of surgical masks.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

1. Filtering efficiency.
2. Flammability.
3. Dermatological reactions.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:33Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6515-01-003-3142

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:33Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6515-5484

Location: MASK15C.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54372A	2. DOCUMENT DATE (YYMMDD) 15 January 1993

3. DOCUMENT TITLE
MASK, SURGICAL (Pleated)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870	(2) AUTOVON AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.