A-A-54377
22 April 1991

# COMMERCIAL ITEM DESCRIPTION

## FILE, BONE,(DENTAL, DOUBLE-ENDED) Seldin, No. 11

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-F-36733A which is cancelled.

This commercial item description covers a file of the Seldin design suitable for filing the bone processes of the jaw in oral surgery.

Salient characteristics:

Material. Files shall be fabricated entirely of class 4, type 440A martensitic stainless steel in accordance with ASTM standard F899, for chemical composition only.

Hardness. After proper stress relieving, the operative ends of the file shall have a Rockwell hardness of C52 to C56, when tested as specified in ASTM E18 or E92. Handles and shanks shall be of proper temper and shall have a hardness suitable for the purpose intended.

Style, design, and dimensions. The bone file shall be of Seldin No. 11 design. The two blades shall be contra-angled from the centerline of the handle and shall have deep teeth faced toward the handle for a pull stroke. The design and dimensions of the file shall conform to Defense Personnel Support Center (DPSC) Drawing No. 10072.

Construction. The file shall be of one-piece or three-piece, double-ended construction. The handle shall be solid. Each file shall be wrapped in neutral tissue or other suitable anti-corrosive material and shall be packaged in a close fitting container to prevent damage to the operative ends.

Finish. The teeth of the operative ends shall be sharp and clean cut with a fine, honed finish. The remainder of the operative ends and the handles and shanks shall have a satin or bright polished finish.

Corrosion resistance. Corrosion resistance tests in accordance with ASTM F1089 shall be applicable. The instruments shall be properly passivated and shall meet the boil test requirement and copper sulfate test requirement.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Identification marking. Files shall be permanently and legibly marked on the handle of the instrument. Marking shall include the name or registered trademark of the manufacturer or contractor, the instrument number, and the word "stainless". Instruments of foreign manufacture shall bear the name of the country of origin.

Workmanship. The files shall be free from cracks, burrs, nicks, pits, or other defects which detract from their appearance or may impair their serviceability.

Unit. Each (EA). One file, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Certificates of quality. Certificates of quality for the raw material, supplied by the manufacturer of the steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one file. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, except for dimensions, which shall be conducted at inspection level S-1, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-2. For tests, the acceptance number shall be zero (0) for all sample sizes.

Examination. Files shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Number is covered by this document:

| National Stock Number | Item Identification |
|---|---|
| 6520-00-528-5050 | FILE, BONE, Seldin, No. 11 |

Ordering data. (Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2353
LOCATION: WORD/FILENO11

PGC 47966



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

| I RECOMMEND A CHANGE: | 1. DOCUMENT NUMBER: A-A-54377 | 2. DOCUMENT DATE (YYMMDD): 22 April 1991 |
|---|---|---|

3. DOCUMENT TITLE: FILE, BONE (Dental, Double-Ended) Seldin, No. 11

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

| a. NAME (Last, First, Middle Initial) | b. ORGANIZATION |
|---|---|
| c. ADDRESS (Include Zip Code) | d. TELEPHONE (Include Area Code)<br>(1) Commercial<br>(2) AUTOVON (If applicable) |
| | 7. DATE SUBMITTED (YYMMDD) |

8. PREPARING ACTIVITY

| a. NAME | b. TELEPHONE (Include Area Code)<br>(1) Commercial | (2) AUTOVON |
|---|---|---|
| Defense Personnel Support Center | (215)737-2870 | 444-2870 |
| c. ADDRESS (Include Zip Code)<br>ATTN: DPSC-MSTE<br>2800 S. 20th Street<br>Philadelphia, PA 19101 | IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:<br>Defense Quality and Standardization Office<br>5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466<br>Telephone (703) 756-2340 AUTOVON 289-2340 | |

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.