A-A-54378
22 April 1991

COMMERCIAL ITEM DESCRIPTION

CURETTE SET, ANTRUM, COAKLEY

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-C-42020A which is cancelled.

This commercial item description covers a set of six antrum curettes of the Coakley design suitable for use in ENT procedures

Salient Characteristics:

General The set shall consist of the following curettes with the designated style:

Style 1 - Large, light curve
Style 2 - Large, strong curve
Style 3 - Large, acute angle
Style 4 - Small, light curve
Style 5 - Small, strong curve
Style 6 - Triangular

Material Curettes shall be manufactured from Class 4 martensitic corrosion-resistant steel. As an alternate, handles may be manufactured from Class 3 austenitic corrosion-resistant steel Steels shall be as specified in ASTM Designation F899

Hardness The loops of the curette set shall have a Rockwell Hardness C scale (HRC) of 40 minimum when tested in accordance with ASTM Designation E18 or E92 whichever is applicable. Conversion of hardness values shall be in accordance with ASTM Designation E140

Style, design and dimensions The curette set shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No 24538 for style, design and dimensions.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A Approved for public release, distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Construction Construction shall be in accordance with DPSC Drawing No 24538 and as indicated herein

Handles and shanks The handles and shanks shall be fabricated as a single component or may be designed as a two piece assembly, with solid or hollow handles When constructed of more than one piece, the shanks shall be securely attached to the handles.

Loops The loops shall be integral to or securely and firmly attached to the shanks Where the loops enter the shanks and for all curettes, there shall be no crevices or spaces where foreign material may collect or which may be subject to corrosion

Passivation The curette set shall be properly passivated.

Finish. The shanks and loops shall have a mirror finish Handles shall be knurled to ease in the handling of the curettes The finish shall be uniform and free from grind or tool marks

Corrosion resistance. Curettes shall show no signs of corrosion or copper plating when subjected to the boil and copper sulfate tests, respectively, specified in ASTM Designation F1089.

Identification marking and labeling The curettes shall be legibly marked with the contractor's name or trademark and the word 'stainless' Foreign manufactured curettes shall bear the name of the country of origin.

Workmanship The curettes shall be free from any defect which detract from its appearance or impair its serviceability The curettes shall be without rough areas, edges, burrs, cracks, pits, dents, and corrosion.

Product Quality Level

Visual and dimensional examination The curette set shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements

Unit Set (SE) One curette set, as specified, constitutes one unit

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one curette set. Sampling shall be inspection level II with an acceptable quality level (AQL) of 1.0.

Examination Examination shall be conducted to determine compliance with specification.

Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein Unit of product for test purposes shall be one curette set. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for two or more test characteristics

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and therefore as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U.S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, and labeling and marking
Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable quantities, labels and marking shall be as specified in the contract and/or order.

NOTE The following National Stock Number is covered by this document

NSN	ITEM IDENTIFICATION
6515-01-110-1382	CURETTE SET, ANTRUM, COAKLEY Set of 6

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent.

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No 6515-5048

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54378	2. DOCUMENT DATE (YYMMDD) 4-22-91

3. DOCUMENT TITLE
CURETTE SET, ANTRUM, COAKLEY

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center	b. TELEPHONE (Include Area Code) (1) Commercial: (215)737-2870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-VSIE 2800 S. 20th Street Philadelphia, PA 19101	d. IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous edition is obsolete. — 9870/0

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.