A-A-54381
22 April 1991

COMMERCIAL ITEM DESCRIPTION

CRUTCH, CANE TYPE, ADJUSTABLE, CHILD

The General Services Administration has authorized the use of this comercial item description as a replacment for Military Specification MIL-C-37277 (DSA-DM) which is cancelled.

This commercial item description covers an adjustable child's aluminum, cane type crutch suitable for use in ambulation.

Salient Characteristics:

Material. Crutch components material shall be furnished as shown on Defense Personnel Support Center (DPSC) Drawing No. 24320. Rubber and vinyl parts shall be black in color.

Construction. The crutch shall consist of a telescoping aluminum tube construction with projecting plastic covered hand grip, and a C-shaped arm support. Bottom of crutch shall be fitted with a removeable rubber crutch tip. Starting above hand grip, tubing shall bend slightly to provide comfortable positioning of arm support which shall be attached by means of an anchor block to an adjusting post telescoping into the upper end of tubing. Adjusting post shall be adjustable to three different heights. The lower portion of the crutch shall also be adjustable in length, to eight positions. The crutch shall conform to Defense Personnel Support Center Drawing 24320. The telescoping crutch sections shall move freely and not bind when moved to all positions. The positioning buttons shall move easily and shall not bind in lower and upper crutch height adjustment. Crutch shall be provided assembled as shown on the drawing. Crutch shall have weight distribution and constuction to permit hanging from a chair or table.

Finish. All exposed parts of the crutch shall have a smooth finish. Aluminum components shall be suitably anodized.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Workmanship. The crutch shall be free from defects which detract from its appearance or impair its serviceability. Crutch shall be free from burs, sharp edges, scratches, deformity, corrosion or foreign material.

Unit. Pair (PR). One pair of crutches, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one crutch. Sampling shall be inspection level II with an acceptable quality level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one crutch. Sampling shall be inspection level S-2 with an AQL of 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-116-0563	CRUTCH, CANE TYPE, ADJUSTABLE, CHILD

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-5004

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMENDED CHANGE:	1. DOCUMENT NUMBER: A-A-54361	2. DOCUMENT DATE (YYMMDD): 4-22-91

3. DOCUMENT TITLE: Crutch, Cane Type, Adjustable, Child

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Street, City, State, ZIP Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME: Defense Personnel Support Center	b. TELEPHONE (Include Area Code) (1) Commercial: (215)737-2870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code): ATTN: DPSC-MSTE, 2800 S. 20th Street, Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office, 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466, Telephone (703)756-2340, AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete. 98-290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.