A-A-54382
22 April 1991

COMMERCIAL ITEM DESCRIPTION

ANESTHESIA SET SPINAL (Saddle Block, Disposable)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile, disposable, saddle block spinal anesthesia set suitable for use in anesthetizing patient for surgical procedures.

Salient Characteristics:

General. The set shall consist of a compartmented plastic tray or trays, containing the following minimum components, overwrapped with a durable plastic material

One (1) each 25 gage by 3-1/2 inch Quincke needle with stylet.
One (1) each 19 gage by 1-1/2 inch thin wall needle, spinal.
One (1) each 20 gage by 1-1/2 inch hypodermic needle.
One (1) each 25 gage by 5/8 inch hypodermic needle.
One (1) each ampule Lidocaine HCl 5% with Dextrose 7.5%, 2 ml.
One (1) each ampule Epinephrine 0.1%, 1ml.
One (1) each ampule Ephedrine Sulfate 5%, 1 ml.
One (1) each ampule Procaine HCl 1%, 2ml.
One (1) each 5 ml, glass, Luer-slip syringe.
One (1) each 3.0 ml, plastic, Luer-slip syringe.
One (1) each 5 micron filter.
Three (3) each sponge assembles.
One (1) solution cup.
Three (3) each 3 by 3 inch gauze sponges.
Two (2) each paper towels.
One (1) each fenstrated, non-woven drape.

The tray/s shall be recessed to receive and retain components in position for proper use sequence.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Material. The glass ampuls shall be Type I or Type III, style 7, prescored, meeting the requirements of the Pharmacopeia of the United States (USP). The ampuls shall be made of molded glass or drawn glass tubing, with walls of uniform thickness with drawn, rolled-in, or stuck-on necks. Ampuls shall stand upright, at right angles to the supporting surface, without wobbling. Ampuls shall be supplied ready for opening and shall break cleanly.

The plastics shall meet the requirements of the USP for Class II plastics.

The needles shall be austenitic and in accordance with ASTM Designation F899. The needles shall be furnished with protectors.

The components of the set shall be within opaque packaging capable of maintaining the sterility of the tray and contents until manually open.

The anesthesia sets shall be furnished with no less than 20 months to the expiration date at the time of delivery to the Government.

Sterility Components shall be sterile in accordance with USP or AAMI guidelines and have a minimum sterility assurance level of 0.000001.

Workmanship. The anesthesia sets shall be free from defects which defect from their appearance or impair their serviceability. Needles shall be without burrs or signs of corrosion.

Unit Package (PG.) One package containing ten packaged sterile anesthesia sets, as specified, constitutes one unit. The package container shall be capable of maintaining sterility of contents unless the package is damaged or opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL_STD-105, and as specified herein. Unit of product for examination purposes shall be one anesthesia set. Sampling shall be inspection level II with an acceptable quality level (AQL) of 2.5.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one anesthesia set. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-075-6600	ANESTHESIA SET, SPINAL, Saddle Block, Disposable, Sterile, 10s

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-5014

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54382	2. DOCUMENT DATE (YYMMDD): 4-22-91

3. DOCUMENT TITLE: ANESTHESIA SET SPINAL (Saddle Block, Disposable)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) / e. DATE SUBMITTED (YYMMDD)

7. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code): ATTN DPSC-MSTE, 2800 S. 20th Street, Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office, 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466, Telephone (703) 756-2340, AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 96/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.