A-A-54383
10 June 1991

COMMERCIAL ITEM DESCRIPTION

FACE-BOW, ARTICULATOR, DENTAL

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-A-664 which is cancelled.

This Commercial Item Description covers a lightweight dental articulator face-bow.

Salient Characteristics:

Material. The face-bow shall be fabricated from aluminum.

Style and design. Shall be a lightweight face-bow suitable for registering and transferring anatomical relationships from a patient to the adjustable articulator. Shall be designed to incorporate calibrated, adjustable condylar rods to permit symmetrical placement of the face-bow on the face of a patient in relationship to the center of the mandibular condyle heads. The face-bow shall be supplied with a bitefork, orbital pointer, 2 condylar rods, orbital point clamp, bitefork clamp and bow.

The orbital pointer clamp shall be adjustable and shall be designed to permit locking of the orbital pointer in relationship to the infraorbital notch of the patient. The adjustable bitefork clamp shall permit locking of the bitefork in relationship to the maxillary arch of the patient. The bitefork clamp shall have an elevating screw designed to adjust the orbital pointer to the level of the orbital indicator of the articulator and to support the forward portion of the face-bow during mounting of the maxillary cast to the articulator. The bitefork shall be designed for use in creating a positive imprint in softened wax, when the prongs are covered with wax and are interposed between the occlusal rims of the mouth of the patient.

The style, design and dimensions of the face-bow shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24135.

Finish. All screws used to make manual adjustments to the face-bow shall be knurled. The rest of the face-bow shall have a smooth finish.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Marking. The condylar rods shall be graduated in millimeters over a 5 centimeter range from 4 to 9 centimeters. Each whole centimeter shall be marked with the applicable number.

Workmanship. The face-bow shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One face-bow, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and Inspection level of S-2 with an AQL of 1.0 (percent defective). The unit of product for test purposes shall be one face-bow.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and an Inspection level of II with an AQL of 2.5 (percent defective) except for dimensions, which shall be conducted at inspection level S-1 with an AQL of 1.0 (percent defective). Unit of product for examination purposes shall be one face-bow.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by it subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statues, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging packing, labeling, and marking. Unless other wise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

National Stock Number	Item Identification
6520-00-935-5826	FACE BOW, ARTICULATOR

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6520-2385
Location: ML#1/FACE-BOW

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54383	2. DOCUMENT DATE (YYMMDD) 10 June 1991

3. DOCUMENT TITLE
FACE-BOW, ARTICULATOR, DENTAL

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. [redacted]	b. ORGANIZATION
c. ADDRESS (Include Zip Code) [redacted]	d. TELEPHONE (Include Area Code) (1) Commercial [redacted] (2) AUTOVON (If applicable) [redacted]
	7. DATE SUBMITTED (YYMMDD) [redacted]

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center	b. TELEPHONE (Include Area Code) (1) Commercial: (215)737-2870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSTE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 90/790

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.