INCH-POUND

A-A-54384
15 April 1994

COMMERCIAL ITEM DESCRIPTION

POINTS, ENDODONTIC, PAPER, STERILE

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Interim Federal Specification GG-P-00581D, which is cancelled.

This Commercial Item Description covers sterile, absorbent, paper points suitable for use during endodontic procedures. Points shall be one of six sizes, and packaged in quantities of 200, package size 1, or 2400, package size 2 (see notes).

Salient characteristics:

Material. The points shall be fabricated of a fine grade of white bibulous paper which has a smooth, lintless surface and a close texture. The paper shall not contain supplemental chemicals or starches. When tested in accordance with the American Society for Testing and Materials (ASTM) D646, the paper shall have a basis weight as follows:

Point size	Basis Weight in Pounds
Size XXF - Extra Extra Fine	8.5 ± 0.9
Size XF - Extra Fine	10.0 ± 1.4
Size F - Fine	10.0 ± 1.4
Size M - Medium	10.0 ± 1.4
Size C - Coarse	10.0 ± 1.4
Size XC - Extra Coarse	15.0 ± 3.1

The ream size shall be 480 to 500 sheets, with each sheet 24 inches by 36 inches, for all weights.

Style, design, and dimensions. The points shall be of a design which facilitates rapid, extensive absorption of blood, exudates, and medicaments. The points shall be uniformly tapered and machine rolled tightly. Point dimensions shall be as follows:

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Size	Largest diameter of Distal End (Inches)	Smallest diameter of Operative End	Length ± 1/16 (Inches)
XXF	0.023 ± 0.004	0.007 ± 0.004	1
XF	0.028 ± 0.004	0.009 ± 0.004	1
F	0.033 ± 0.004	0.012 ± 0.004	1
M	0.038 ± 0.005	0.016 ± 0.005	1
C	0.045 ± 0.008	0.019 ± 0.006	1-1/8
XC	0.055 ± 0.011	0.022 ± 0.006	1-1/8

The sterile points shall be packaged in sealed cards which are divided into individual cells to preserve sterility. Each cell shall contain not more than 5 sterile paper points. Packaging shall be such that points may be used in multiples of 5 without impairing the sterility of points in other cells. Each package shall contain points of one size only, in accordance with the applicable NSN (see notes). There shall be no holes, channels or tears opening into any cell.

Physical properties. The sterile points shall be firm and stiff enough to permit insertion into a cavity or pulp canal without collapsing or breaking. The points shall retain their form (with no signs of disintegration) after having remained completely immersed in 100 ± 2°F water for 15 minutes.

A solution of Mercurochrome shall rise to a height of not less than the height shown in the table below when the narrowest end of the point is inserted to a depth of 1 mm in an aqueous solution of 2% Mercurochrome.

Point Size	Minimum Stain Height in mm
XXF	2.5
XF	3.5
F	4.0
M	5.0
C	6.0
XC	6.0

The point shall remain in the solution for 10 seconds before it is removed, and the height of Mercurochrome absorption measured. A minimum of 5 points from each lot or batch number shall be subjected to this procedure and the results averaged.

Instructions. Each unit of issue shall include adequate and accurate instructions for opening the immediate container while maintaining the maximum possible sterility.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Sterility. The points shall be sterile when tested in accordance with current USP or AAMI guidelines.

Workmanship. Sterile points shall be free of disagreeable odor and any occluded particles. Points shall be free from defects which may detract from their appearance or impair their serviceability.

Unit. Package. (PG). For size 1, one package containing a minimum of 200 sterile points as specified, constitutes one unit. For size 2, one package containing a minimum of 2400 sterile points, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one package of 200 sterile points. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one package of 200 sterile points. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6520-01-112-4054	POINT, ENDODONTIC, PAPER, MEDIUM, STERILE, 2400's
6520-01-112-4055	POINT, ENDODONTIC, PAPER, COARSE, STERILE, 2400's
6520-01-112-4056	POINT, ENDODONTIC, PAPER, EXTRA COARSE, STERILE, 2400's
6520-01-112-4057	POINT, ENDODONTIC, PAPER, EXTRA EXTRA FINE, STERILE, 2400's

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:33Z
Check the source to verify that this is the current version before use.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6520-01-114-1535	POINT, ENDODONTIC, PAPER, EXTRA FINE, STERILE, 2400's
6520-01-114-1536	POINT, ENDODONTIC, PAPER, FINE, STERILE, 2400's
6520-00-904-1977	POINT, ENDODONTIC, PAPER, EXTRA FINE, STERILE, 200's
6520-00-904-1978	POINT, ENDODONTIC, PAPER, FINE, STERILE, 200's
6520-00-904-1979	POINT, ENDODONTIC, PAPER, MEDIUM, STERILE, 200's
6520-00-904-1980	POINT, ENDODONTIC, PAPER, COARSE, STERILE, 200's
6520-00-904-1981	POINT, ENDODONTIC, PAPER, EXTRA COARSE, STERILE, 200's
6520-00-926-9124	POINT, ENDODONTIC, PAPER, EXTRA EXTRA FINE, STERILE, 200's

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS
GSA-FSS-FCGC

Project No. 6520-2402
Location: paul/point/dawne

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.