A-A-54385
22 April 1991

COMMERCIAL ITEM DESCRIPTION

CRIMPER, MESH, Maxillofacial Implant

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a titanium mesh implant crimper used in maxillofacial surgical implant procedures.

Salient characteristics:

General. Shall be a titanium mesh implant crimper designed for use in modifying preformed titanium mesh implants and custom making implants from flat sheet titanium mesh in oral and maxillofacial surgical implant procedures using autogenous marrow and cancellous bone for discontinuity defects of the mandible.

Material. The crimper tips shall be fabricated from corrosion resistant type 440 steel in accordance with ASTM Designation F899 covered with a titanium nitride coating. The rest of the crimper shall be fabricated from class 4 corrosion resistant steel in accordance with ASTM Designation F899.

Style and design. The crimper shall be designed for use with the maxillofacial mesh implants (see notes). The style and design shall be in accordance with (DPSC) Defense Personnel Support Drawing No. 24565. The edges shall be properly beveled to prevent pinching of tissue. The crimper shall operate smoothly. The crimper shall be free from excessive stiffness and shall not have excessive play or indication of looseness when the crimper is in the open or closed position. The handle shall provide a secure grip. The tips shall be permanently secured to the crimper. The crimper shall be packaged in a container designed to prevent damage during shipment or storage.

Passivation. The crimper shall be properly passivated. As an alternate, if the entire crimper is coated with titanium nitride, passivation is not required.

Sterilization. The crimper shall be capable of withstanding repeated steam autoclaving and ethylene oxide sterilization without showing any signs of corrosion.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Identification marking. The crimper shall be marked with the name or registered trademark of the contractor. If the crimper is of foreign manufacture, the instrument shall be marked with the country of origin.

Finish. The crimper except for the knurled portion of the handle shall have a smooth finish.

Workmanship. The crimper shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One mesh implant crimper, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one crimper.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified below. The unit of product for test purposes shall be one crimper. Sampling shall be level S-1, AQL 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sterilization. The crimper shall be autoclaved at 275°F for twenty minutes. Remove the crimper from the sterilizer and allow it to cool to room temperature. Repeat this procedure 3 times.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6515-01-163-8080	CRIMPER, MESH, Maxillofacial Implant

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

NOTES. The following is a list of the maxillofacial implant meshes and their corresponding NSN's.

Left mandibular body	-	6515-01-163-3762
Left mandibular angle	-	6515-01-163-3763
Flat sheet	-	6515-01-163-7935
Right mandibular angle	-	6515-01-163-8078
Mandibular symphysis	-	6515-01-163-8079
Right mandibular body	-	6515-01-164-2678

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4231
Location: ML#1/Crimper

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54385	2. DOCUMENT DATE (YYMMDD): 4-22-91

3. DOCUMENT TITLE: CRIMPER, MESH, Maxillofacial Implant

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	e. DATE SUBMITTED (YYMMDD)

7. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code): ATTN: DPSC-MSTE, 2800 S. 20th Street, Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office, 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466, Telephone (703) 756-2340, AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/196

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.