A-A-54387
22 April 1991

COMMERCIAL ITEM DESCRIPTION

MIRROR, MOUTH EXAMINING, (Plastic Handle, Disposable)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a disposable mouth examining mirror suitable for patient use in preventive dentistry programs.

Salient characteristics:

The disposable mirror shall consist of a continuous reflective glass or acrylic disc securely encased in a one-piece molded plastic frame and handle. There shall be no movement of mirror in frame. To provide for disc security, the top edge of the plastic frame shall be bent over the disc or, as an alternate, the frame shall be tapered inwardly to provide a force fit. The mirror face shall have a smooth reflective surface free from pits, scratches, cracks, ripples or other surface imperfections affecting reflectivity. The mirror frame and handle shall have a smooth finish free from sharp edges or projections.

The mirror glass shall be of good quality, polished glass or crystal clear acrylic. The composition used shall be capable of withstanding five (5) repeated soap and water washings without dulling the surface of the mirror. The back of the glass shall be coated with chemically deposited silver, protected by a film of electrically deposited copper. The copper surface shall be provided with a heavy coating of water-resistant enamel. The back of the acrylic disc shall be coated with pure aluminum (at least 99.9%). The glass, silver, copper and enamel interfaces and the acrylic and aluminum interface shall be firmly adherent. The acrylic and aluminum interface shall be coated with a water-resistant non-toxic lead free acrylic meeting ASTM D3359 for adhesion. The front and back surfaces of the glass and acrylic shall be flat and parallel. The front surface of the acrylic mirror shall have a poly-film protectant with a tab for easy removal. The reflective surface shall provide a clear undistorted image so that a patient can interpret results of flossing and brushing stained tooth areas under normal unintensified illumination.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Plastic used for the mirror frame and handle shall be acrylonitrile-butadiene-stryrene (ABS) cellulose acetate, or polystyrne. The plastics shall have a minimum average Izod impact strength at 23° Centigrade of 1.5 foot-pounds per inch of notch when tested in accordance with ASTM D256 and a minimum average tensile of 4000 pounds per square inch when tested in accordance with ASTM D638.

The overall length of the mirror shall be 6±1/2 inch. The plane of the mirror head (face) shall be 30 degrees ±10 degrees to the major plane of the handle. The cross section of the handle shall be suitable for finger manipulation. The cross section of the neck shall be sufficient to maintain mirror integrity. The proximal end of the handle shall have a suitable hole or slot or both. The mirror glass and coatings shall be a minimum of 0.050 inches to a maximum of 0.075 inches in thickness. The mirror head shall have a thickness of 0.10 to 0.15 inch. The diameter of the exposed mirror face shall be 5/8 to 1 inch.

Workmanship. The mirror shall be free from defects which may affect proper use, appearance, or serviceability.

Unit. Package (PG). One package containing seventy-two mirrors, as specified constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, characteristics. The unit of product for inspection purposes shall be one mirror.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-00-935-4135	MIRROR, MOUTH EXAMINING, PLASTIC HANDLE, DISPOSABLE, 72s

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2472
LOCATION: PB#2/Mirror

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54387	2. DOCUMENT DATE (YYMMDD): 4-22-91

3. DOCUMENT TITLE: MIRROR, MOUTH EXAMINING, (Plastic, Plane Glass, Disposable)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) / 7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSTE, 2800 S. 20th Street, Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office, 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466, Telephone (703) 756-2340, AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.