COMMERCIAL ITEM DESCRIPTION

CURETTES, PERIODONTAL (McCall)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-C-36507C which is cancelled.

This commercial item description covers double-ended periodontal dental curettes of McCall design.

Salient characteristics:

The double-ended periodontal dental curettes of McCall design shall be as specified below:

McCall No. 13s and 14s
McCall No. 17s and 18s

Material.

Operative ends. The operative ends shall be Class 4, Type 440A, 440B, 440C stainless steel.

Handles. Handle A shall be fabricated of Class 4, Type 416 martensitic stainless steel or Class 6, Type 430F ferritic stainless steel as specified in ASTM Designation F899 for composition only. Handle C shall be fabricated of Class 3, Type 303 austenitic stainless steel as specified in ASTM Designation F899 for composition only. Handle B and D shall be chromium over nickel plated brass. Brass shall conform to copper alloy No. 260 or 330 as specified in ASTM Designation B135 for composition only.

Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C52 to C56 for Type 440A, C54 to C58 for Type 440B, and C56 to C60 for Type 440C in accordance with ASTM Designation E18 (Rockwell) or E92 (Vickers) and E140. Handles and shanks shall be properly tempered and shall have a hardness suitable for the purpose intended.

Design and dimensions. The curettes shall be double-ended McCall design curettes. The curettes shall be furnished as specified herein. The curettes shall be suitable for use in

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

subgingival curettage during periodontal treatment. The design and dimensions of the curettes and handles shall conform to Defense Personnel Support Center (DPSC) Drawing 24034 for the applicable design and handles designations (A, B, C, or D).

Construction. The curettes shall be of double-ended construction consisting of a handle and an operative end at each end. The operative ends shall be permanently secured to the handle. The blades of the operative ends of each curette shall face in opposite directions. Handles A and C shall be of solid construction. Handles B and D shall be of hollow construction. Junction of the operative end to the handle shall be smoothly finished and with no crevices evident. There shall be no movement, loosening, or separation between the operative end and the handle. The handle and shank of the operative end shall be securely held in separate chucks or other holding devices. A slowly increasing pull out force, increasing from zero to 85 pounds (minimum), shall be applied to the shank of the operative end. The force shall be maintained at the high value for a minimum of one second. The instrument shall be examined for movement, looseness, or separation of the operative end with respect to the handle. The test shall be repeated for each operative end of the double ended instrument.

Finish. The operative end including shank shall have a smooth mirror or smooth bright finish, except for the surfaces that form the cutting edges which shall have a finely honed finish. The operative end (entire blade) shall have a sharp cutting edge and shall be free of defects such as rough grind marks, cracks, gouges, burrs, feathers, or other imperfections when viewed under four power (4X) magnification. Handles shall have a bright finish, except for the serrated or knurled portion. Finger grip or serrated portion of handles shall be free from barbs and shall provide adequate finger grip during use.

Corrosion resistance. Corrosion resistance tests in shall be in accordance with ASTM Designation F1089. The copper sulfate test shall apply to the stainless steel operative ends and handles. The boil test shall apply to the entire curette (operative ends and handles).

Identification marking. Each curette shall be marked with the identification "MC 13s-14s" or "MC 17s-18s" for the applicable National Stock Number on the smooth portion of the handle. Each curette shall be marked with the country of origin, the name or registered trademark of the manufacturer (and

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

contractor, if other than the manufacturer) and the word "stainless" or the use of a registered trade name identifying the steel as stainless, provided prior approval is received from the cognizant contracting officer. All markings shall be permanent and legible.

Individual Package. Each curette shall be wrapped in neutral tissue or other suitable anticorrosive material and enclosed in a close fitting box to prevent damage to the operative ends. As an alternate, the individual package shall be a rigid plastic tube having a metal screw cap or a plastic friction plug type closure. The curette shall be cushioned within the tube in such a manner to prevent damage to the operative ends.

Workmanship. The curettes shall be free from all defects such as pits, burrs, feathers or cracks which detract from their appearance or impair their serviceability, proper function or intended use.

Unit. Each (EA). One periodontal curette of the McCall design No., as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

(determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, examination, and test characteristics. The unit of product for inspection, examination, test purposes shall be one curette. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, except for dimensions which shall be conducted at inspection level S-1, AQL 2.5 (percent defective) and sampling for tests (when applicable) shall be conducted at inspection level S-1. For tests the acceptance number shall be zero (0) for all sample sizes.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Chemical analysis. In the absence of contractor certification for the raw material, a chemical analysis shall be conducted. The chemical analysis shall be performed by reagent solutions, combustion methods, nonemission methods, or by spectrochemical analysis. Any ASTM method shall be acceptable.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statues, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulations Subpart 23.4 to the maximum extent practical .

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-00-890-2050	CURETTE, PERIODONTAL, McCall, No. 13s and 14s
6520-00-890-2051	CURETTE, PERIODONTAL, McCall, No. 17s and 18s

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2417
LOCATION: KAD2/CURETTES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54396	2. DOCUMENT DATE (YYMMDD): 4-30-91

3. DOCUMENT TITLE: CURETTES, PERIODONTAL (McCall)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) / 7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial / (2) AUTOVON
Defense Personnel Support Center	(215)737-2870 / 444-2870
c. ADDRESS (Include Zip Code): ATTN: DPSC-MSTE, 2800 S. 20th Street, Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office, 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466, Telephone (703) 756-2340, AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.