MIL-C-36961A
11 June 1975
SUPERSEDING
MIL-C-36961
11 July 1974

MILITARY SPECIFICATION

CASE, MEDICAL INSTRUMENT AND SUPPLY SET PLASTIC, RIGID

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers two sizes of plastic cases that are suitable for the protection of medical supplies and equipment in the field, in vehicles, small water craft, liferafts and aircraft.

1.2 Classification. Plastic cases covered by this specification shall be of the following sizes:

Size A	Nominal inside dimensions: 7-1/2 inches Long by 4-1/2 inches Wide by 2-3/4 inches High.
Size B	Nominal inside dimensions: 9-1/2 inches Long by 9 inches Wide by 7 inches High.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

FSC 6545

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:39Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

L-P-375	Plastic, Film, Flexible, Vinyl Chloride.
L-P-1183	Plastic Molding Material, Acrylonitrile-Butadiene-Styrene (ABS), Rigid.
QQ-A-225/8	Aluminum Alloy Bar, Rod, Wire and Special Shapes, Rolled, Drawn, or Cold Finished, 6061.
QQ-A-250/8	Aluminum Alloy, 5052, Plate and Sheet.
QQ-W-321	Wire, Copper Alloy.
QQ-W-423	Wire, Steel, corrosion-resisting.
TT-E-527	Enamel, Alkyd, Lusterless.
TT-P-1757	Primer Coating, Zinc Chromate, Low Moisture Sensitivity.
PPP-B-636	Boxes, Shipping, Fiberboard.
PPP-T-76	Tape, Pressure-sensitive Adhesive Paper, Water Resistant, (For Carton Sealing).

Military

MIL-C-5541	Chemical Films and Chemical Film Materials for Aluminum and Aluminum Alloys.
MIL-R-6130	Rubber, Cellular, Chemically Blown.
MIL-C-8514	Coating Compound, Metal Pretreatment, Resin-acid.
MIL-A-8625	Anodic Coatings, for Aluminum and Aluminum Alloys.
MIL-M-20693	Molding Plastic, Polyamide (Nylon) Rigid.
MIL-P-55010	Plastic Sheet, Polyethylene Terephthalate.

STANDARDS

Federal

FED-STD-595	Color.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:39Z
Check the source to verify that this is the current version before use.

DRAWINGS

Defense Personnel Support Center (DPSC).

24125	CASE, MEDICAL INSTRUMENT AND SUPPLY SET, Plastic, Rigid, Size A.
24163	CASE, MEDICAL INSTRUMENT AND SUPPLY SET, Plastic, Rigid, Size B.

(Copies of specifications, standards and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Component parts.

3.1.1 Size A case. The size A case shall consist of a body, and a cover with two hinge pins, two molded retaining clips, and a gasket.

3.1.2 Size B case. The Size B case shall consist of a body, cover, six drawbolts, two of which secure the cover to the body and serve as hinges, a gasket that is retained in the cover, and a carrying handle.

3.2 Materials. Except as specified herein, materials shall conform to the applicable portions of the specifications, standards, and drawings referenced herein. When materials to be used are not specifically designated, those used shall be of the highest commercial quality, and suitable for the purpose intended.

3.2.1 Aluminum rod. The aluminum alloy rod for the handle of the Size B case shall be alloy designation 6061, temper T 4, of QQ-A-225/8.

3.2.2 Aluminum sheet. The aluminum sheet for the drawbolt draw of the Size B case shall be alloy designation 5052, temper H 24, or H 34, of QQ-A-250/8.

3.2.3 Brass wire. The brass wire for the hinge pins of the Size A case shall be 1/8 hard, 70-30 brass wire of QQ-W-321.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:39Z
Check the source to verify that this is the current version before use.

3.2.4 Corrosion resisting steel wire.

3.2.4.1 Wire for hinge pin. Wire for the drawbolt hinge pin of the Size B case shall be Form I, Composition 316, Condition B, of QQ-W-423. As an alternate, a corrosion resisting steel tubular rivet may be used as a hinge pin.

3.2.4.2 Wire for the draw ring. Wire for the drawbolt draw ring of the Size B case shall be Form I, Composition 302, Condition B, of QQ-W-423.

3.2.5 Plastic.

3.2.5.1 Plastic for the cases. Plastic for all plastic components, except the clips of the Size A case, shall be an acrylonitrile-butadiene-styrene material conforming to L-P-1183, Type V.

3.2.5.2 Plastic for the clips. Plastic for the retaining clips of the Size A case shall be Type IVA of MIL-M-20693.

3.2.5.3 Plastic for alternate Red Cross. The plastic for the alternate Red Cross (see 3.8.1) shall consist of a sheet of flexible vinyl chloride sandwiched between two (2) sheets of polyethylene terephthalate.

a. The flexible vinyl chloride sheet (film) shall conform to the requirements of L-P-375, Type II, Class 2, 0.020 inches thick. Color shall be white (see 3.6).

b. The polyethylene terephthalate sheet shall conform to the requirements of MIL-P-55010. Thickness shall be 0.002 inches; commercial tolerances shall apply.

3.2.6 Adhesive. The adhesive used to bond the alternate Red Cross (see 3.8.1) to the cover of the case shall be Tadco Mfg. Co., Inc's Tadco No. 89, or equal.

3.2.7 Rubber, cellular, chemically blown. Rubber for the gaskets of both cases shall conform to Type II, Grade A, Condition Medium of MIL-R-6130 except for sheet stock thickness tolerances. Rubber for the Size A case shall be closed cell neoprene cord. Rubber for the Size B case shall be cut from nominal 1/4 inch thick sheet stock. The tolerance on the 1/4 inch sheet stock shall be plus (+) 1/16 inch and minus (-) 1/32 inch.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:39Z
Check the source to verify that this is the current version before use.

3.2.8 Primer. The primer used on those metal components of the Size B case requiring a prime coat shall conform to TT-P-1757, and shall be color Y of the same specification.

3.2.9 Enamel. The enamel used on all exposed metal components of the Size B case shall conform to TT-E-527.

3.3 Construction.

3.3.1 Size A case. The body and the cover shall each be a molded one-piece construction in accordance with Drawing No. 24125. The retaining clips and hinge pins shall be permanently secured. The wall thickness of the cover and bottom shall be within specified tolerances, and shall be symmetrical. The finished case shall show no evidence of cracking when tested as specified in 4.3.1. The gasket shall render the case watertight when the case is tested as specified in 4.3.2.1.

3.3.2 Size B case. The body and cover shall each be of a molded one piece construction in accordance with Drawing No. 24163. Drawbolt supports and cover strikes shall be integrally molded. The wall thickness of the cover and bottom shall be within specified tolerances, and shall be symmetrical. The finished case shall show no evidence of cracking when tested as specified in 4.3.1. The gasket shall render the case watertight when the case is tested as specified in 4.3.2.2.

3.4 Gasket.

3.4.1 Size A case. The gasket shall be fabricated of the material specified in 3.2.7. The gasket shall be either one-piece endless construction or the ends shall be cut on a nominal 45 degree angle and securely cemented together. The cemented joint shall not separate when the gasket is tested as specified in 4.3.3.

3.4.2 Size B case. The gasket shall be fabricated of the material specified in 3.2.7. The gasket shall be of one piece construction and shall conform to the contour of the gasket recess as shown in the drawing. The surface of the gasket against which the body bears shall be suitably treated to prevent adhesion. The gasket shall remain in place and be suitable for sealing the case during normal use.

3.5 Handle. The handle of the Size B case shall rotate freely in its bracket.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:39Z
Check the source to verify that this is the current version before use.

3.6 Color. The color of the finished case, and the color of the case markings shall be uniform and shall approximate the specified colors of Fed. Std No. 595 indicated below, see 6.3 and 6.4:

Component or Marking	Color No.	Nominal color
All molded plastic components (3.7.1)	34087	Olive drab
All exposed surfaces of metal components (3.7.2)	34087	Olive drab
Background for Geneva Red Cross: (3.8.1)	17886	White
Geneva Red Cross (3.8.1)	11105	Red
Identification marking - exterior of case cover (3.8.1)	17038	Black
Identification marking - exterior of case bottom (3.8.2)	34087	Olive drab

3.7 Finish.

3.7.1 Plastic. The clips, handle bracket, and interior and exterior of the cases shall have a lusterless finish.

3.7.2 Metal. All observable metal surfaces shall be finished with the enamel specified in 3.2.9. Surfaces to be enameled shall be thoroughly cleaned and degreased, then prepared for top-coating by one of the following methods:

For aluminum alloys:

(1) Chemical treatment in accordance with MIL-C-5541.

(2) Anodic treatment in accordance with MIL-A-8625.

(3) Priming treatment, using material specified in 3.2.8, thinned and applied by spray in accordance with the directions specified under "Preparation for Delivery" in TT-P-1757, and air-dried for not less than 30 minutes before top-coating.

For corrosion resisting steel alloys:

(1) Pretreatment in accordance with MIL-C-8514.

3.8 Identification marking.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:39Z
Check the source to verify that this is the current version before use.

3.8.1 Cover marking. The top exterior surface of the case cover shall have molded on to it the National Stock Number of the First Aid Kit, the kit item identification, and a square on which will be mounted a Geneva Red Cross on a white background, as indicated on the Drawing. Height of characters above surface shall be uniform and shall be 0.035 ± .005 inches. Color of characters shall be as specified in 3.6. As a partial alternate, the molded square shall be eliminated and replaced with a 0.024 ± .002 inch thick 3-ply plastic laminate depicting a Geneva Red Cross printed on a white background (see 3.2.5.3). The laminate shall be permanently adhered to the top surface of the cover with the waterproof adhesive specified in 3.2.6.

3.8.2 Body marking. The information indicated below shall be permanently and legibly embossed on the exterior bottom surface of the case body. Minimum letter size shall be 1/8 inch. Color of letters and numerals shall be the color of the case body:

6545-00-113-3722 or 6545-00-131-6992, as applicable.

CASE, MEDICAL INSTRUMENT AND SUPPLY SET,
Plastic, Size A, or Size B, as applicable.

The name or registered trademark of the manufacturer.

The contract number.

3.9 Workmanship. Workmanship shall be first class throughout. Completed cases shall be well made and free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:39Z
Check the source to verify that this is the current version before use.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the plastics, metals, and rubber, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables I and II (see 4.2.1 and 4.2.2) and 4.2.3.

4.2.1 Sampling for examinnation. Sampling for *examination* shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in table I. Unit of product for examination purposes shall be one complete case.

Table I. Sampling for Examination

Examination	Inspection level	AQL (percent defective)
For visual *examination*		
Major defects	S-2	1.0
Minor defects	S-2	6.5
For dimensional examination	S-2	6.5

4.2.2 Examination.

4.2.2.1 Case. The cases shall be examined for defects including, but not limited to, those listed in table II.

Table II. Classification of Defects

Major	
101	Any component missing.
102	Case not completely molded.
103	Cover does not fit body.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:39Z
Check the source to verify that this is the currentversion before use.

Table II. Classification of Defects (cont'd)

104	Drawbolts/Latches do not engage.
105	Hinge pins of Size A case do not engage distal hinge support.
106	Drawring/Latch not secured.
107	Drawbolt hinge pins or drawbolt rivets of Size B case not secured.
108	Color of case, excluding markings on cover, does not approximate 34087.
109	Marking incorrect, incomplete, illegible, or missing.
110	Red Cross does not register on white background within the tolerance specified.
111	Handle of Size B case not located at balance point.
112	Cemented gasket joint in Size A case not aligned.
Minor	
201	Case not free of flash.
202	Finish not as specified.
203	Exterior surface not free of blemishes.
204	Colors of cover markings not colors specified.
205	Handle of Size B case does not rotate freely in its bracket.

4.2.2.2 Preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table III. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 (defects per hundred units).

Table III. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:39Z
Check the source to verify that this is the current version before use.

4.2.3 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, Inspection Level S-1, AQL 4.0, except that the acceptance number shall be zero for all sample sizes. Unit of product for test purposes shall be one case.

4.3 Tests. Tests shall be conducted as necessary to determine compliance with specification requirements.

4.3.1 Drop tests. The weight indicated below for the size case under test shall be evenly distributed over the bottom of the case, and the case cover shall be secured completely. The case shall be conditioned at 70°F ± 5° for a minimum of 3 hours, and then dropped once on a bottom corner onto a concrete floor from a height of 36 inches. The case shall be examined, both inside and outside, for evidence of cracking.

Test weights shall be as follows:

Size A Case	5 pounds of lead shot.
Size B Case	15 pounds of lead shot or sand.

4.3.2 Leakage tests.

4.3.2.1 Simulated rain precipitation test, Size A case. For the simulated rain precipitation test, set a shower head 6 feet ± 6 inches above the tops of the cases to be tested, provide as a support for sample cases a horizontal platform directly below the shower head, and adjust the rate of flow to deliver a minimum of 30 gallons of tap water per hour. The diameter of the shower spray hitting the platform shall not be greater than 30 inches. Position the bottoms of sample cases on the platform in such a manner that there is a minimum distance of 2 inches between cases on all sides. All cases under test at any time shall be entirely within the perimeter of the shower streams. Sample cases shall be subjected to a continuous water spray under the aforementioned conditions for a period of 15 minutes. The exterior surfaces of the cases shall then be carefully blotted dry, the cases opened, and the interiors examined for water.

4.3.2.2 Watertightness, Size B case. Sufficient weight shall be placed in an empty case to ensure that the closed case will remain submerged in water. The case shall then be closed at room temperature and the cover secured by all six drawbolts. The closed case shall be placed with its cover horizontal and uppermost in a container filled with tap water to a level 2 feet above the top of the case. The water level shall be maintained for 15 minutes. Thereafter, the case shall be removed from the water, the outside carefully blotted dry, and the case then opened to determine whether any water has entered.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:39Z
Check the source to verify that this is the current version before use.

4.3.3 Gasket joint adhesion test, Size A case. Remove the gasket from the case and suspend it vertically with the cemented joint at approximately the midpoint of one of the two parallel sides. Attach a one (1) pound weight to the lower extremity of the loop. The weight shall remain on the gasket a minimum of five (5) seconds.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Each case shall be closed and the lid secured. Each case shall be wrapped in kraft paper and the wrap secured. As an alternate, the case may be inserted into a snug fitting paperboard sleeve to prevent marring or scratching of the case.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Forty eight Size A cases or four Size B cases, packaged as specified in 5.1.1.1, shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, type CF, class weather-resistant. Closure, strapping, and waterproofing shall be as specified in the appendix of the box specification. Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution. Size B cases shall be separated within the exterior container with fiberboard partitions or pads.

5.2.2 Level B. Forty eight Size A cases or four Size B cases, packaged as specified in 5.1.1.1, shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, type CF, class domestic. Closure shall conform to method II as specified in PPP-B-636. Size B cases shall be separated within the exterior container with fiberboard partitions or pads.

5.2.3. Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking. Interior packages and exterior containers shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in c rner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the *referenced document(s)* or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-C-36961A CASE, MEDICAL INSTRUMENT AND SUPPLY SET PLATIC, RIGID

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:39Z
Check the source to verify that this is the current version before use.