MILITARY SPECIFICATION

FORCEPS, NASAL SEPTUM STRAIGHTENING, ASCH, 8-1/2 INCH

This Amendment forms a part of Military Specification MIL-F-36965, dated 2 May 1973, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 4:

4.2.2 Sampling for tests. Delete Table II and notes under Table II, and substitute:

"Table II. Sampling for Tests

Component (unit of product)	Characteristic	Requirement	Test procedure	Inspection level	AQL
Forceps	Hardness test	3.2 and GG-I-526	GG-I-526	S-2	1.5
	Dimensions	3.4	———	S-2	2.5
	Boil test	GG-I-526	GG-I-526	S-2	1.0
	Copper sulfate test	GG-I-526	GG-I-526	S-2	1.0
	Distortion test	GG-I-526	GG-I-526	S-2	2.5
	Lock stiffness test	3.6	GG-I-526	S-2	2.5

."

Add following paragraph:

"4.2.3 Raw materials. A sufficient quantity of each lot of metal used in the construction of the forceps shall be selected for chemical analysis."

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:40Z
Check the source to verify that this is the current version before use.

Page 6:

5.1.3.2 Preservative. Delete "Preservative shall not be applied" and substitute: "No preservative shall be applied."

6.3 Delete "Federal Stock No. 6515-336-5800", and substitute: "National Stock No. 6515-00-336-5800".

6.4 Delete entirely.

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1868

Review activities:

Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:40Z
Check the source to verify that this is the current version before use.