MIL-F-36984
Amendment-1
9 October 1974

MILITARY SPECIFICATION

FORCEPS, BULLET EXTRACTING, SENN, 8 INCH

This Amendment forms a part of Military Specification MIL-F-36984, dated 28 June 1973, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 2:

3.5 Style, design and dimensions. Delete period at end of paragraph, and substitute: ",except that the finger rings shall have a width of 3/32 ± 1/64 inch and thickness of 1/8 ± 1/64 inch in cross section."

Page 5:

4.3 Sampling for tests. Delete Table II entirely, including notes under table, and substitute:

"Table II. Snapling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Forceps	Hardness	3.3	GG-I-526	S-2	1.5
	Dimensions	3.5	----	S-2	2.5
	Copper sulfate test	GG-I-526	GG-I-526	S-2	1.0
	Boil test	GG-I-526	GG-I-526	S-2	1.0

"

Add the following paragraph:

"4.3.1 Raw materials. A sufficient quantity of each lot of metal used in the construction of the forceps shall be selected for chemical analysis."

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:42Z
Check the source to verify that this is the current version before use.

Page 6:

In Table III, for Major category 101, delete "or" and substitute "and". Delete description for Major category 102, and substitute: "Finish not uniform, and not of finish specified."

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1876

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:42Z
Check the source to verify that this is the current version before use.