MIL-I-36969A
6 April 1976
SUPERSEDING
MIL-I-36969 (DSA-DM)
31 May 1973

MILITARY SPECIFICATION

IRRIGATION KIT, PATIENT

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a sterile, disposable, prepackaged irrigation kit.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-636	-	Boxes, Shipping, Fiberboard.

MILITARY

MIL-S-36586	-	Sterilization Test Strip Set, Bacterial Spore.
MIL-S-36876	-	Syringe, Irrigating, Surgical, Catheter Tip, Disposable, 50 CC (2 Fl. oz.).

STANDARDS

MILITARY

MIL-STD-105	-	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	-	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:40Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

U. S. PHARMACOPEIAL CONVENTION, INC. (U.S.P.)

Pharmacopeia of the United States

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

3. REQUIREMENTS

* 3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

* 3.2 Materials. Materials used in the fabrication of the components of the irrigation kit shall be of the highest quality and suitable for the intended purpose. All material shall comply with the applicable requirements of the Federal Food, Drug and Cosmetic Act.

* 3.3 Design and construction. The irrigation kit shall consist of a tray, a hospital overwrap, a towel, a syringe, and a solution container. The components shall be assembled in the tray and the tray shall be wrapped in the hospital overwrap in a manner which will facilitate optimum aseptic technique. The irrigation kit shall be sterile.

* 3.3.1 Tray. The tray shall be fabricated of polyethylene, polypropylene or polystryrene and shall have a minimum capacity of 1200 cubic centimeters (cc). The tray shall be graduated at each 100 cc interval and shall be numbered at a minimum of 200, 400, 600, 800, and 1000 cc.

* 3.3.2 Overwrap. The hospital overwrap shall be waterproof, white in color, and shall be a minimum of 17 by 18 inches. The overwrap shall consist of a minimum of 4 plies, one ply continuous plastic and 3 plies of facial-quality absorbent tissue or similar nonwoven absorbent towel. The overwrap shall be suitable for providing a sterile field for assembling components or as an underpad for the patient.

* 3.3.3 Cleaning towel. The cleaning towel shall be a non-woven absorbent fabric not less than 5 by 7 inches in size. The cleaning towel shall be saturated with povidone-iodine antiseptic solution. The towel shall be folded flat and sealed in a plastic-foil laminate packet.

3.3.4 Syringe. The syring shall be of a three piece design consisting of a bulb, a syringe, and a tip cap. The syringe shall be provided assembled ready for use.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:40Z
Check the source to verify that this is the current version before use.

3.3.4.1 Bulb. The bulb shall be fabricated of opaque flexible plastic or synthetic rubber. The bulb shall be spherical and shall have a neck for insertion into the syringe barrel. The neck shall provide an air tight seal with the syringe barrel during normal use in irrigating and aspirating. A plastic insert in the neck shall be acceptable if provided. The bulb shall be such a size as to permit complete filling of the syringe to at least the 60 cc level with one squeeze. The force required to depress the bulb one inch shall not exceed 8 pounds when tested as specified in 4.4.3. The bulb shall be sufficiently rigid to provide a strong suction force when the syringe is used for aspirating.

* 3.3.4.2 Syringe barrel. The syringe barrel shall meet the requirements of MIL-S-36876 for material, design and construction.

3.3.4.3 Tip cap. The tip cap shall be fabricated of polypropylene or polyethylene. The cap shall serve to protect the catheter tip from damage and to keep it clean.

* 3.3.5 Solution container. The container shall be fabricated of polypropylene, polyethylene or polystyrene. Container shall be provided with a snap-on-cap, or telescoping cover. The telescoping cover shall have a hole to accommodate the syringe barrel. The container shall have a minimum capacity of 500 cc and shall be graduated and numbered at a minimum of each 50 cc interval.

3.4 Sterility. The irrigation kit shall be sterile when tested with a biological indicator as specified.

3.4.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National stock number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and date of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:40Z
Check the source to verify that this is the current version before use.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-AIQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA, 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.4.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.2. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.5 Instructions for use. Instructions for use shall be included with each kit. The instructions shall be either printed on the package or provided on a separate sheet.

3.6 Workmanship. Workmanship shall be first class throughout. The plastic tray, solution container and syringe units shall be free of burrs, cuts, and cracks, or other defects which detract from their appearance or may impair their serviceability. The overwrap shall be free of holes and tears. The bottoms of the plastic tray and container shall be so formed that the tray and container are stable and do not rock when placed on a flat surface. The components shall be free from dirt, dust, oil, and other foreign matter.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:40Z
Check the source to verify that this is the currentversion before use.

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the irrigation kits when a first article sample is required (see 3.1). This inspection shall include examination of 4.3 and tests of 4.4.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.3 and tests of 4.4.

4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the components or materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for examination shall be the kit.

TABLE I. Sampling for examination

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major	II	1.0
Minor	I	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for sterility tests shall be as indicated in 4.4.1. The acceptance number shall be zero for sterility tests. Sampling for 4.4.2 and 4.4.3 shall be S-4 of MIL-STD-105, AQL shall be 1.0. The unit of product for tests shall be the individual kit.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:40Z
Check the source to verify that this is the currentversion before use.

4.3 Examination. The kits shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects, as shown in table II. Examination shall not be restricted to the classified possible defects listed below.

TABLE II. Classification of defects

Categories	Defect
Major	
101	Kit not complete with all components.
102	Components not free from damage.
103	Trays or containers not stable on flat surface.
104	Graduations not present or not as specified.
105	Materials not as specified.
106	Components not free from foreign matter.
Minor	
201	Towel packet not free of open seams.
202	Components not free of scratches, dents, smudges, or flash.
203	Components not free of burrs and sharp edges.
204	Instructions for use not present.

4.3.2 Dimensional examination. The irrigation kit shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

* 4.3.3 Examination of packaging, packing and marking. An examination shall be made to determine compliance with the packaging, packing and marking requirements of section 5. Defects shall be scored as specified in table III. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 with AQL of zero defects per hundred units.

TABLE III. Examination of packaging, packing and marking

Examine	Defect
Contents	not as specified.
Individual container	not capable of maintaining sterility of contents.
Container	not as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:40Z
Check the source to verify that this is the current version before use.

TABLE III. Examination of packaging, packing and marking - continued

Examine	Defect
Marking	omitted; incorrect; illegible; improper size, location, sequence or method of application.
Material	component missing or damaged.
Plastic bag	not heat sealed or peel type.
Workmanship	bulging or distoration of containers.

4.3.3.1 Examination of closure, waterproofing and banding of container. When the shipping container is required to comply with PPP-B-636, the examination for defects in closure, waterproofing or banding shall be in accordance with the appendix of the box specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Sterility test. The irrigation kits shall be tested for sterility in accordance with the method specified in the U.S.P. for solids using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days of inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.1.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for type I sterilization and 30 for type II sterilization including inoculated carrier and product test units.

4.4.1.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam (Type I)	B. stearothermophilus	55° - 65°C
Dry Heat (Type II)	B. stearothermophilus, or	55° - 65°C
	B. subtilis var. niger (or globigii)	30° - 35°C
Gas (Type II)	B. subtilis var. niger (or globigii)	30° - 35°C
Radiation (Type II)	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:40Z
Check the source to verify that this is the current version before use.

4.4.1.2 Inoculated carrier. Inoculated carriers (paper spore strips), if used, shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.1 and 4.4.1.1. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.1.1.
NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the U.S.P.

4.4.1.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility shall be made to the Contracting Officer and approval obtained before employment of the alternate method on a Government contract.

4.4.1.4 Product testing. Testing for the sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the U.S.P.

4.4.2 Syringe test. Water shall be drawn into the syringe by submerging the tip and squeezing and releasing the bulb once. The water level in the syringe barrel shall be observed to assure that the level is at least at the 60 cc graduation. The syringe shall then be emptied. With the bulb compressed the tip shall be sealed to prevent the entrance of air. After 15 minutes the syringe bulb shall be examined for evidence of air leakage between the bulb neck and the syringe. The syringe shall again be filled with water and inverted to permit the water to enter the bulb. When the water is emptied from the syringe it shall be visually observed for any foreign matter.

4.4.3 Bulb depression test. The bulb shall be secured in a suitable test apparatus. Force shall be applied normal to the rear of the bulb until the bulb has been depressed one inch. The magnitude of the force shall be noted at this point. The face of the instrument applying the force shall be circular and shall have a diameter not less than 1/2 inch nor more than 1 inch.

5. PACKAGING.

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One package containing twenty individually packaged kits as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Individual package. Components shall be arranged within the plastic tray in a neat and orderly manner. The tray shall be wrapped in the overwrap, inserted into a paperboard box or sleeve, and packaged into a clear plastic peel type bag and the closure effected by heat sealing. The assembly shall be so packaged as to maintain sterility of contents unless the package

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:40Z
Check the source to verify that this is the current version before use.

is damaged or opened. As an alternate, the wrapped tray shall be inserted into a sealed peel open type bag and the bag packaged in a snug fitting box of an appropriate design. The wrapped tray may also be inserted into a sealed snug polyethylene coated paperboard box of an appropriate design. The wrapped tray may also be inserted into a formed blister pack with a peelable, tear resistant, moisture resistant, heat sealable lid closure. All alternate packs shall be capable of maintaining sterility of contents unless packages are damaged or opened.

5.1.2.2 Unit package. Twenty individually packaged kits shall be packaged in a snug fitting box constructed in accordance with PPP-B-636, type CF, class weather resistant. Closure and waterproofing shall be as specified in the appendix of PPP-B-636.

5.1.3 Level C. Kits shall be individually packaged in such a manner so as to be capable of maintaining sterility of contents unless the package is damaged or opened. The packaged kits shall be further packed in shipping containers that will afford adequate protection against damage during direct shipment from the supply source to the first receiving. These packs shall conform to the applicable carrier rules and regulations.

5.2 Packing.

5.2.1 Levels A and C. Units packaged as specified in 5.1.2 and 5.1.3 shall require no further packing.

5.3 Marking.

5.3.1 Individual package. In addition to contractors commercial markings, each individual package shall be marked in accordance with requirements of the Federal Food, Drug and Cosmetic Act. Marking shall include the item identification, instructions for use, and the legend: "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED" or similar statement.

5.3.2 Unit package. Each unit package shall be marked in accordance with the requirements for exterior containers as specified in MIL-STD-129 and the requirements of the Federal, Food, Drug and Cosmetic Act. The method of sterilization (G-Gas, R-Radiation, C-Cold, S-Steam, D-Dry Heat Sterilization) shall be shown following the level of packaging and the date of packaging. Example "A-3/76-G". Final sterilization shall be on or after indicated packaging date. In addition, the lot (control) number shall be shown.

6. NOTES

6.1 Intended use. The irrigation kit is intended for use in bladder, gastrointestinal and wound irrigation and for preparation of the patient, protection of contiguous areas of the bed, obtaining and transporting specimens and clean-up of the patient in nursing unit procedures.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:40Z
Check the source to verify that this is the current version before use.

6.2 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging (see 5.1).
c. When a first article is required (see 3.1, 4.1.2.1 and 6.3).

* 6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions of 7-104.55 of the Armed Services Procurement Regulation. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

6.4 This specification covers the following item appearing in the Federal Supply Catalog:

NSN	Item Identification
6530-00-117-8423	Irrigation Kit, Patient, Disposable, 20s.

6.5 Changes from previous issue. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DSA-DM

Project No. 6530-1338

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS. The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-I-36969A IRRICATION KIT, PATIENT

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

* GPO 1974-- 603 680/6126 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:40Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:40Z
Check the source to verify that this is the current version before use.