MIL-I-36996
Amendment-2
10 October 1974
SUPERSEDING
Amendment-1
1 March 1974

MILITARY SPECIFICATION

INTRAVENOUS INJECTION SET

This Amendment forms a part of Military Specification MIL-I-36996, dated 8 October 1973, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 1:

1.1 At end of paragraph delete period (.) and add ", or with an integral airway connector."

Page 3:

3.1.1.1 Line 2, delete "75" and substitute "66 or 75".

Page 4:

3.2 Line 12, delete "and shall not be solvent welded or cemented" and substitute "and shall be solvent sealed."

3.2.2 Lines 6 and 7, delete "nominal 1/4 inch" and substitute ".260 $\pm$ .002 or .265 $\pm$ .002."

Page 5:

3.2.2 Lines 34, 35 and 36, delete sentence beginning with, "The air inlet etc." and substitute "The fit of the Luer cavity shall be tested by means of the plug gage (without feeler gage) as per Figure 5 of ANSI Z70.1 with a notch (denoting tolerance) extending for the entire 0.057 inch thickness of the feeler gage which includes the 'GO-NO-GO' thickness. As per Figure 5, with the feeler gage removed, the fit of the female Luer cavity shall not bottom out on the plug gage and must be within a tolerance including the "GO-NO-GO' dimensions of the feeler gage." Lines 43, 44 and 45, delete "15-20" and substitute "14-23". The instructions supplied with the set (see 3.10) shall include the drop size (i.e. number of drops per ml.). The drop size of each set shall not vary by more than plus or minus 3 drops per ml. from the drop size indicated in the instructions".

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:44Z
Check the source to verify that this is the current version before use.

Page 6:

* 3.2.3 Line 2, delete "78" and substitute "60".

3.2.4 Line 3, after 45 ± 5° insert "or 30 ± 5°".

* 3.2.5 Line 2, delete period (.) and add ", or a roller type clamp."

* 3.3 Line 8, delete "71" and substitute "70".

Add the following paragraph:

* 3.3.1 Drip chamber. The drip chamber housing shall have an effective volume between 4.0 ml minimum and 7.0 maximum. This volume to be measured from the lower end of the extended micro-tip to the distal end of the chamber.

Page 10:

* 4.2.1.1 Line 3 after "conditions" insert period (.) and delete remainder of paragraph.

Page 17:

* 4.4.9 Line 4, delete "for 30 seconds."

Page 19:

5.1.3.1 Individual box. At end of paragraph add, "As an alternate, each set shall be packaged in an open tray, overwrapped in cellophane. Wrapper of tray shall be sealed to prevent accidental opening. Wrapper shall have a tear strip for ease of opening."

Page 20:

5.2.1.1 Exterior container. Line 4, delete "grade V3c."

5.3.1 Individual box. Line 1, following "box" insert: "or cellophane wrapped tray". Line 2 following "box", insert "or cellophane wrapped tray".

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:44Z
Check the source to verify that this is the current version before use.

Page 21:

Add the following paragraph:

"6.4 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue."

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1837

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:44Z
Check the source to verify that this is the current version before use.