A-A-0030120(VA-OSS)
January 8, 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

SPONGE, SURGICAL

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Veterans Administration Marketing Center, Marketing Division for Medical Supplies (904A), P.O. Box 76, Hines, Illinois, 60141, based upon currently available technical information.

This interim commercial item description covers a tamperproof sealed tray containing a sterile x-ray detectable surgical gauze sponge suitable for use in surgical procedures.

Salient characteristics. The sponge shall be in a tamperproof sealed tray containing 10 sterile 4 x 8 inches, 12 ply, x-ray detectable surgical gauze sponges, suitable for use in surgical procedures. The sponge shall be 12 ply, USP Type VII absorbent gauze.

The sponge is to be designed so that it can be opened to 4 x 24 inches without exposing raw edges and shall not contain nor be prone to develop in use, loose threads or filaments. The sponge shall not fray or shred when wet.

The x-ray detectable insert shall be fabricated of plastic material and shall be impregnated with an inert, nontoxic substance which is opaque to x-rays. The radiopaque insert may be monofilament not less than 0.381 mm in thickness and equal to the length of gauze in the warp direction, and the length shall be a minimum of four times the width. The radiopaque or x-ray detectable insert shall be anchored between the layers of gauze in such a manner that it will not fall out when the sponge is unfolded and vigorously shaken.

The tray shall be moisture resistant and of sufficient rigidity to serve as a vessel for fluids and as the immediate container for 10 sponges.

The sponges in the immediate container shall be precounted by the contractor. The sponges shall be banded or tied with a colorfast string into a precounted bundle.

The x-ray detectable insert shall not bleed nor show signs of deterioration, the precount band shall show no sign of decomposition or parting at the seal, and the precount string shall show no sign of decomposition, no appreciable change from its original color and no staining of the cloth when they are subjected to saturated steam at a temperature of 250°F for one hour.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification: The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements

Federal Food, Drug and Cosmetic Act - If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials - The offeror/contractor is encouraged to use recovered material in accordance with Public Law 580, as amended and as implemented by Federal Acquisition Regulation 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking - Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:36Z
Check the source to verify that this is the currentversion before use.

NOTE. The following National Stock Number is covered by this document:

6510-00-116-1285

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

PHS
FDA-MPQAS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92B)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:36Z
Check the source to verify that this is the currentversion before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.