INCH-POUND

A-A-30172
April 20, 1993

COMMERCIAL ITEM DESCRIPTION

SPONGE, SURGICAL (STERILE)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes a surgical sponge, for general use in the dressings of wounds and for postoperative wound dressings.

Salient characteristics. Sponge, Surgical, White, absorbent, plain, all gauze, sterile, (package/product must be marked sterile, no strike through or bleeding of the labeling), sealed, peel-open design. Shall meet the requirements of USP Type VII, for Absorbent Gauze, except the filling thread count may be 8 to 11. Sponge shall be sterile when tested in accordance with USP. The sponge shall be folded and shall not have any exposed raw edges and shall be free of loose threads, ravelings and foreign matter. Package must be clean and without holes or tears.

The sponges shall be of the size and ply as listed below:

SIZE	PLY	DIMENSIONS INCHES	UNIT PACKAGE QUANTITY FOR DOD
1	8	2" x 2" (5.08 cm x 5.08 cm) ± 1/8" (0.32 cm)	A - Two sponges shall be sealed in an envelope, 1500 envelopes (3000 sponges) constitutes one unit. OR B - Two sponges per envelope, 50 envelopes per package, 30 packages per case.

Shall be capable of being opened to approximately 2" x 4" (5.08 cm x 10.2 cm) without exposing raw edges or loose threads. When envelopes are opened the sponges shall fall freely or the envelopes shall provide positive control for dispensing.

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

SIZE	PLY	DIMENSIONS INCHES	UNIT PACKAGE QUANTITY FOR DOD
2	12	4" x 4" (10.2 cm x 10.2 cm) ± 1/4" (0.064 cm)	B - Ten sponges shall be sealed in a tamperproof (not capable of being resealed) tray. One package containing 128 trays (1280 sponges) OR C - One package containing 144 trays (1440 sponges) constitutes one unit.

Shall be capable of being opened to approximately 4" x 12" (10.2 cm x 30.5 cm) without exposing raw edges or loose threads. Tray covers shall be peel-open and the trays shall be waterproof and be of sufficient rigidity to serve as a vessel for fluids and also as the immediate container for 10 sponges.

Workmanship. The sterile surgical sponge shall be free from defects which detract from its appearance or impair its serviceability.

Unit. VA requirements shall be as stated in the solicitation.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for test shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one sponge. Sampling shall be inspection level S-1 with an AQL of 2.5 (percent defective).

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination shall be one sponge. Examination shall be inspection level of S-1 with an AQL of 2.5 (percent defective).

Examination. Examination shall be conducted to determine complaince with specification requirements.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document.

NSN	ITEM DESCRIPTION	SIZE	PLY	UNIT PACKAGE QUANTITY FOR DOD
6510-00-058-4421	SPONGE, SURGICAL 2" x 2" (5.08 cm x 5.08 cm)	1	8	A
6510-00-559-3219	SPONGE, SURGICAL 4" x 4" (10.2 cm x 10.2 cm)	2	12	
6510-00-116-1311	SPONGE, SURGICAL 4" x 4" (10.2 cm x 10.2 cm)	2	12	B or C

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government, may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Size and ply

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VANAC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6510-0822

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.