A-A-30181A
July 15, 1992
SUPERSEDING
A-A-30181
APRIL 30, 1992

# COMMERCIAL ITEM DESCRIPTION

## URINAL INCONTINENT

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes a day and night device, worn by male incontinent patients.

Salient characteristics. Urinal, Incontinent consists of two component parts:

TYPE I Urinary Collection Appliance, complete shall consist of a suspensory and urine reservoir or penile sheath. A 1-1/2 inch (38 mm) diameter latex cone shaped sheath to fit snugly around penis to prevent leakage, adjustable elastic waist belt 1 inch (25 mm) ± 1/4 inch (6 mm) wide by 32 inches (81 cm) ± 1 inch (25 mm) long (unstretched); two rubber thigh straps with buckles. The collection appliance shall be provided with instructions for use.

TYPE II Latex Leg Bag 32 oz. (946 ml)

Leg bag or collection unit shall be compatible with top portion, and shall allow the connection of an external male catheter, conical shaped to the top of the bag, and shall have a step down connector that connects the tubing to the male catheter (the larger dimension shall be the top). Bag shall have a 32 oz. (946 ml) capacity with reinforced rolled seams or seamless one piece latex, with connectors at each end and an inner valve to prevent back-flow, supplied with two leg straps, each 21 inches (53 cm) ± 2 inches (51 mm) long (unstretched), having a minimum of 5 holes, (1/8 inch diameter) (3.17 mm) equally spaced at each end to provide a secure fit; a screw drainage valve for bottom connector and an extension drainage tube 6 inches (152 mm) long with connector and leakproof shut-off clamp; or an extension drainage tube 6 inches (152 mm) long, attached to the bottom of the bag with leakproof shut-off clamp; inlet connector shall be threaded or O-ring sealing type for top of bag. The bag shall be a minimum of .035 (0.89 mm) inch thick. The leg bag or collection unit shall have a minimum life expectancy of three months, or shall be provided with sufficient replacement parts



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

to achieve a minimum life expectancy of three months. Pinch points, protrusions, sharp edges, or other features which contribute to patient discomfort are not allowed. Instructions shall be provided for the use of the bag or collection unit and all component parts.

TYPE I - Urinary Collection Appliance, Complete

TYPE II - Latex Leg bag, 32 oz (946 ml).

Workmanship. The Urinal, Incontinent shall be free from defects which detract from its appearance or impair its serviceability.

Unit. The unit of issue shall appear in the VA solicitation.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be one incontinent urinal as specified in this description.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

Type I - NSN 6530-00-723-6782 Urinary Collection Appliance Complete

TYPE II - NSN 6530-00-723-6783 Latex Leg Bag - 32 oz (946 ml)

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Type

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VANAC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.