INCH/POUND

A-A-30189
September 18, 1992

COMMERCIAL ITEM DESCRIPTION

RESPIRATOR, PARTICULATE HEALTHCARE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes a disposable particulate respirator suitable for use by medical personnel in areas where exposure to patients with infectious tuberculosis is anticipated.

Salient characteristics. Respirator, Particulate Healthcare shall be a disposable, cup shaped, molded design, with two elastic bands secured with staples. The area of the body of the respirator which fits over the bridge of the nose shall have an aluminum malleable nose-piece centrally located on the respirator 1/4 in. (6mm) from the perimeter seal and firmly cemented to the outside surface. A foam strip that fits over the bridge of the nose shall be attached on the inside of the respirator. The nose-piece shall be easily formed to the contours of the wearers nose to minimize the loss of expired air between the respirator and the face in this area. The respirator shall be NIOSH/MSHA certified as a particulate respirator for use against dusts and mists with a PEL (permissible exposure limit) not less than 0.05 milligrams per cubic meter. The filtering efficiency of the respirator shall not be less than 99% when tested by the method of Greene and Vesley (J Bacteriol., vol. 83, No. 3, p. 663, March 1962 , and no less than 80% efficiency (20% penetration) when tested by DOP (dioctyl phthalate) submicron filtration efficiency test, ASTM D2986-71. Delta P shall be no more than 2.2 ± 0.2 (MMH2O). The respirator mask shall be fluid resistant as defined by meeting the following requirement when tested in accordance with hydrostatic pressure test method ASTM D583-63 Method II (AATCC 127-1985):

Minimum average = 6.0 cm
Sample size = 15 Respirators

Elastic Bands:
Top: Minimum 12-1/2 inches (32cm) long
Bottom: Minimum 10-1/4 inches (26cm) long

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Aluminum nose-piece 3-1/2 inches (9cm) long by 3/16 inch (0.476cm) wide.

The respirator shall meet the requirements of Class 1 Code of Federal Regulation CFR Part 16. The mask shall not cause irritation to the skin and eyes when tested using human panel and rabbit subjects in compliance with Federal Regulation CFR Part 16, Section 1500.41 and 1500.42.

Workmanship. The respirator shall be free from defects which detract from its appearance or impair its serviceability

Unit. The unit of issue shall appear in the VA solicitation.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be one respirator as specified in this description.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE The following National Stock Number is covered by this document:

NSN 6515-01-363-1641

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Military Coordinating Activity.

DoD-MB

AGENT.

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VANAC (904)
Agent VA-OSS VACC (92A)

Project No. 6515-5501

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.