INCH-POUND

A-A-30193
January 6, 1993

COMMERCIAL ITEM DESCRIPTION

CHAIR, BATHTUB AND SHOWER

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes a chair used in the transfer of patients in and out of the tub.

Salient characteristics:

CHAIR, BATHTUB AND SHOWER, shall be a minimum of 7/8 inch (2.2 cm) diameter rust resistant anodized aluminum tubing; reversible patient seat and back; 4 adjustable legs level to the floor and provided with non-marring suction cups and/or rubber tips (or both) that are skid resistant; 3 vinyl cushion seat pads and 1 vinyl cushion backrest; with the following measurements:

backrest height a minimum of 11-3/4 inches (measured from top of seat to top of backrest) (29.8 cm);
seat height (in lowest position) minimum 16 inches to a minimum of 19 inches in highest position (measured from floor to top of seat) (41 cm to 48 cm);
seat length, minimum 15 inches (38 cm) (seat only measured);
bench length, minimum 24 inches (61 cm);
seat depth/width, 16 inches, plus or minus 1-1/4 inch (41 cm plus or minus 3 cm);
overall bench depth, minimum 20 inches (51 cm) (measured from front leg to backrest);
overall bench length, minimum 29 inches (measured at outside of bottom leg) (74 cm);
weight of chair, 10 to 15 pounds (4.5 kg to 6.8 kg);
weight capacity a maximum of 250 pounds (113 kg).

Workmanship. The chairs shall be free from defects which detract from their appearance or impair their serviceability.

Unit. EACH (EA).

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

NOTE. The following National Stock Number is covered by this document:

NATIONAL STOCK NUMBER	NOMENCLATURE
6530-01-191-5431	CHAIR, BATHTUB AND SHOWER

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Copies of Federal Test Method Standard 191 may be obtained from your local General Services Administration.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description

MILITARY INTERESTS: CIVIL AGENCY COORDINATING ACTIVITIES.

Military Coordinating Activity: FDA-MPQAS
PHS

DoD-MB

PREPARING ACTIVITY:

VA-OSS VANAC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6530-2392

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.