MIL-I-23870(Weps)
19 September 1963

MILITARY SPECIFICATION

INDICATOR, FUEL RATE OF FLOW, 0 TO 15 GPM

This specification has been approved by the
Bureau of Naval Weapons, Department of the Navy

1. SCOPE

1.1 This specification covers one type of fuel flow indicator for use with 2-1/2 inch inside diameter (ID) fuel lines in aircraft ground refueling systems.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids, form a part of this specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

P-S-661 - Solvent, Dry Cleaning
QQ-C-320 - Chromium Plating (electrodeposited)
QQ-P-416 - Plating, Cadmium (electrodeposited)

MILITARY

MIL-P-116 - Preservation, Methods of
MIL-S-3136 - Standard Test Fluids Hydrocarbon and Iso-Octane
MIL-E-5272 - Environmental Testing, Aeronautical and Associated equipment, General Specification for
MIL-E-5557 - Enamel; heat-resisting, Clycerylphthalate, Black
MIL-G-5572 - Gasoline, Aviation: Grades 80/87, 91/96, 100/130, 115/145
MIL-J-5624 - Jet Fuel, Grades JP-4 and JP-5
MIL-P-6906 - Plates, Information and Identification
MIL-F-7024 - Fluids, Calibrating, for Aircraft Fuel System Components
MIL-S-7742 - Screw Threads, Standard Optimum Selected series; General Specification for
MIL-P-7936 - Parts and Equipment, Aeronautical, Preparation for delivery
MIL-P-8585 - Primer Coating, Zinc Chromate, Low-Moisture Sensitivity
MIL-A-8625 - Anodic Coatings, for Aluminum and Aluminum Alloy
MIL-N-5877 - Nozzle, Pressure Fuel Servicing, locking, Type D-1
MIL-C-19179 - Coupling Assembly, quick disconnect, fuel
MIL-F-8615 - Fuel System Components; general specification for
MIL-N-25027 - Nut, Self Locking, 250°F, 500°F, and 800°F
MIL-D-70327 - Drawings, Engineering and Associated List

FSC 6680

Source: https://assist.dla.mil -- Downloaded: 2026-10-19T18:09Z
Check the source to verify that this is the current version before use.

STANDARDS

MILITARY

| | |
|---|---|
| MIL-STD-105 | - Sampling Procedures and tables for inspection by attributes |
| MIL-STD-129 | - Marking for Shipment and Storage |
| MIL-STD-130 | - Identification Marking for U.S. Military Property |
| MS20995 | - Wire, Lock |
| MS29513 | - Packing - "O" Ring Hydrocarbon Fuel Resistant |
| MS33540 | - Safety Wiring, General Practices for |
| MS33586 | - Metals, definition of Dissimilar |
| MS33588 | - Nuts and Plate Nuts, Self-Locking, Functional Limitations of |
| MS33786 | - Fitting Installation, Flared Tube and Hose, Swivel |

PUBLICATIONS

AIR FORCE-NAVY AERONAUTICAL BULLETINS

No. 143
No. 438

(Copies of specifications, standards, drawings, and publications required by Contractor in connection with specific procurement functions should be obtained from the procuring activity or as directed by the Contracting Officer)

3. REQUIREMENTS

3.1 Preproduction: Fuel flow indicators furnished under this specification shall be of a product that has passed the preproduction tests specified herein (see paragraphs 4.2 and 6.2)

3.2 Materials and Processes: All materials used in the indicator shall be sufficiently resistant to fuels conforming to Specification MIL-G-5572 and MIL-J-5624 with Aromatics of 0 to 30 percent to assure satisfactory operation as herein defined. Except as provided in the following paragraph, specifications and standards for all materials, parts, and processes which are not specifically designated herein and which are necessary for the execution of this specification, shall be selected in accordance with ANA Bulletin 143.

3.2.1 Standard Parts: Standard parts (MS and AN) shall be used wherever they are suitable for the purpose, and shall be identified on the drawing by their part numbers. Commercial utility parts such as screws, bolts, and nuts, and cotter pins may be used, provided they possess suitable properties and are replaceable by the standard parts (MS and AN) without alteration, and provided the corresponding standard part numbers are referenced in the parts list, and if practicable, on the Contractor's drawings. In the event



Source: https://assist.dla.mil -- Downloaded: 202610-19T18:09Z
Check the source to verify that this is the current version before use.

there is no suitable corresponding standard part in effect on date of invitation for bids, commercial parts may be used, provided they conform to this specification.

3.2.2 Metals: Metals shall be of corrosion-resistant type or shall be suitably treated to resist corrosion during storage and service use. The use of magnesium or any alloy thereof is prohibited.

3.2.2.1 Dissimilar Metals: The use of dissimilar metals shall be avoided wherever practicable or used in accordance with Standard MS33586.

3.2.2.2 Sparking: Metals that generate sparks when struck by steel or flint shall not be used.

3.2.3 Castings: Castings shall be of uniform quality, clean, sound, and free from blowholes, porosity, cracks, and any other defects.

3.3 Design and Construction

3.3.1 Design: The fuel flow indicator shall visibly indicate the flow of fuel in one direction and shall withstand the pressures that may be encountered in fueling services as specified herein. The flow indicator entrance and exit flanges shall conform to Standard MS33786.

3.3.1.1 Weight: Minimization of weight shall be a design consideration to the maximum degree obtainable consistent with all other requirements of this specification.

3.3.1.2 Lubrication: The flow indicator shall not require the use of lubricants other than fuel in the indicator.

3.3.1.3 "O" Rings: All "O" Rings shall conform to Standard MS29513.

3.3.1.4 Screw Threads: All screw threads shall be in accordance with Specification MIL-S-7742.

3.3.1.5 Locking of Parts: All threaded parts shall be securely locked by cotter pins, safety wiring, self-locking nuts, or other approved method. Self-locking nuts shall be of the all-metal type conforming to specification MIL-N-25027, and shall be used in accordance with Standard MS33588. Self-locking nuts shall not be used when loosening or disengagement of the nut could result in nut or other parts entering the fuel system. Safety wire shall conform to Standard MS20995 and shall be installed in accordance with Standard MS33540. The use of staking or lock washers is prohibited.

3.3.1.6 Dimensions: All parts shall conform to the manufacturer's drawings, as approved by the procuring activity. Where dimensions and tolerances may affect interchangeability, operation, or performance of the flow indicator, they shall be held or limited accordingly.

3.3.1.7 Interchangeability: All parts having the same manufacturer's

Source: https://assist.dla.mil -- Downloaded: 202610-19T18:09Z
Check the source to verify that this is the current version before use.

part number shall be directly and completely interchangeable with each other with respect to installation and performance. Changes in manufacturers' part numbers shall be governed by the drawing number requirements of Specification MIL-D-70327, or as otherwise specified by the procuring activity.

3.3.2 Construction: The flow indicator shall be so constructed as to withstand the strains, jars, vibrations, and other conditions incident to shipping, storage, installations, and service usage, particularly dropping on a steel plate from a height of 6 feet without any damage affecting operation.

3.3.2.1 Cleaning: All parts shall be clean and free from dirt, sand, and metal chips while being assembled and after assembly.

3.4 Synthetic Rubber Parts

3.4.1 Age Control: Age control of all synthetic rubber parts shall be in accordance with ANA Bulletin 438.

3.4.2 Serviceability: All synthetic rubber parts shall be readily replaceable.

3.5 Performance

3.5.1 Internal Pressure: The flow indicator shall be capable of withstanding a sustained internal pressure of 55 psi at a flow rate of 225 GPM and a momentary internal pressure of 240 psi at no flow.

3.5.2 Pressure Loss: The hydraulic pressure loss shall be less than 2 psi at a flow rate of 225 GPM.

3.5.3 Flow Indication

3.5.3.1 Range and Scale: The range of flow indication shall be from zero to 15 GPM. The scale shall be marked in increments of one (1) GPM and beginning with Zero (0), shall be marked numerically at 5 GPM intervals.

3.5.3.2 Accuracy: The error in flow indication with the indicator in any attitude shall not exceed one (1) GPM, or 10 percent of scale reading, whichever is greater, for JP-5 fuel at temperatures between 15°F and 100°F. The error shall not exceed one (1) GPM or 15 percent of scale reading, whichever is greater, for JP-4 fuel at temperatures between -30°F and 60°F. The effect of service use, as simulated by the preproduction test (paragraph 4.2), shall be less than 2 percent.

3.5.3.3 Readability: The scale length and graduations shall permit reading to the nearest GPM from a distance of 18 inches.

3.6 Finish

3.6.1 Anodizing: All aluminum-alloy parts in contact with the fluid or

Source: https://assist.dla.mil -- Downloaded: 2026-10-19T18:09Z
Check the source to verify that this is the current version before use.

atmosphere shall be anodized in accordance with Specification MIL-A-8625 or shall be adequately treated in some other acceptable manner for corrosion prevention.

3.6.2 Surface Treatment of Parts: Steel parts, other than corrosion-resistant steel not in moving contact, shall be cadmium plated in accordance with Type II, of Specification QQ-P-416. Cadmium plating shall not be used on surfaces which would normally be in contact with fuel, unless specifically spproved by the procuring activity. Wearing surfaces shall be chromium plated in accordance with Specification QQ-C-320, or adequately treated in some other acceptable manner for corrosion prevention where the galling of cadmium plating would cause malfunction. Brass, bronze, and copper-alloy parts shall be cadmium plated or adequately treated in some other acceptable manner for corrosion prevention. Stainless steel parts shall be passivated.

3.6.3 Paint Finish: Any part painted shall be finished with one coat of primer conforming to Type I of Specification MIL-P-8585, and one coat of enamel conforming to Type II or Type IV of Specification MIL-E-5557. Surfaces that come in direct contact with fuel shall not be painted.

3.7 Markings: All markings shall be permanent to prevent obliteration resulting from service use. The body of the indicator shall be marked with an arrow to indicate direction of flow.

3.8 Identification of Product: Equipment, assemblies, and parts shall be marked in accordance with Specification MIL-STD-130.

3.8.1 Name Plate: A name plate conforming to Specification MIL-P-6906 and containing the following information, properly and legibly filled in, shall be securely attached to the indicator or the same information may be etched, engraved, or embossed in a suitable location on the body of the indicator:

Indicator, Fuel Rate of Flow, 0 to 15 GPM
Specification MIL-I-23870(Weps)
Purchaser's Part No. (If applicable)
Manufacturer's Part No.
Manufacturer's Serial No.
Manufacturer's Name or Trademark
US

3.9 Workmanship: Workmanship shall be of uniform quality and such that the finished product is free of surface defects and sharp corners or other proturbances likely to cause personal injury or property damage in the course of normal use and handling in service.

## 4. QUALITY ASSURANCE PROVISIONS

4.1 Inspection Responsibility: The supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own or any other inspec-



Source: https://assist.dla.mil -- Downloaded: 2026-10-19T18:09Z
Check the source to verify that this is the current version before use.

tion facilities and services acceptable to the Government. Inspection records of the examination and tests shall be kept complete and available to the Government as specified in the contract or order. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Rejection and Retest: Items which have been rejected may be reworked or replaced to correct the defects, and resubmitted. Before submitting, full particulars concerning previous rejection and the action taken to correct the defects found in the original shall be furnished the inspector. Units rejected after retest shall not be resubmitted without the proper approval of the procuring activity.

4.1.2 Classification of Inspections and Tests: The inspection and testing of flow indicators shall be classified as follows:

a. Preproduction Tests (Para. 4.2)

b. Acceptance Tests (Para. 4.10)

c. Preparation for Delivery (Para. 5)

4.2 Preproduction Tests: Preproduction tests shall consist of Examinations (See Para. 4.10), and the following additional tests in the order shown:

a. Calibration and Pressure Loss (Para. 4.3)

b. Fuel Resistance (Para. 4.4)

c. Accelerated Corrosion (Para. 4.5)

d. Maximum Pressure and Leakage (Para. 4.6)

e. Endurance Test (Para. 4.7)

f. Recalibration (Para. 4.8)

4.3 Calibration and Pressure Loss Test

4.3.1 Test Conditions: A test fluid conforming to Specification MIL-J-5624, Grade JP-4 or JP-5, Specification MIL-F-7024, Type II or P-S-661 and a test set-up similar to that shown in Figure 1 shall be used. Fluid temperature shall be maintained at 75 plus or minus 15°F.

4.3.2 Test Procedure: Test fluid shall be passed through the flow indicator in one (1) GPM increments of flow rate from 0 to 15 GPM and increments of 100 GPM of flow rate from 25 to 225 GPM, while maintaining the inlet static pressure at 45 plus or minus 5 psi. Inlet and outlet static pressures and volumetric flow rate shall be recorded at the stabilized test points.



Source: https://assist.dla.mil -- Downloaded: 202640-19T18:09Z
Check the source to verify that this is the current version before use.

4.4 Fuel Resistance Test: If the flow indicator incorporates synthetic rubber parts in contact with fuel, the indicator shall be subjected to the accelerated aging test outlined in Table II.

4.5 Accelerated Corrosion Test

4.5.1 Test Procedure: Spray the interior of the indicator (through the inlet and outlet openings) with a saturated solution of sodium chloride and distilled water. Allow the indicator to drain at room temperature for 20 minutes with the longitudinal axis inclined 45 degrees from horizontal. Then dry the indicator in an air oven at 130°F for a period of 20 minutes. Repeat the spray-drain-dry procedure to obtain a total of 50 cycles. Upon completion of the cycles, wash the indicator thoroughly with hot water and dry in a moving stream of air.

4.6 Maximum Pressure and Leakage Test

4.6.1 Preparation of Specimen: The interior of the indicator shall be sprayed with a fluid conforming to Specification MIL-J-5624, Grade JP-4 or JP-5, Specification MIL-F-7024, Type II, or Specification P-S-661 and then capped with end fittings, one of which shall have provisions for admitting air under pressure which can be measured by a suitable gage.

4.6.2 Test Procedure: The prepared specimen shall be allowed to "soak" for a period of one hour at room conditions. It shall then be submersed in a tank of water and, while submersed, shall be subjected to an internal air pressure of 240 psi for a period of at least one (1) minute. After this period and before releasing the air pressure, the submerged specimen shall be examined for evidence of leakage.

4.6.3 Examination after Test: Upon completion of the test, the indicator shall be examined for evidence of distortion or other indications of non-compliance with requirements of paragraphs 3.3.2 and 3.5.1.

4.7 Endurance Test (Cyclic Operation)

4.7.1 Test Conditions: Test conditions shall be as specified in paragraph 4.3.1. The flowmeter used for calibrating the flow indicator and establishing test flow rates may be isolated from the system during the endurance portion of the test.

4.7.2 Test Procedure: With the main outlet valve closed and the inlet and vernier valves adjusted to give an inlet static pressure of 55 plus or minus 5 psi at a flow rate of 15 GPM, the flow rate shall be cycled between 15 GPM and 225 plus or minus 25 GPM by alternately opening and closing the main outlet valve for a total of 10,000 complete cycles (a complete cycle consists of one opening and one closing).

4.8 Recalibration: Upon completion of the endurance test the indicator shall be subjected to the calibration and pressure loss test (paragraph 4.3) to determine compliance with paragraph 3.5.3.2.



Source: https://assist.dla.mil -- Downloaded: 202610-19T18:09Z
Check the source to verify that this is the current version before use.

4.9 Environmental

4.9.1 Fungus Resistance: This test shall be conducted in accordance with the fungus resistance tests of Specification MIL-E-5272, except that substitution for the specified fungi may be made with the approval of the procuring activity.

4.9.2 Sand and Dust: The sand and dust test shall be conducted in conformance with the sand and dust tests, Procedure I of Specification MIL-E-5272.

4.10 Acceptance Tests: Acceptance tests shall consist of the following tests conducted on samples selected from individual lots submitted for acceptance:

| | | |
|---|---|---|
| a. | Examination | (Para. 4.10.1) |
| b. | Maximum Pressure and Leakage | (Para. 4.6) |
| c. | Calibration and Pressure Loss | (Para. 4.3) |

4.10.1 Examinations (of product): Samples selected for acceptance tests shall be visually examined for compliance with the following requirements:

| | | |
|---|---|---|
| a. | Markings | (Para. 3.7) |
| b. | Identification of Product | (Para. 3.8) |
| c. | Workmanship | (Para. 3.9) |

4.10.2 Lot: An individual lot shall consist of 50 units or, at the contractor's option, the number of flow indicators produced in one or more full working days so chosen as to result in an accumulation of 41-65 units.

4.10.3 Sampling: Sampling shall be in accordance with Standard MIL-STD-105.

4.10.4 Acceptable Quality Level (AQL): An AQL of 4% shall be used for major defects. An AQL of 10% shall be used for minor defects.

4.10.5 Classification of Defects:

| Major (AQL-4) Nature of Defect | Inspection or Test |
|---|---|
| 1. Leakage or distortion under pressure | Para. 4.6 |
| 2. Excessive error in flow indication | Para. 3.5.3.2 |

| Minor (AQL-10) | |
|---|---|
| 1. Workmanship | Para. 3.9 |
| 2. Marking (missing of illegible) | Para. 3.7 |



Source: https://assist.dla.mil -- Downloaded: 2026-10-19T18:09Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Preservation and Packing

5.1.1 Level A: When this level is required, preservation and packaging shall be in accordance with the applicable method or sub-method of Specification MIL-P-116, utilizing the appropriate containers in accordance with Specification MIL-P-7936.

5.1.2 Level C: When this level is required, preservation and packaging shall be in accordance with standard commercial practice.

5.2 Packing: Packing shall be in accordance with Specification MIL-P-7936.

5.3 Marking of shipment: In addition to any special markings required by the contract or order, unit packages, intermediate packages, and shipping containers shall be marked in accordance with Standard MIL-STD-129.

6. NOTES

6.1 Intended Use: The flow indicator is intended for installation between the hose and the nozzle of pressure-fueling line to indicate the combined leakage rate of all fuel level control valves in an aircraft undergoing refueling.

6.2 Definitions

6.2.1 Preproduction Test: The preproduction test is that group of tests conducted to demonstrate the suitability of a product for production and service use.

6.2.2 Acceptance Test: The acceptance test is that group of test conducted on samples from production lots submitted for acceptance under contract to demonstrate end item quality control, correct assembly, and performance.

6.3 Ordering Data: Procurement documents should specify the following:

a. Title, Number, and Date of this specification.
b. Quantity of flow indicators required.
c. Time and place (if other than contractor's plant) for preproduction testing (See Paragraphs 3.1 and 4.2).
d. Applicable levels of packaging and packing. (See Section 5.)



Source: https://assist.dla.mil -- Downloaded: 20260-19T18:09Z
Check the source to verify that this is the current version before use.

NOTICE: When government drawings, specifications, or other data are used for any purpose other than in connection with definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise and in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.



Source: https://assist.dla.mil -- Downloaded: 2026-10-19T18:09Z
Check the source to verify that this is the current version before use.

TABLE II

Fuel resistance and extreme temperature test schedule
for components incorporating moving parts but no electrical equipment

| Test | Fuel resistance | | | | |
|---|---|---|---|---|---|
| Period 1/ | Phase I soak | Phase I dry | Phase II soak | Phase II dry | Low temperature |
| Component configuration | 2/ | Drained and blown dry, normal condition as would be expected under service conditions, ports open. | 2/ | Drained and blown dry, normal condition as would be expected under normal service conditions, ports open. | Mounted as would be expected under normal service conditions. 2/ |
| Test fluid | MIL-S-3136, type III | None | MIL-S-3136, type III | None | MIL-S-3136, type I |
| Period duration | 96 hours (4 days) | 24 hours | 18 hours | 30 hours | 18 hours |
| Ambient and test fluid temperature. | 158° ±2°F or the normal operating temperature of the system in which the component is used, whichever is higher. | Circulating air at 158° ±2° F or the normal operating temperature of the system in which the component is used, whichever is higher. | 158° ±2°F or the normal operating temperatore of the system in which the component is used, whichever is higher. | Circulating air at 158° ±2°F or the normal operating temperature of the system in which the component is used, whichever is higher. | Lower the fluid temperature to -67° ±2°F, then maintain the fluid temperature at -67° ±2°F for a minimum of 18 hours. 3/ |
| Operation or tests during period. | Actuate component at least 4 cycles per day in a normal manner. | None | Actuate component at least 4 cycles in a normal manner. | None | None |





Source: https://assist.dla.mil -- Downloaded: 2026-10-19T18:09Z
Check the source to verify that this is the current version before use.

TABLE II (Continued)

Fuel resistance and extreme temperature test schedule
for components incorporating moving parts but no electrical equipment

| Test | Fuel resistance | | | | |
|---|---|---|---|---|---|
| Period 1/ | Phase I soak | Phase I dry | Phase II soak | Phase II dry | Low temperature |
| Operation or tests immediately after period. | Conduct leakage test, using MIL-S-3136, type III fluid. 4/ | (a) Actuate components for 5 cycles. (b) Conduct functional and leakage tests, using MIL-S-3136 type I fluid. 4/ | Conduct leakage test, using MIL-S-3136. type III fluid. 4/ | (a) Actuate components for 5 cycles. (b) Conduct functional and leakage tests, using MIL-S-3136, type I fluid. 4/ | With temperature not higher than -65 F, conduct functional and leakage tests, using MIL-S-3136 type I fluid. 4/ |

1/ Each period shall follow immediately after the preceding one in the order noted.
2/ During periods of soaking in the test fluid, the component shall be maintained in such manner as to insure complete contact of all non-metallic parts with the fluid as would be expected under normal service conditions.
3/ Unless an increased test period is specified by the procuring activity.
4/ Leakage shall be kept within the allowable limits of this specification and the model specification or drawing, except for the first 15 minutes of the leakage tests conducted after the two dry periods.

Source: https://assist.dla.mil -- Downloaded: 2026-01-19T18:09Z
Check the source to verify that this is the current version before use.

# SPECIFICATION ANALYSIS SHEET
NAVWEPS FORM 4121/8 (8-62)

FORM APPROVED
BUDGET BUREAU NO. 48-R308

## INSTRUCTIONS

This sheet is to be filled out by Government or contractor personnel involved in the use of this specification in procurement of products for ultimate use by the activity shown on the reverse of this sheet.

This sheet is provided for obtaining information which will aid the activity shown in insuring that through use of this specification suitable products can be procured with a minimum amount of delay and at the least cost.

Comments and the return of this sheet will be appreciated.

Fold on dotted lines on reverse side, staple in corner, and send to the activity shown.

SPECIFICATION *(No. and abbrev. title)*

| ORGANIZATION | CITY | STATE |
|---|---|---|
| | | |

| CONTRACT NUMBER | QUANTITY OF ITEMS PROCURED | DOLLAR AMOUNT |
|---|---|---|
| | | $ |

MATERIAL PROCURED UNDER A DIRECT GOVERNMENT CONTRACT ☐ OR A SUBCONTRACT ☐

1. HAS ANY PART OF THIS SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
*(If so, give paragraph number and wording, and recommendations for correcting the deficiencies.)*

COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

. IS THE SPECIFICATION RESTRICTIVE?
☐ YES ☐ NO IF YES, IN WHAT WAY?

. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If additional papers are attached, send this form and papers in an envelope. This form is addressed for use in window envelope when appropriate.)*

| SUBMITTED BY *(Print name and title)* | DATE |
|---|---|
| | |

Source: https://assist.dla.mil -- Downloaded: 20261O-19T18:09Z
Check the source to verify that this is the current version before use.

Fold

DEPARTMENT OF THE NAVY
BUREAU OF NAVAL WEAPONS
WASHINGTON 25, D.C.

OFFICIAL BUSINESS

POSTAGE AND FEES PAID
NAVY DEPARTMENT

Chief, Bureau of Naval Weapons (RREN-5)

Navy Department

Washington 25, D.C.

Fold

Source: https://assist.dla.mil -- Downloaded: 2026-10-19T18:09Z
Check the source to verify that this is the current version before use.