A-A-50831
January 8, 1982

COMMERCIAL ITEM DESCRIPTION

SLIDE, MICROSCOPE

The General Service Administration has authorized the use of this commercial item description in preference to Federal Specification NNN-S-450B, for Type I, Size 1 slides.

This commercial item description covers plain 75 millimeter (mm.) by 25 mm. pre-cleaned glass microscope slides for use in laboratory procedures.

Salient characteristics:

Microscope slides shall be supplied with the faces finished in the following styles: Clear (i.e. without sectors for marking) and Frosted end (i.e. with sectors for marking).

Microscope slides shall be supplied pre-cleaned, free of detergent, and ready for use.

Microscope slides shall be fabricated from clear, transparent glass. The dimensions shall be 75 mm. long by 25 mm. width, with tolerances of ± 1.5 mm. for the length and + 1.0 for the width. The thickness of the slides shall be within either of the following ranges without overlapping: 0.8 mm. to 1.1 mm. or 1.1 to 1.3 mm.

The corners of the microscope slides shall be square, rounded, or bevelled.

To facilitate blood smear work, the surface of the short dimension at both ends of the microscope slides (i.e. 25 mm. working edges) shall be flat and perpendicular to the slide faces. As an alternate, the working edges shall be bevelled at approximately a 45 degree angle, with tip in the form of a knife edge, flat surface, or rounded surface. The short edges may also be partially bevelled and need not be further ground. In addition, the faces of the microscope slides shall be such that, when dipped into a beaker of distilled water, both surfaces immediately after removal exhibit a uniform dispersion of liquid and absence of smears.

Frosted end microscope slides shall be supplied with one or both faces at one end finished to give a marking sector(s) that allows legible markings to be applied with a ball point pen or a No. 2 pencil. Clear microscope slides shall be supplied without frosted ends.

In addition to blood smear work, frosted end microscope slides shall be suitable for mounting histological and cytological preparations. Tissue sections shall adhere to the microscope slides which have been prepared with proper binding agents such as gelatin or albumin, prior to affixing of the tissues.

Workmanship. The microscope slides shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing seventy-two slides, as specified, constitutes one unit.

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

A-A- 50019

Source of Information for Non-Government Association Documents.

National Motor Freight Traffic Association, Inc., Agent
National Motor Freight Classification
American Trucking Associations, Inc., Traffic Department,
1616 P Street, N.W., Washington, DC 20036

Uniform Classification Committee, Agent
Uniform Freight Classification
Uniform Classification Committee, Room 1106
222 South Riverside Plaza, Chicago, IL 60606

MILITARY INTERESTS:

Custodians

Army - GL
Air Force - 99

Review Activities

Army - MD

CIVIL AGENCY COORDINATING ACTIVITIES:

GSA - FSS

PREPARING ACTIVITY:

Army - GL

Project No. 7320-0712

GSA DC-01201779

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.