COMMERCIAL ITEM DESCRIPTION

INTRAVENOUS INJECTION SET, WITH METERING CHAMBER, WITHOUT NEEDLE, DISPOSABLE, STERILE

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-I-36708 and Amendment 4.

This commercial item description covers a sterile, disposable intravenous injection set with metering chamber suitable for use in the administration of parenteral solutions, without the use of an airway cannula.

Salient characteristics:

Shall be a sterile, non-pyrogenic, nontoxic, disposable type infusion set having a cap with piercing device, two drip chambers, an in-line, graduated metering chamber, a medicinal entry assembly, 52-56 inch length of tubing, a molded rubber part, two flow control clamps, air-inletting bacterial filter, a needle adapter, and end protectors.

Shall be suitable for accurate administration of parenteral solutions in controlled volumes from 10 ml to 100 ml at 60 drops per ml.

The set shall have a sterile, pyrogen-free fluid path completely sealed to maintain sterility.

The bacterial filter shall pass air at a rate exceeding 100 ml per minute under a pressure equal to a column of 16 inches of water.

The resealing rubber stopper on the medicinal entry shall be a sleeve style stopper molded from natural, non-oxidizing, low sulfur, wax-free rubber of 40-50° shore A Durometer with a tapered plug, diaphragm and sleeve design which, when assembled onto the Y-part, allows both outside and inside circumference contact to assure secure retention of the stopper upon withdrawal of the medicinal needle and syringe. The stopper diaphragm shall reseal without a falling drop within 10 seconds subsequent to puncture by a nonlubricated 20 gage needle when tested at a hydrostatic pressure of 15 pounds per square inch.

The intravenous injection sets shall meet the requirements for transfusion and infusion assemblies and the added requirements for Class II plastics as stated in the current USP. The needle adapter proximal to the needle shall be tapered to fit a standard Luer hub in accordance with ANSI Standard Z70.1.

Workmanship. The intravenous injection set shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Set (SE). One intravenous injection set, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following product, as of the date of this document, is purported by the following company to meet the salient characteristics of the item described herein:

Volutrole "20", Catalog No. 886-00
Source: Cutter Laboratories

NOTE: The following National Stock Number is covered by this document:

6515-00-088-8868

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARAING ACTIVITY:

DoD - MB

Agent:

DLA - DM

44595

Project No. 6515-2297

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.