A-A-50836
July 27, 1982

COMMERCIAL ITEM DESCRIPTION

BASIN, EMESIS, CORROSION-RESISTING STEEL

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification RR-B-120.

This commercial item description covers an emesis basin suitable for use in departments of military facilities and having the following characteristics.

Salient characteristics:

Shall be a conventional kidney-shaped corrosion-resisting steel emesis basin. Shall be fabricated seamlessly of Type 301, 302 or 304 corrosion-resisting steel and after forming shall have a minimum wall thickness of 0.020 inches. The basin shall be approximately 10 +1/4, -1/2 inches long (overall) and 5 + 5/8 inches wide across the rolled flange when measured along the center line of the basin. The entire top of the basin shall have a half-round, open bead edge. The inside joint of the side walls with flat bottom shall be rounded or curved to allow easy cleaning. The basin shall have a minimum capacity of 500 ml.

Finish. Shall have a satin finish or a No. 4 commercial finish as a minimum.

Workmanship. The basin shall be free from defects which detract from its appearance or impair its serviceability. All burrs and sharp edges shall be removed from the basin.

Unit. Each (EA). Each basin, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6530-00-770-9220

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians

Army - MD
Navy - MS
Air Force - 03

50358

PREPARING ACTIVITY:

DoD-MB

AGENT

DLA-DM

Project No 6530-1511

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.