A-A-50840A
24 March 1989
SUPERSEDING
A-A-50840
27 July, 1982

# COMMERCIAL ITEM DESCRIPTION

## CARRIAGE, PAIL

## Corrosion-Resisting Steel

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-C-126A which is cancelled.

This Commercial Item Description covers a one piece corrosion-resisting steel frame mounted on casters, suitable for use with a standard 12 quart pail.

Salient characteristics:

The pail carriage shall be fabricated of corrosion-resisting steel of the American Iron and Steel Institute (AISI) 300 series and suitable for surgical operating room use. It shall be mounted on electrically conductive casters and have electrically conductive rubber bumpers inside the frame and in a continuous ring around the frame. Electrical conductivity shall be according to National Fire Protection Association (NFPA) code 99, Health Care Facilities, as it refers to flammable anesthetics and conductivity.

The carriage shall be circular in design and shall consist of a welded corrosion-resisting steel frame mounted on either three or four smooth-rolling ball bearing swivel casters. The continuous rubber bumper shall be securely fastened to the frame. The rubber bumpers inside the frame shall support a standard 12 quart pail in a rattle-free manner when in use. The casters shall be either welded, bolted, or riveted to the carriage frame and shall have a minimum diameter of 2 inches. The carriage shall be mobile and free moving.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

All welding shall be in accordance with the latest practices recommended by the American Welding Society. No welds shall be cracked or broken. All welds shall be continuous, clean, and free of pitting, splatter, or discoloration. All butt welds shall be ground smooth to blend with adjacent surfaces. All fillet welds shall be properly sized and applied.

Metal parts of the carriage shall be free of cracks, tears, or ruptures. Any exposed or accessible edge, corner, or surface shall be free of burrs, sharpness, or sharp projections. Carriage finish shall be uniform and free of coarse grind marks and scratched areas.

The carriage shall be clean and free of oil, grease, dirt, or other contamination.

Workmanship. The pail carriage shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One pail carriage, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one pail carriage as specified in this description.

Defects.

Any carriage component that is missing or broken.

Carriage is not fabricated of AISI series 300 corrosion-resisting steel.

Caster diameter is less than 2 inches.

Rubber bumper does not fit tightly at all points.

Caster does not swivel freely.

Metal parts are not free of cracks, tears, or ruptures.

Any weld is cracked, broken, or not sound.

Any weld is not continuous, not clean, or not free of pitting, splatter, or discoloration.

Any exposed butt weld not ground smooth to blend with adjacent surfaces.

Any exposed or accessible edge, corner, or surface, that is not free of burrs, sharpness, or sharp projections.

Finish is not uniform and free of coarse grind marks and scratched areas.

Carriage is not clean and free of oil, grease, dirt, or other contamination.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for test's including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be one pail carriage.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Tests

Electrical Conductivity. The assembled carriage, complete with rubber bumpers and casters, shall be tested in accordance with National Fire Protection Association (NFPA) code 99, Health Care Facilities, as it applies to flammable anesthetics and conductivity.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the opetion of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-00-773-1450

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

A-A-50840A

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-2021

Location: CID/PAIL/CARRIAGE/H13-32



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*(See Instructions – Reverse Side)*

| 1 DOCUMENT NUMBER | 2 DOCUMENT TITLE |
|---|---|
| A-A-50840A | CARRIAGE, PAIL, Corrosion-Resisting Steel |

| 3a. NAME OF SUBMITTING ORGANIZATION | 4 TYPE OF ORGANIZATION *(Mark one)* |
|---|---|
| | ☐ VENDOR |
| b ADDRESS *(Street City State ZIP Code)* | ☐ USER |
| | ☐ MANUFACTURER |
| | ☐ OTHER *(Specify)* |

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

| 7a NAME OF SUBMITTER *(Last First MI)* – Optional | b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional |
|---|---|
| c MAILING ADDRESS *(Street City State ZIP Code)* – Optional | 8 DATE OF SUBMISSION *(YYMMDD)* |

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2020-10-26T08:47Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, were too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

*NOTE* This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

*(Fold along this line)*

*(Fold along this line)*

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.