A-A-50841A
29 July 1988
SUPERSEDING
A-A-50841
27, July 1982

COMMERCIAL ITEM DESCRIPTION

ENEMA ADMINISTRATION SET

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-S-36459 which is cancelled.

This Commercial Item Description covers a completely assembled disposable enema administration set suitable for the introduction of barium sulfate for radiographic examination of the colon.

Salient characteristics:

The enema administration set shall consist of:

1. A clear, dull finish, leakproof, polyvinyl calendered sheet, plastic bag.
2. A minimum of 72 inches of transparent sterilizable, non-toxic vinyl plastic tubing.
3. A rectal tip, and a metal or plastic shut-off clamp.

The plastic tubing shall have a minimum wall thickness of 0.006 inch with an approximate OD of 5/16 inch. It shall be flexible and pliable, but shall be rigid enough to prevent collapse during normal use.

The bag shall have a minimum capacity of 2500 ml with graduations at 500 ml intervals up to 2500 ml. The graduations shall be embossed on the surface of the bag. The bag shall be provided with a filling spout and an outlet suitable for introducing a solution of barium into the colon for radiographic examination. The filling spout shall have a double tie-wrap closure, snap-cap closure, or other suitable secure closure. The spout shall be flexible and pliable but yet rigid enough to maintain its shape when being grasped to be filled.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

The tubing shall be equipped with a metal or plastic clamp, of the clip or slide type, attached to the upper part of the tubing. Construction of clamp shall be such as to permit single-handed, positive activation or cessation of flow of fluids in tubing without distortion or breakage of the clamp or incision of tubing. If made of metal, the clamp shall be appropriately plated if necessary to resist corrosion.

The tubing shall have an adult rectal tip made from polyethylene. Its surface shall be smoothly polished and shall have a standard bore.

The plastic components of the enema set shall contain not more than 5 parts per million (ppm) of lead.

The enema set shall be free from sharp edges and free from foreign material inside the bag and tubing.

The outer surfaces of the enema set shall be free of oil, grease, dirt, or other contamination.

The enema set shall be provided with instructions for use printed on the exterior of the packaging.

Workmanship. The enema administration set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Set (SE). One package containing forty-eight enema administration sets, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation witnout the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one enema administration set as specified in this description.

Defects.

Enema set broken, or missing components.

Bag capacity is less than 2500 ml.

Bag has no graduations.

Graduations are present, but not embossed on surface of bag.

Graduations are at incorrect intervals.

Plastic tubing is less than 72 inches long.

Wall thickness of tubing is less than 0.006 inch.

Enema bag and tubing is not transparent.

Foreign material inside enema bag or tubing.

Enema set not free from sharp edges.

Enema set outer surfaces not free from oil, grease, dirt, or other contamination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests including dimensional test shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one enema administration set.

Tests

Leakage Test

The plastic bag with the shut-off clamp tightly closed, shall be filled with water at 100°F and suspended in the inverted position for a period of 1 hour. Any evidence of leakage shall constitute a defect.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-016-7693

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-2004

PGC 70072

Location: CID/ENEMA/ADM/H13-32

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

See Instructions - Reverse Side

1 DOCUMENT NUMBER
4-30841A

2 DOCUMENT TITLE
ENEMA ADMINISTRATION SET

3a NAME OF SUBMITTING ORGANIZATION

b ADDRESS (Street, City, State, ZIP Code)

4 TYPE OF ORGANIZATION (Mark one)

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER (Specify)

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last, First, MI) – Optional

b WORK TELEPHONE NUMBER (Include Area Code) – Optional

c MAILING ADDRESS (Street, City, State, ZIP Code) – Optional

8 DATE OF SUBMISSION (YYMMDD)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

Fold along this line

Fold along this line

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT [illegible] WASHINGTON D.C.

[illegible]
Defense Personnel Support Center
ATTN: [illegible]
[illegible]
[illegible]
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.