A-A-50842
July 27, 1982

COMMERCIAL ITEM DESCRIPTION

URINAL, MALE PATIENT, GLASS

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification GG-U-675

This commercial item description covers the conventional design for glass urinal suitable for use by male patients.

Salient characteristics:

The urinal shall be fabricated from a good quality clear glass. It shall have an integral handle, and shall be free standing and stable when standing in the upright position. It shall have a minimum capacity of either 28 ounces or 825 milliliters and shall have appropriate and evenly spaced graduations up to and including the stated capacity. The contents shall not spill from the urinal when resting in either the upright or horizontal position. The urinal shall have a finished surface appropriate for patient contact.

Workmanship. Urinals shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One urinal, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

For information only. The following product, as of the date of this document, is purported by the following companies to meet the salient characteristics of the item described herein:

Mercer Glass Works, Inc. - Cat. No. 1095
General Medical - Cat. No. 5117 - 00110

NOTE: The following National Stock Number is covered by this document:

6530-00-794-4094

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

50690

Project No. 6530-1542

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.