A-A-50843A
31 March 1986

# COMMERCIAL ITEM DESCRIPTION

## FLASHLIGHT, (Patient Examining)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-F-36678B.

This commercial item description covers a disposable flashlight suitable for use in the examination of medical and dental patients.

Salient characteristics:

The disposable flashlight shall consist of a case, two batteries, one lamp, pocket clip, and other internal components necessary for proper operation. The case shall be a thermoplastic tube with a wall thickness not less than 0.023 inch. The length of the case shall not exceed 5 inches and the diameter shall not exceed 5/8 inch. The ends of the case shall be rolled or otherwise formed to secure the lamp, batteries, and other components in proper alignment. The color of the case shall be white or light blue. The flashlight shall be powered by two batteries. The batteries shall be Leclanche cylindrical carbon zinc dry cell batteries. The batteries shall be size AAA, general purpose flashlight cells conforming to the applicable paragraphs of American National Standards Institute (ANSI) Standard C18.1. The cells may be unjacketed. The lamp shall be a commercial number 222 miniature prefocused incandescent lamp. The pocket clip shall be fabricated of a suitable metal appropriately finished to resist corrosion. Pressing the pocket clip shall actuate the flashlight and cause the lamp to light. The internal components shall hold the batteries and lamp in proper alignment and maintain good electrical contact. The batteries shall be new. Not more than seventy-five (75) days shall have elapsed from date of manufacture of the batteries to the date of delivery of the flashlights to the Government.

Workmanship. The flashlights shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing twelve flashlights, as specified, constitutes one unit. Flashlights shall be packaged in such a manner as to prevent actuation of the contact switch within the unit package.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

AMSC N/A FSC 6230

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:47Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following products, as of the date of this document, are purported by the following companies to meet the salient characteristics of the item described herein:

Concept, Incorporated
Wonder Corporation of America
Medical Products Development, Inc.

NOTE: The following National Stock Number is covered by this document:

6230-00-125-5528

Copies of ANSI standard C18.1 may be obtained from the Amercian National Standards Institute, Inc., 1430 Broadway, New York, NY 10018.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PGC 41190

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6230-0307
Location: CID H1-1-42

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.