A-A-50851A
12 August 1986
SUPERCEDING
AA50851, 23 AUG 1982

COMMERCIAL ITEM DESCRIPTION

NEBULIZER, MEDICINAL

The General Services Administration has authorized the use of this commercial item description in preference to commercial item description #AA50851, dated 23 August 1982.

This commercial item description covers a a disposable manifold and nebulizer suitable for use with Intermittent Positive Pressure Breathing (IPPB) equipment.

Salient characteristics:

The nebulizer shall essentially consist of the following items:

a. Manifold
b. Nebulizer
c. Flexible corrugated mainstream tubing at least 39 inches (99.06 cm) long with two supply lines
d. Flexible corrugated tube approximately 6 inches (15.24 cm) long
e. Tubing adapters
f. Mouth piece

Materials. All components shall be fabricated of a suitable plastic material, except for the diaphragm and o-ring, which shall be fabricated of a suitable rubber or latex material.

Liquid level. The nebulizer shall have a liquid level capacity of at least 10 ml.

Performance. The nebulizer shall produce a fine mist with particles which primarily fall in the size range of 0 to 10 microns. The device shall be suitable for use with either single or double supply lines.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Additional requirements. The nebulizer shall comply with the requirements of ANSI Standard Z79.9 dated 1979. The following paragraphs are applicable to the nebulizer:

3.1.1	Fluid Output
3.1.3	Accuracy of Operator Controls and Indicators
3.2.1	Liquid Level
3.2.4	Test for Biocompatability
3.3.1.1	Liquid Level
3.3.1.2	Pressure Relief
3.3.1.3	Flow Direction
3.3.3	Instructions
3.3.4	Disposability
3.3.5	Intended Use
3.3.8	Flow Resistance
3.3.9	Internal Compliance
3.3.10	Internal Leakage
3.4.1	Connections
3.4.3.1	Tilting
3.4.3.2	Reverse Connections
3.4.4	Legibility of Markings

Workmanship. The nebulizer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package, containing fifty nebulizers, as specified, constitutes one unit of issue.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Testing shall include but not be limited to the following applicable portions of ANSI Standard Z79.9 dated 1979:

4.1	Performance
4.2.2	Overpressure Relief
4.24	Biocompatability
4.3	Marking and Labeling
4.4	Human Engineering

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one nebulizer.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-00-528-4092

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS

PGC 46822

Project No. 6515-3584

Location: CID/NEBUL-MED/H10

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-50851A	NEBULIZER, MEDICINAL

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*: _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.