A-A-50856
October 29, 1982

COMMERCIAL ITEM DESCRIPTION

WATER BATH, ELECTRIC, Serological

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-W-37045(DSA-DM).

This commercial item description covers an electrically heated, thermostatically controlled, five rack capacity water bath for use in serological testing.

Salient characteristics:

Shall be a five rack (removable type) serological water bath with a removable gable cover, suitable for use in serological laboratory procedures.

Shall be operable on nominal 115 volt, 50-60 Hz, AC supply lines with a 1,000 watt capacity.

Shall include the following temperature ranges:

1. Slightly above ambient to 60 degrees Centrigrade (C) when the water bath is uncovered.
2. Slightly above ambient to temperature of boiling water when the water bath is covered.

Shall use a mercury-in-glass thermometer to indicate the temperature of the water bath.

Shall be fitted with a temperature controller that is capable of maintaining a selected water bath temperature constant within ± 1/2 degree C (i.e. - highest temperature not exceeding lowest temperature by more than 1 degree C) when the water bath is covered. A safety thermostat shall also be supplied that provides an over-temperature limit control in the event that the temperature controller should fail.

Shall have a double wall construction with insulation material between walls including the bottom of the water chamber. The insulation material shall be moisture-resisting and nonpacking. The exterior walls and water chamber shall be fabricated of corrosion-resisting metals, and shall have nominal dimension (inches) that, as a minimum, include the following:

1. Water chamber: 12-3/4 long by 11-1/2 wide by 5 deep.
2. Outside housing: 15 long by 16 wide by 8 deep.

Shall include the following electrical controls:

1. An "on-off" line type switch to control the electrical power to the heating elements.
2. A pilot light for indicating when heating elements are energized.
3. Thermostat dial selectors with reference markings for water bath temperature settings of the temperature controller and safety thermostat.
4. A 3-wire power cord that terminates with a 3-prong plug that is suitable for use with receptacles specified in American National Standards Institute (ANSI) C73.11 (Dimensions of Caps, Plugs and receptacles, 125 volts, 12 Amperes, 2 Pole, 3 Wire, Grounding Type, General Purpose).

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:48Z
Check the source to verify that this is the current version before use.

Shall be supplied with the following:

1. A single gable cover.
2. A detachable partial immersion mercury-in-glass thermometer that includes a range of 0 to 100 degree C.
3. A thermometer holder for the mercury-in-glass type thermometer that is in the form of a bracket attached to the water chamber or a grommet in the gable cover.
4. Five (5) each - 20 tube capacity corrosion resisting steel test tube racks that accommodate 13 millimeter diameter test tubes, and include two rows of 10 holes with numbering from 1 to 10.

Workmanship. The water bath shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One water bath, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following product, as of the date of this document, is purported by the following companies to meet the salient characteristics of the item described herein:

Boekel Industries	Cat. No. 148003 Water Bath with Cat. No. 148003 gable cover and 5 each Cat. No. 1492 test tube racks.
Lab-Line Instruments	Cat. No. 18005 Imperial III Water Bath with Cat. No. 18005-2 gable cover and 5 each Cat No. 56100 test tube racks.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:48Z
Check the source to verify that this is the current version before use.

The following National Stock Number is covered by this document: 6640-00-765-0621

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

55866

Project No. 6640-1161

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:48Z
Check the source to verify that this is the current version before use.