A-A-50861A
27 October 1989

COMMERCIAL ITEM DESCRIPTION

COVERS, DENTAL BRACKET TABLE, Paper

The General Services Administration has authorized the use of this Commerical Item Description.

Salient characteristics:

This commercial item description covers disposable, paper covers suitable for use on bracket trays of dental operating units. The covers shall be rectangular in shape and of the following sizes:

Size 1: 17-1/2 by 11 inches

Size 2: 12 by 8-1/4 inches

Material: Covers shall be fabricated of fully bleached sulfite paper, in accordance with the physical properties specified herein.

Physical properties:

Property	Minimum Requirement	TAPPI* Test Method
Brightness (white only)	75%	T452
Weight	20.0 lb (17x22-500)	T410
Bursting strength	25 points, average	T403
Thickness	0.0040 inch	T411

* Technical Association of the Pulp and Paper Industry (TAPPI).

Color. The covers shall be white or any light pastel shade.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Dimensions. The size 1 cover shall be a rectangle measuring 17-1/2 + 1/4, -1/8 inches in length and 11 ± 1/8 inches in width, with rounded corners each having a 1-3/4 ± 1/8 inch radius. The size 2 cover shall be a rectangle measuring 12 + 3/8 - 1/8 inches in length and 8-1/4 + 1/4, -1/8 inches in width, with rounded corners of 1-3/16 ± 1/8 inches in radius. The covers shall be flat on a flat surface.

Finish. The covers shall have a plain or embossed finish.

Workmanship. Covers shall be free from defects which may detract from their appearance or impair their serviceability. The covers shall be free of cuts, holes, and wrinkles, and shall have neatly trimmed edges. The paper shall be free from shives, ground-wood, and unbleached fibers. The covers shall be clean and without objectionable odor.

Unit. Package (PG). One package containing 1000 covers of one size and one color only, as specified, constitutes one unit.

Unit package. Each unit package shall be plastic shrink wrapped.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the used of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, at inspection level S-2. The unit of product for examination purposes shall be one unit of issue from which 10 covers shall be examined for defects, except for count. Examination for defects shall include but shall not be limited to those listed below:

Defects.

Shape and size not as specified.
Color not as specified; color not uniform.
Covers not clean, not free of dirt or grease.
Covers wrinkled or creased.
Finish not uniform.
Workmanship not as specified.
Corners not rounded and clean cut.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and of regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Number	Item Identification
6520-00-139-4344	COVER, DENTAL, BRACKET TABLE, Rectangular, 17-1/2 by 11 inches
6520-00-241-2431	COVER, DENTAL, BRACKET TABLE, Rectangular, 12 by 8-1/4 inches

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

05184

PROJECT NO. 6520-2316
LOCATION: NCR - DISK 8

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.