A-A-50862A
12 June 1985
SUPERSEDING
A-A-50862
26 January 1983

COMMERCIAL ITEM DESCRIPTION

MIXING SLAB, DENTAL

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification DD-M-00440a(DSA-DM) and Federal Specification DD-M-440.

This commercial item description covers a glass slab suitable for use in preparing dental cements, restorative, and other materials requiring spatulation and shall have the following salient characteristics:

Salient characteristics:

The mixing slab shall be fabricated from clear, colorless, plate glass, except the appearance of a green tint is acceptable. Shall be rectangular in shape. Dimensions shall be 6 ± 3/16 inches long, 3 ± 3/16 inches wide, and 1/2 to 7/8 inch thick. All surfaces shall be flat and smooth. Working surfaces shall be parallel and shall be highly polished. Side faces shall be polished or shall have a fine grind finish. All edges shall be beveled. The working surfaces shall be free from chips, cracks, grind marks, scratches, and open gaseous inclusions. The mixing slab shall be suitable for mixing dental cements and other manipulatives on both the upper and lower mixing surfaces (6 by 3 inch faces).

Material. Material shall be of the highest quality throughout and suitable in all respects for the purpose intended.

Workmanship. The mixing slab shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One mixing slab, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one mixing slab. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract..

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

The following National Stock Number is covered by this document:

6520-00-556-2000

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6520-1813

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.