A-A-50863
March 4, 1983

COMMERCIAL ITEM DESCRIPTION

# URINAL, MALE PATIENT, PLASTIC

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification GG-U-675.

This commercial item description covers a single patient use plastic male urinal.

Salient characteristics:

The urinal shall be fabricated from a translucent, chemically resistant, semirigid plastic material. The plastic may be a blend of virgin material and not more than 30 percent reworked material.

The urinal shall consist of a container, an integral handle, and an attached tight-fitting lid. It shall have a nominal capacity of 1000 milliliters (ml). The urinal shall be capable of standing in an upright position and have the ability to rest on its side with the opening angled upwards. No contents shall spill from the urinal when filled to capacity while resting in a horizontal position.

Workmanship. Urinals shall be free from blooming, chalking, crazing, degradation, weld-lines, or other defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing fifty urinals, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following products, as of the date of this document, are purported by the following companies to meet the salient characteristics of the item(s) described herein:

| | |
|---|---|
| Vollrath Company | Cat. No. 04090 |
| Bemis Health Care | Cat. No. SP-93 |

NOTE: The following National Stock Number is covered by this document:

6530-01-042-2485

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

Project No. 6530-1534

50691

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.