A-A-50869
June 1, 1983

COMMERCIAL ITEM DESCRIPTION

PIPET, SEROLOGICAL (Disposable, Sterile)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-P-36668B.

This commercial item description covers a disposable, sterile, plugged, 1 milliliter capacity serological glass pipet, suitable for use in laboratory procedures.

Salient characteristics:

Shall be a disposable, sterile, plugged, 1 milliliter (ml) capacity, borosilicate glass, serological pipet that is calibrated "to deliver".

Shall be in accordance with ASTM E-714 (Disposable Glass Serological Pipets), as applicable to 1.0 ml. capacity and type I glass pipets. The mouthpiece end shall include a plug fabricated from cotton or synthetic fiber.

Shall be supplied in a sterile condition. The method and efficacy of sterilization shall be acceptable to the Food and Drug Administration (FDA).

The volumetric accuracy shall meet the requirements set forth in ASTM E-714. A certificate of quality may be furnished in lieu of actual performance of individual testing provided the lot/sample identities can be maintained. The certificate shall contain the lot number, lot size, sample size and test method.

Workmanship. The pipet shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package of twenty five pipets, as specified, constitutes one unit. Twenty five pipets shall be packaged in a suitable transparent container which shall maintain sterility until manually opened. As an alternate, each pipet shall be packaged in a sealed peel open type container which shall maintain sterility until manually opened.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians

Army - MD
Navy - MS
Air Force - 03

56085

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

Project No. 6640-1150

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.