A-A-50875B
5 October 1990
SUPERSEDING
A-A-50875A
22 December 1986

# COMMERCIAL ITEM DESCRIPTION

## TUBE, BLOOD COLLECTING (Vacuum)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers vacuum blood collecting tubes of the following types, aspiration capacities, grades of additives and tube sizes.

| | | |
|---|---|---|
| Type I | - | With anticoagulant |
| Type II - | - | Silicon coated without anticoagulant |
| Aspiration capacity | - | 3,5,7,15 milliliter (ml) |
| Grade D additive | - | Sodium Fluoride - 30 milligrams (mg) |
| Grade F additive | - | Sodium Citrate (buffered or non-buffered) - 0.5 ml, 3.8% solution |
| Grade G additives | - | Tri-potassium EDTA-0.06 ml, 7.5% solution (or EDTA (K3) - 0.06 ml, 7.5% solution). |
| | - | Thimerosal 50 ppm (or Potassium Sorbate 0.012 mg, with an antimycotic agent) |
| Grade H additives | - | EDTA (K3) - 0.07 ml, 15% solution (or Tri-potassium EDTA - 0.07 ml, 15% solution) |
| | - | Thimerosal 50 ppm (or Potassium Sorbate 0.012 to 0.014 mg, with an antimycotic agent) |
| Grade I additive | - | Sodium Heparin - 143 USP units |
| Tube sizes (diameter X length) | - | 10 25 X 64, 13 X 75, 13 X 100, 16 X 75, 16 X 125 millimeters (mm) |

Commercially designated 4.5 ml aspiration capacities may be offered as an alternate to the specification designated 5 ml aspiration capacity.

The Thimerosal and Potassium Sorbate antimicrobal agents shall be optional with the grade G and H additives if the anticoagulant is supplied in a sterile condition.

Commercially designated 10.25 X 65 and 16 X 127 mm tube sizes may be offered as alternates to the specification designated 10 25 X 64 and 16 X 125 mm tube sizes respectively.

AMSC N/A FSC 6630

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Salient Characteristics:

Shall be non-sterile or sterile vacuum blood collecting tubes suitable for collecting, containing and transporting blood samples

Shall comply with the applicable requirements of NCCLS (National Committee for Clinical Standards) Approved Standard ASH-1 (Standard for Evacuated Tubes for Blood Specimen Collection). Commercial tolerances on the additives for type I tubes shall be acceptable.

Shall consist of evacuated transparent glass tubes with diaphragm type self sealing rubber stoppers that form airtight seals. The type I tubes shall be supplied without a non-wettable lining on the inside surfaces.

Shall be affixed with a label that includes a blank space for use by hospital personnel.

Shall be suitable for use with a standard double-ended blood-collecting needle/holder, described in the ASH-1 standard, that contains a bevel edge needle in its open end When the stopper ends of the tubes are inserted into the needle/holder to the depth even with the needle/holder's reference mark, the needle point shall be imbedded in the stoppers without puncturing the stoppers.

The stopper of each tube shall be color coded in accordance with table I

Table I. Stopper Color Coding

| Stopper color | Type tube | Grade tube |
|---|---|---|
| Red | II | --- |
| Grey | I | D |
| Blue | I | F |
| Lavender | I | G,H |
| Green | I | I |

Workmanship. The blood collecting tubes shall be free from defects which detract from their appearance or impair their serviceability

Unit Package (PG) One package containing 100 or 1000 tubes, as specified, constitutes one unit.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted on the blood collecting tubes in accordance with commercial practice to determine compliance with specification requirements. The sampling for inspections and tests shall be conducted in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

NOTE: The following National Stock Numbers are covered by this document.

| National Stock No. | Item Identification | Specification Designation Type | Grade | Tu Sz |
|---|---|---|---|---|
| 6630-00-145-1137 | TUBE, BLOOD COLLECTING, Vacuum, 7 ml, without Anticoagulant, 100s | II | - | 13X100 |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

| National Stock No. | Item Identification | Specification Designation Type | Grade | Tu | Sz |
|---|---|---|---|---|---|
| 6630-00-145-1143 | TUBE, BLOOD COLLECTING, Vacuum, 15 ml, without Anticoaqulant, 100s | II | - | | 16X125 |
| 6630-00-145-1534 | TUBE, BLOOD COLLECTING, Vacuum, 7 ml, with Sodium Fluoride, 100s | I | D | | 13X100 |
| 6630-00-250-4264 | TUBE, BLOOD COLLECTING, Vacuum, 5 ml, with Sodium Citrate, 100s | I | F | | 13X75 |
| 6630-01-047-6716 | TUBE, BLOOD COLLECTING, Vacuum, 7 ml, with Sodium Heparin, 1000s | I | I | | 13X100 |
| 6630-01-049-1842 | TUBE, BLOOD COLLECTING, Vacuum, 7 ml, with Sodium Heparin, 100s | I | I | | 13X100 |
| 6630-01-069-6231 | TUBE, BLOOD COLLECTING, Vacuum, 7 ml, EDTA (K3) 0.07 ml, 15% Solution, 1000s | I | H | | 16X75 |
| 6630-01-119-8575 | TUBE, BLOOD COLLECTING, Vacuum, 7 ml, EDTA (K3), 0.07 ml, 15% solution, 100s | I | H | | 13X100 |
| 6630-01-221-0330 | TUBE, BLOOD COLLECTING,- Vacuum, 3 ml, with Tri-Potassium EDTA, 100s | I | G | 10 | 25X64 |
| 6630-01-221-3320 | TUBE, BLOOD COLLECTING Vacuum, 3 ml, without Anticoagulant, 100s | II | - | 10 | 25X64 |

Ordering data. (Expiration dating required, intermediate/ exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

MILITARY INTERESTS: PREPARING ACTIVITY:

Custodians: DoD-MB



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

A-A-50875B

Preparing Activity:
DOD - MB

Army - MD
Navy - MS
Air Force - 03

Agent:
DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No: 6630-0509

Location: ENABLE/TUBECID2/S31



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*(See Instructions – Reverse Side)*

| 1. DOCUMENT NUMBER | 2. DOCUMENT TITLE |
|---|---|
| A-A-50875B | TUBE, BLOOD COLLECTING (Vacuum) |

| 3a. NAME OF SUBMITTING ORGANIZATION | 4. TYPE OF ORGANIZATION *(Mark one)* |
|---|---|
| | ☐ VENDOR |
| | ☐ USER |
| b. ADDRESS *(Street, City, State, ZIP Code)* | ☐ MANUFACTURER |
| | ☐ OTHER *(Specify)*: |

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation

6. REMARKS

| 7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional | b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional |
|---|---|
| c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional | 8. DATE OF SUBMISSION *(YYMMDD)* |

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

**INSTRUCTIONS:** In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

*NOTE:* This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

*(Fold along this line)*

*(Fold along this line)*

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.