MIL-L-16446A
16 May 1966
SUPERSEDING
MIL-L-16446 (BuMed)
15 July 1951

MILITARY SPECIFICATION

LITTER, FOLDING, FOLDING POLE

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers one type of folding, aluminum pole, folding litter.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

V-T-276	Thread, Cotton.
QQ-A-200/3	Aluminum Alloy Bar, Rod, Shapes, and Tube, Extruded, 2024.
QQ-A-225/5	Aluminum Alloy Bar, Rod, and Wire, Rolled, Drawn, or Cold Finished, 2017.
QQ-A-225/6	Aluminum Alloy Bar, Rod, and Wire, Rolled, Drawn, or Cold Finished, 2024.
QQ-A-225/8	Aluminum Alloy Bar, Rod, Wire, and Special Shapes, Rolled, Drawn, or Cold Finished, 6061.
QQ-B-613	Brass, Leaded and Non-Leaded, Plate, Rolled Bar, Sheet and Strip.
QQ-S-763	Steel Bars, Shapes, and Forgings - Corrosion Resisting.
TT-E-527	Enamel, Alkyd, Lustreless.
TT-P-659	Primer-Surfacer, Synthetic, Tints and White (for Metal and Wood Surfaces).
WW-T-700/6	Tube, Aluminum Alloy, Drawn, Seamless, 6061 and 6062.
CCC-C-419	Cloth, Duck, Cotton, Unbleached, Plied-Yarns, (Army and Numbered).
CCC-D-950	Dyeing and Aftertreating Processes for Cotton Cloths.
CCC-T-191	Textile Test Methods.
PPP-B-576	Box, Wood, Cleated, Veneer, Paper Overlaid.
PPP-B-601	Box, Wood, Cleated-Plywood.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

Military

MIL-F-495	Finish, Chemical, Black, for Copper Alloys.
MIL-C-496	Clips, End, Strap.
MIL-W-530	Webbing, Textile, Cotton, General Purpose, Natural or in Colors.
MIL-B-543	Buckles, Tongueless and Web Strap.
MIL-T-3530	Thread and Twine, Mildew Resistant or Water Repellent Treated.
MIL-C-5541	Chemicals Films and Chemical Film Materials for Aluminum and Aluminum Alloys.
MIL-P-8585	Primer Coating, Zinc Chromate, Low Moisture-Sensitivity.
MIL-L-10547	Liners, Case, and Sheet, Overwrap, Water-Vaporproof or Waterproof, Flexible.
MIL-S-13518	Sealer, Surface, Wood Preservative.

STANDARDS

Federal

FED-STD-151	Metals, Test Methods.
FED-STD-595	Colors.
FED-STD-751	Stitches, Seams, and Stitchings.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER

1769	Litter, Folding, Folding Pole, Aluminum Pole.

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publication. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids shall apply.

COLOR ASSOCIATION OF THE UNITED STATES, INC.

United States Army Standard Color Card for the Official Standardized Shades of Sewing Threads.

(Application for copies of the color cards should be addressed to the Color Association of the United States, Inc., 200 Madison Avenue, New York, N.Y. 10016).

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material.

3.1.1 Litter poles. Litter poles shall be made of aluminum-alloy tubing 6061-T6, 1-1/2 inches outside diameter, a wall thickness of 0.120 inch (No. 11 Birmingham Wire Gage), conforming to WW-T-700/6.

3.1.2 Stirrups. Stirrups shall be formed from 6061-T6 aluminum, conforming to QQ-A-225/8.

3.1.3 Spreader bar, coupler, bolt block, plug insert, and saddle. Spreader bar, coupler, bolt block, plug insert, and saddle shall be formed from 2024-T4 aluminum, conforming to QQ-A-200/3 or QQ-A-225/6.

3.1.4 Stop lug. Stop lug shall be formed from 2017-T4 aluminum, conforming to QQ-A-225/5.

3.1.5 Bolt, rivets, and drive pin. Bolt, rivets, and drive pin shall be formed from corrosion-resisting steel, comforming to class 302, or 304, of QQ-S-763.

3.1.6 Nut for bolt. Nut for bolt shall be elastic stop, self-locking type, made of carbon steel, 18 thread, with fiber insert an integral part of the nut.

3.1.7 Bushing. Bushing shall be made of brass, suitable for the purpose intended.

3.1.8 Litter cover. Litter covers shall be made of cotton duck.

3.1.8.1 Cotton duck. The cotton duck, before finishing, shall be type I, No. 10, hard texture, conforming to CCC-C-419. The unfinished and finished duck shall be tested as specified in 4.4.2.

3.1.8.1.1 Color and treatment. Cotton duck shall be vat-dyed olive drab No. 7, conforming to type 1, class B of CCC-D-950, except that the cloth shall be made mildew resistant by treatment limited to one of the following mildew inhibitors:

(a) 2,2'methylene-bis (4-chlorophenol).
(b) Copper-8-quinolinolate.

3.1.8.2 Thread. The thread shall be cotton conforming to type IIIA and IIIB, 3 ply, ticket No. 10/3 of V-T-276. All references to approved shade standard, standard sample, standard shade, and the paragraph titled certification shall not apply. The untreated and treated thread shall be tested as specified in 4.4.3.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

3.1.8.2.1 Color and treatment. The thread shall be dyed to match olive-drab shade S-1 (Cable No. 66022) of the United States Army Standard Color Card for the Official Standardized Shades of Sewing Threads. The cotton thread shall be mildew resistant in accordance with type 1, class 1, of MIL-T-3530.

3.1.9 Securing straps. Securing straps shall be made of cotton webbing, type III, class 4 or 7, one inch wide, vat dyed olive drab No. 7, conforming to MIL-W-530. If copper-8-quinolinolate is used as the mildew inhibitor, the maximum percent copper-8-quinolinolate shall be 1.28. The straps shall be tested as specified in 4.4.4.

3.1.10 End clips. End clips shall be type 1, brass, one inch, conforming to MIL-C-496, except the material shall conform to QQ-B-613. The finish shall be as specified in MIL-F-495.

3.1.11 Buckles, tongueless, 2-bar. Buckles for securing straps shall conform to type II, style 4, class 1, one inch, conforming to MIL-B-543.

3.1.12 Handles. Handles shall be made from thoroughly seasoned, kiln dried, hard (sugar) maple, ash, birch, beech, oak, true hickory, or pecan hickory free from knots and checks.

3.1.13 Binding. Binding for the openings in the litter cover shall be cotton tape conforming to type 1, class 4 or 7, of MIL-W-530. The tape shall be 3/4 inch wide, vat dyed olive drab No. 7, and shall be tested as specified in 4.4.4.

3.2 Construction. Construction shall conform to Defense Personnel Support Center (DPSC) Drawing No. 1769. Each litter shall be furnished assembled with two complete spreader bar and stirrup assemblies, two poles with handles, canvas cover, and two securing straps. The complete unit shall pass the tests specified in 4.4.1.

3.2.1 Spreader bar, stirrup assemblies and poles. Spreader bar, stirrup assemblies, and poles shall be formed as shown on the drawing. All sharp edges shall be removed, and corners shall be slightly rounded.

3.2.2 Handles. Handles shall be turned from hardwood specified in 3.1.12. The end fitted to the inside circumference of the pole shall be smooth, true, and not tapered, except for the 1/8 inch for starting into the pole, which shall be tapered. The handle shank shall be driven the entire length into place in the litter pole, and shall be secured in position with a drive pin passing through one thickness of the tubing. The pin shall be set so that it shall be flush with the outside of the pole.

3.2.2.1 Drive pin. Drive pin shall be a corrosion-resisting steel headless nail as shown on drawing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

3.2.3 Canvas cover. Canvas cover shall be fabricated from one piece of cotton duck specified in 3.1.8.1, and shall conform in all details to the drawing. The cover shall be so constructed that when placed on the litter, the warp shall run parallel to the long axis of the litter. Turnunders and seams shall be neatly made. No loose or raveled edges shall be exposed. When bought separately, covers shall be furnished without bolt holes.

3.2.3.1 Stitching. All stitching shall be machine stitching conforming to stitch type 301 of FED-STD-751. There shall be a minimum of eight stitches per inch. The type IIIB thread shall be used for needle or top thread, and type IIIA thread for bobbin or bottom thread in the machine. When machine stitching is not a continuous thread, it shall be backstitched not less than one inch at each break. All stitching shall be backstitched at the ends to prevent raveling, except where ends are turned under in a hem or held down by other stitching. Thread tension shall be properly maintained so that the lock will be embedded in the center of the materials sewed. All edges which are not selvaged shall be hemmed with stitching type EFb-1. All seams shall be straight and evenly spaced.

3.2.4 Securing straps. Securing straps shall be clamped between the bolt block and the litter cover. A hole shall be punched in the strap as shown in the drawing for the induction of the bolt shaft. The buckles shall be fastened to the short end of the strap by stitching (see drawing). The free end of the strap shall be provided with a ball-type end clip, so constructed that it shall pass freely through the buckle. The end clip shall be firmly attached without cutting the webbing. Webbing shall be inserted the full depth and lie flat at the point where it enters. The other free end of the strap shall be impregnated with a hardening material to prevent raveling. When bought separately, securing straps shall be furnished without bolt holes.

3.3 Finish.

3.3.1 Degreasing. All surfaces shall be thoroughly cleaned and chemically degreased prior to chemical treatment or plating.

3.3.2 Aluminum parts. All aluminum surfaces shall be finished with a chemical film conforming to MIL-C-5541.

3.3.3 All steel parts, other than corrosion-resisting steel parts, shall be zinc coated.

3.3.4 Primer. After degreasing, chemical treatment, or coating, a primer coat shall be applied to all metal parts other than corrosion-resisting steel parts. The primer shall be zinc-chromate conforming to MIL-P-8585. The primer coat shall form a grounding bond and base for the finishing enamel.

3.3.5 Enamel. The primer coat shall be followed by one coat of lustreless olive-drab enamel, conforming to TT-E-527. The resultant finish shall be hard, smooth, and highly resistant to abrasion, chipping, and cracking.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

3.3.5.1 Color. The color of the final finish shall be lustreless olive-drab conforming to color X34087 (see 6.2) of FED-STD-595.

3.3.6 Handles. Prior to finishing, the handles shall be sanded smooth, then immersed in wood sealer, conforming to type I or type II of MIL-S-13518, for three minutes. Handles shall be in final form before dipping. After dipping, handles shall be allowed to air-dry approximately one hour for rapid-drying seals, or 18 hours for slow-drying seals. The handles shall then receive a coat of refinishing primer, conforming to TT-P-659, 0.7 mils. minimum dry film thickness. This coat shall be allowed to air-dry until dry enough to recoat (approximately four hours). A minimum dry-film thickness of 1.5 mils. of lustreless olive drab enamel, conforming to TT-E-527 and to color X34087 (see 6.2) of FED-STD-595, shall be applied to the primed surface and baked.

3.3.7 Lubrication. All parts shall be free of oil and grease, except the points of moving parts, which shall be lubricated with an intimate mixture of equal parts, by weight, of zinc dust and petroleum jelly.

3.4 Identification marking. The litter frame, canvas cover, and under surface of the litter securing straps shall be permanently and legibly marked with the name or registered trademark of the manufacturer. On the underside of the litter cover, approximately four inches from one end and centrally located, shall be placed the letters "U.S." four inches high. The litter cover shall also be marked with the contract number. Marking shall be done with black waterproof ink or waterproof stencil paint.

3.5 Workmanship. Workmanship shall be first class throughout. Complete litters shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory and investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties)of the item.

Source: https://assist.dla.mil -- Downloaded: 202610-18T09:01Z
Check the source to verify that this is the currentversion before use.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.1.3 Government furnished material. Government furnished material shall not be sampled or tested unless specifically requested by the Government.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL (percent defective)
For visual examination *		
Major defects (each component).	II	1.0
Minor defects (each component).	II	2.5
For dimensional examination	S-3	2.5

* When the complete litter assembly is examined, each component shall be examined in accordance with the classification of defects for that component in addition to that for the complete litter assembly.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in table I.

Table I - Tests.

Component (unit of product)	Characteristic	Req.	Test procedure	Sample Unit	Insp. level	AQL
Duck, treated	Mildew resistance	3.1.8.1.1	4.4.2	3	S-1	*
Webbing	Mildew resistance	3.1.9	4.4.4	3	S-1	*
Litter (assembled)	Performance tests	3.2	4.4.1	Each	S-2	1.5

* Acceptance number shall be zero for all sample sizes.

4.2.3 Components. Sufficient samples of the component materials shall be selected from each lot of material used in the manufacture of the litter to perform tests as required by the applicable subsidiary specification or by tests specified herein. When sample size for component materials covered by subsidiary specifications is not specified, the sample size for tests shall be as indicated in 4.2.2. The unit of product for testing shall be the quantity of material necessary to perform all required tests one time each, taking into

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

consideration the fact, that where possible, the same material shall be used to perform two or more tests. Lot sizes are expressed as follows:

Duck	- - - - - Yards.
Webbing	- - - - - Yards.
Threads	- - - - - Cones or Spools.
Buckles	- - - - - Each.
End clips	- - - - - Each.

4.2.4 For chemical testing of metals. Four ounces of each lot of each metal used in the construction of the litters shall be selected for chemical analysis as specified in 4.4.5.

4.3 Examination. The litters shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Categories and defects *

Major A Unit of product - litter (AQL 1.0 percent defective)

101. Litter not complete with all component parts, assemblies and subassemblies.
102. Components not properly assembled or aligned.
103. Exposed and accessible edges and surfaces not free of sharp edges and burrs.
104. Center-to-center distance between poles (when litter is spread) not within tolerances specified.
105. Longitudinal distance between bolt holes on pole not within tolerances specified.
106. Bolt holes in poles not centered.
107. Wood handle not a secure fit in pole.
108. Headless nail protrudes from pole.
109. Canvas cover not a taut fit when litter is spread.
110. Litter cannot be spread.

Major B Unit of product - litter cover (AQL 1.0 percent defective)

121. Cover not fabricated from one piece of cotton duck.
122. Cover not constructed so that warp runs parallel to the long axis of the litter.
123. Unselvage edges not hemmed.
124. Width of the cover not within tolerances specified.
125. Cover pockets do not accommodate poles.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

Major C Unit of product - spreader bar and stirrup assembly (AQL 1.0 percent defective)

131. Spreader bar and stirrup assembly not complete with all component parts.
132. Component parts not properly assembled or aligned.
133. Exposed and accessible corners and surfaces not free from sharp edges and burrs.
134. Spreader bars bind in coupler.
135. Center-to-center distance between bolt holes not within tolerances specified.
136. Nut for stirrup bolt not elastic stop, self-locking type.

Major D Unit of product - securing strap (AQL 1.0 percent defective)

141. Strap not complete with all component parts (buckle and end clip).
142. End clip end of strap does not pass through buckle readily.
143. Raw end of strap not impregnated with a hardening material.
144. Strap not free of cuts and tears.

Major E Unit of product - wood handle (AQL 1.0 percent defective)

151. Handle not free of splits, slivers, checks, and knots.
152. Handle not turned for a drive fit into poles.
153. Handle not drilled for drive pin.

Major F Unit of product - headless nail (AQL 1.0 percent defective)

161. Style and design not in accordance with drawing.

Minor A Unit of product -litter (AQL 2.5 percent defective)

201. Style, design, and construction not as shown on drawing.
202. Litter does not rest evenly on a flat surface.
203. Joints of moving parts not lubricated.
204. Enamelled surfaces not free of chipped, cracked, nonadherent, and bare surfaces.
205. Identification marking not complete, not legibly, and not waterproof.

Minor B Unit of product - litter cover (AQL 2.5 percent defective)

221. Style, design, and construction not as shown on drawing.
222. Turnunders, seams, and stitch lines not straight or evenly spaced.
223. Cover not free from loose or ravelled edges.
224. Stitching less than eight stitches per inch.
225. Stitching not backstitched at the ends.
226. Cover not permanently and legibly identified with the name or registered trademark of the manufacturer, the letters U.S., and contract number with black waterproof ink or paint.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

Minor C Unit of product - spreader bar and stirrup assembly (AQL 2.5 percent defective)

231. Style, design, and construction not as shown on drawing.
232. Aluminum parts not chemically treated as specified.
233. Enamelled surfaces not free of chipped, cracked, nonadherent, and bare surfaces.
234. Joints of moving parts not lubricated.
235. Identification marking not permanently or legibly accomplished.

Minor D Unit of product - securing strap (AQL 2.5 percent defective)

241. Style, design, and construction not as shown on drawing.
242. Strap not permanently and legibly identified with the name or registered trademark of the manufacturer in black waterproof ink.

Minor E Unit of product - wood handle (AQL 2.5 percent defective)

251. Style, design, and construction not as shown on drawing.
252. Finish not as specified.

*Examination shall not be restricted to the classified possible listed defects.

4.3.2 Dimensional examination. The litters shall be examined for defects in dimensions. Any dimension (except those specified in the classification of defects) not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted,as necessary, to determine compliance with specification requirements.

4.4.1 Performance tests on litters.

4.4.1.1 Two hundred pounds shall be placed on the litter canvas with the litter suspended from four points located 37-1/2 inches from the center of each pole. This weight shall be distributed as follows: 100 pounds shall be distributed over the center 18 inch length (full width) of the canvas; 30 pounds each on the adjacent 15 inch lengths (full width) of the litter canvas; and 20 pounds each on the adjacent 12 inch lengths (full width) of the litter canvas. The downward deflection of the poles shall not exceed 1-1/4 inches.

4.4.1.2 With the litter resting on the stirrups, 1600 pounds shall be placed on a board, 18 inches wide, resting crosswise on the poles at the center of the litter. This weight shall be built up in a period of five minutes, allowed to remain for 5 minutes,then removed. The permanent downward deflection or set in the poles shall not exceed 5/8 inch.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

4.4.1.3 Litter shall be so constructed that the tension of the canvas cover, when dry, shall require no less than 55 and no more than 70 pounds static load placed at the approximate center of the spreader bar, to open one end of the litter.

4.4.2 Cotton duck. The unfinished duck shall be tested in accordance with the test procedure of CCC-C-419. The finished duck shall be tested in accordance with the test procedures of CCC-D-950, except that the performance of the tests for mildew inhibitor content and mildew resistance shall be conducted under the following conditions:

(a) If the material conforms to the requirements for mildew inhibitor content, the mildew resistance test shall not be performed and the material shall be considered as complying with mildew resistance requirements.

(b) If the mildew inhibitor content exceeds the maximum requirement, the material shall be rejected.

(c) If the mildew inhibitor content is found to be below the minimum requirement, the material may be accepted with respect to mildew resistance, provided it passes the mildew resistance test, method 5762, as specified in the test section of CCC-D-950. When the mildew resistance test is performed, the leaching process shall be omitted.

4.4.3 Thread. The untreated thread shall be tested in accordance with the test procedures of V-T-276. The treated thread shall be tested as specified in the test section of MIL-T-3530.

4.4.4 Webbing. The webbing shall be tested in accordance with the test procedures of MIL-W-530, except that the performance of the tests for mildew inhibitor content and mildew resistance shall be conducted under the following conditions:

(a) If the material conforms to the requirements for mildew inhibitor content, the mildew resistance tests shall not be performed and the material shall be considered as complying with mildew resistance requirements.

(b) If the mildew inhibitor content exceeds the maximum requirement, the material shall be rejected.

(c) If the mildew inhibitor content is found to be below the minimum requirement, the material may be accepted with respect to mildew resistance, provided it passes the mildew resistance test.

4.4.4.1 Mildew resistance. The finished webbing shall be tested in accordance with method 4762 of CCC-T-191, except that the leaching process shall be omitted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

4.4.4.2 Water absorption. The finished webbing shall be tested for water absorption in accordance with method 5500 as specified in the test section of MIL-W-530, except that the size of the specimen and the size of the blotting paper shall conform to method 5502 of CCC-T-191.

4.4.5 Chemical analysis of the metals. The samples selected in accordance with 4.2.4 shall be analyzed in conformance with methods 111 or 112 of Federal Test Method Standard No. 151. In case of dispute, method 111 shall be used for referee analysis.

4.4.6 Enamel. Enamel shall be tested in accordance with TT-E-527.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One litter, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit package. Each litter shall be folded and wrapped in kraft paper of suitable weight. Wrapping shall be adequately secured to prevent accidental opening.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 3 load and constructed in accordance with PPP-B-576, style A, class 2, or PPP-B-601, table II, using type III, class 1 plywood.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 3 load and constructed in accordance with PPP-B-576, style A, class 1 or PPP-B-601, table I, using type III, class 1 plywood.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.3 Marking.

5.3.1 Interior packages and exterior container. Each unit package and each exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.

(b) Selection of applicable levels of packing (see 5.1 and 5.2).

6.2 Availability of color chips. The individual 3 by 5 inch chips of the color referenced in 3.3.5.1 and 3.3.6 are available from General Services Administration, Business Service Center, Region 3, Seventh and D Streets, S.W., Washington, D. C. 20704. Price - five cents.

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

CATEGORY: N

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION: MIL-L-16446A

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64 WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

FOLD

DEFENSE SUPPLY AGENCY
DEFENSE PERSONNEL SUPPORT CENTER
2300 SOUTH 20TH ST.
SPHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER

ATTN: DIRECTORATE MEDICAL MATERIEL

2800 SOUTH 20TH STREET

PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:01Z
Check the source to verify that this is the current version before use.