GG-A-616E
July 20, 1978
SUPERSEDING
Fed. Spec. GG-A-616D
September 8, 1965

FEDERAL SPECIFICATION

APPLICATOR, DISPOSABLE

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers single ended applicators suitable for use in medical and dental procedures.

1.2 Classification. The applicators shall be either plain or tipped and furnished either sterile or non-sterile as specified in the invitation for bid or request for proposal (see 6.2).

Type I - Wood, cotton or rayon tipped, 3 inches by 1/12 inch diameter, nonsterile.

Type II - Wood, cotton or rayon tipped, 6 inches by 1/12 inch diameter, nonsterile.

Type III - Wood, cotton or rayon tipped, 6 inches by 1/12 inch diameter, sterile.

Type IV - Wood, non-tipped, 6 inches by 1/12 inch diameter, nonsterile.

Type V - Wood or plastic, cotton tipped, 6 inches by 1/12 inch diameter, sterile.
Class 1 - 100 per package.
Class 2 - 2000 per package (1000 envelopes each containing 2 applicators).

Type VI - Suitable material (see 3.1.4), cotton tipped, 6 inches by 1/12 inch diameter, sterile.

Type VII - Wood, non-tipped, 6-1/4 inches by 1/12 inch diameter, nonsterile.

Type VIII- Wood, non-tipped, 12 inches by 1/12 inch diameter, nonsterile.

Type IX - Wood, non-tipped, 12 inches by 1/8 inch diameter, nonsterile.

Type X - Wood, cotton or rayon tipped, 18 inches by 1/8 inch diameter, nonsterile.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issues in effect on the date of invitation for bid or request for proposal, form a part of this specification to the extent specified herein.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

Federal Specifications:

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-636 - Box, Shipping, Fiberboard.
PPP-B-676 - Boxes, Set-Up, Paperboard.

Federal Standards:

Fed. Std. No. 102 - Preservation, Packaging, and Packing levels.
Fed. Std. No. 123 - Marking for Shipment.

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, US Government Printing Office, Washington, DC 20402.

(Single copies of this specification and other Federal Specifications required by activities outside the Federal Government for bidding purposes are available without charge from Business Service Centers at the General Services Administration Regional Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas City, MO, Fort Worth, Denver, San Francisco, Los Angeles, and Seattle, WA.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications and Standards from established distribution points in their agencies.

Military Specifications:

MIL-S-36586 - Sterilization Test Strip Set, Bacterial Spore.

Military Standards:

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.
MIL-STD-1363 - Measurement of Wood Moisture Content.

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bid or request for proposal shall apply.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

The United States Pharmacopeia Convention, Inc.

The United States Pharmacopeia (USP)

(Applications for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

3. REQUIREMENTS

3.1 Material. The materials for the types of applicators classified in 1.2 shall be as specified herein.

3.1.1 Wood. The wood shall be a tough, flexible, well seasoned wood, such as gumwood, birch, or other wood suitable for the purpose. The wood shall be of high quality, free from cross grains, knots, stains, or fungus growth. The wood shall be finished smooth.

3.1.1.2 Moisture content. Moisture content after fabrication and after sterilization of the wooden applicator has been completed, shall be not less than 6 percent nor more than 12 percent when tested as specified in MIL-STD-1363.

3.1.2 Plastic. The plastic shall be a suitable material and shall permit a bending arc of not less than 80 degrees and shall show no evidence of fracturing, splitting or breaking when tested as specified in 4.4.1.

3.1.3 Cotton. The cotton shall be in accordance with the requirements of the United States Pharmacopeia (USP) for purified cotton, excluding sterility.

3.1.4 Rayon. The rayon shall be in accordance with the requirements of the United States Pharmacopeia (USP) for purified rayon, excluding sterility. In addition, the rayon shall contain not more than 30 parts per million (ppm) of lead and not more than 30 ppm of heavy metals when tested as specified in the USP.

3.1.5 Material for type VI. Type VI applicators shall be fabricated from materials capable of withstanding autoclaving after the initial sterilization, and withstand chemicals such as ether and acrylic monomers used in dental laboratories. When material other than wood are used for the stick, the stick shall permit a bending arc of not less than 80 degrees and shall show no evidence of fracturing, splitting or breaking when tested as specified in 4.4.1. The finished applicator shall not deteriorate, melt, warp, nor shall the cotton tip loosen when autoclaved as specified in 4.4.2. The finished applicator shall not dissolve, deteriorate, become tacky nor shall the cotton tip loosen when tested for chemical resistance as specified in 4.4.3. The material for the type VI applicator shall not be metal.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

3.1.6 Adhesive. The adhesive used in the fabrication of the applicators shall inhibit mold growth or shall contain a preservative which inhibits mold growth. The adhesive shall be nontoxic and capable of withstanding sterilization without affecting the intended use of the item. The adhesive used in the fabrication of the applicators shall not deteriorate during sterilization. The adhesive used in the fabrication of the type VI applicators shall meet the autoclaving and chemical resistance requirements specified in 3.1.5.

3.2 Construction. Construction of the applicators shall be as specified in 1.2 and table I for the type specified. The stick portion of each applicator shall be smooth and symmetrical throughout its length and the ends shall be cut square and smooth. To form the tipped applicators, one end shall be wound with material as specified.

3.2.1 Tipped applicators. The tip shall be pear-shaped and shall cover one end of the stick to a thickness of not less than 1/16 inch. Winding of material to form the tip shall be smooth, with no loose fibers hanging. The tip shall be securely attached to end of applicator with glue (3.1.6) without interference with absorptivity of the fibers.

3.3 Dimensions. Applicators shall conform to the requirements of table I.

3.4 Weight. The average weight of the rayon and cotton tips shall be as specified in table I when tested as specified in 4.4.4.

TABLE I. Dimensions

Type	Stick			Tip		
	Length (inches)	Diameter (inches) + 0.03 - 0.01	Overall length (inches)	Diameter at Large end (inches)	Length of tip (inches)	Weight per 1000 (grains)
I	3 ± 1/4	0.083	3-1/16 ± 1/4	3/16 ± 1/32	5/8 ± 1/16	600 ± 200
II, III	6 ± 1/4	0.083	6-1/16 ± 1/4	3/16 ± 1/32	5/8 ± 1/16	600 ± 200
IV	6 ± 1/4	0.083	6 ± 1/4	not required	-	-
V, VI	6, + 1/4, -5/16	0.083	6-1/16, + 1/4 -5/16	3/16 ± 1/32	5/8,+ 3/16, -1/16	600 ± 200
VII	6-1/4 ± 5/16	0.083	6-1/4 ± 1/16	not required	-	-
VIII	12 ± 1/4	0.083	12 ± 1/4	not required	-	-
IX	12 ± 1/4	0.125	12 ± 1/4	not required	-	-
X	18 ± 1/2	0.125	18-1/4 ± 1/2	1 ± 1/4	2 ± 1/2	6000 ± 400

3.5 Sterility. Types III, V and VI applicators shall be sterile when tested with a biological indicator as specified in 4.4.5.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

3.5.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National stock number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of a laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the procuring activity (6.5), and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.5.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.2.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.6 Workmanship. Workmanship shall be first class throughout. The applicator shall be free from defects which detract from its appearance or impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/ material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table II. The unit of product for examination purposes shall be one applicator or one immediate package as applicable. For defects pertaining only to the envelope, the unit of product for examination purposes shall be the envelope.

TABLE II. Sampling for examination

Type of defect	Inspection level	AQL (percent defective)
Major	S-4	1.0
Minor	S-4	2.5
Dimensions	S-4	6.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in table III. Where feasible, the same sample unit shall be used for the determination of two or more characteristics. The unit of product for testing flexibility, moisture content, autoclaving and chemical resistance shall be one applicator. The unit of product for sterility testing shall be one sealed envelope containing applicators.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

TABLE III. Sampling for tests

Component	Characteristic	Requirement	Test	Inspection level	AQL (percent defective)
Plastic stick	Flexibility	3.1.2	4.4.1	S-2	4.0
Wood stick	Moisture content	3.1.1.2	MIL-STD-1363	S-2	4.0
Cotton	Purified	3.1.3	USP	1/	1/
Rayon	Purified	3.1.4	USP	1/	1/
	lead content	3.1.4	USP	2/	2/
	free heavy metals	3.1.4	USP	2/	2/
Type VI material	Flexibility	3.1.5	4.4.1	S-2	4.0
	Autoclaving	3.1.5	4.4.2	S-2	4.0
	Chemical resistance	3.1.5	4.4.3		
Tips	Weight	3.4	4.4.4	4/	4/
Type III, V, VII applicator	Sterility	3.5	4.4.5	S-3	5/

1/ Each test conducted in accordance with USP shall be run twice per lot of cotton and rayon and the lot shall be rejected when one individual test fails.
2/ The lead content and free heavy metals test shall be run twice per lot of rayon and the lot shall be rejected when either test fails.
3/ The weight test shall be conducted 10 times per lot of tips and the lot shall be rejected when two or more individual weighing tests fail.
4/ One sterility test failure shall reject the entire lot.

4.3 Examination.

4.3.1 Applicator. The applicator shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in tables II and IV. Examination, however, shall not be restricted to defects listed in table IV.

TABLE IV. Classification of defects

Classification	Defect
Major:	
101	Applicator stick not smooth.
102	Applicator stick broken.
103	Incomplete seal of sterile bag or sterile envelope.
104	Sterile bag or sterile envelope torn.
105	Type V, class 2 sterile envelope capable of being resealed.
106	Type V, class 2 sterile envelope lacks "peel-back" opening tabs.
107	Sterile envelope tears when peeled back.
108	Cotton or rayon tip unclean or contains foreign matter.
109	Cotton or rayon tip loose.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

TABLE IV. Classification of defects (Cont'd)

Classification	Defects
Minor:	
201	Dimensions not as specified.
202	Stick portion of applicator not symmetrical throughout.
203	Stick portion of applicator ends not cut square and smooth.
204	Cotton or rayon tip not pear shaped.
205	Cotton or rayon tip does not cover end of applicator to a depth by at least 1/16 inch.

4.3.2 Dimensional examination. The applicators shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.3.3 Examination of preparation for delivery. An examination shall be performed to determine compliance with the packaging, packing, and marking requirements of section 5. Defects shall be scored as specified in table V. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for examination at one time. The inspection level shall be S-2 with an AQL of 4.0 expressed in terms of percent defective.

TABLE V. Examination of preparation for delivery

Examine	Defect
Container	Not as specified.
Content	Not as specified.
Marking	Omitted; incorrect; illegible; improper size, location, sequence or method of application.
Packaging and packing material	Component missing or damaged.
Workmanship	Bulging or distortion of container.

4.3.4 Examination of closure, waterproofing, and banding. When containers are required to comply with PPP-B-566, PPP-B-636, or PPP-B-676, examination for defects in closure, waterproofing, and banding shall be in accordance with the appendix to the appropriate specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements. Separate applicators shall be used for each test except type VI when tested in accordance with 4.4.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

4.4.1 Flexibility. After the initial sterilizing, the applicator shall be tested for flexibility as follows: Grasp the stick with the fingers 3/4 ± 1/4 inch from each end and slowly bend the stick through an arc (3.1.2 and 3.1.5). For type VI, this test shall be performed after the test specified in 4.4.2.

4.4.2 Autoclaving. The type VI applicator shall be autoclaved with steam at 260°F (127°C) for 20 minutes (3.1.5) after their initial sterilization.

4.4.3 Chemical resistance. The type VI applicator shall be submerged in ether for 30 minutes and submerged in liquid methyl methacrylate monomer for 30 minutes (3.1.5 and 6.4). Separate applicators shall be used for each test. This test shall be performed after initial sterilization.

4.4.4 Weight. The weight of the tips shall be determined by weighing 100 tips and multiplying the result by 10. Weight of tips shall be based on dry weight (3.4).

4.4.5 Sterility test. The applicator shall be tested for sterility in accordance with the method specified in the USP for purified cotton and purified rayon using inoculated product or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in 4.4.5.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for type I sterilization and 30 for type II sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in table II, Sampling for tests.

4.4.5.1 Inoculated product. For each type of sterilization, the organism used shall be as listed below:

Type of sterilization	Organism	Incubation temperature
Steam	B. stearothermophilus	55° - 65° C
Dry heat	B. stearothermophilus, or	55° - 65° C
	B. subtilis var. niger (globigii)	30° - 35° C
Gas	B. subtilis var. niger (or globigii)	30° - 35° C
Radiation	B. pumilus	30° - 35° C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:51Z
Check the source to verify that this is the currentversion before use.

4.4.5.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.5 Sterility test and 4.4.5.1 inoculated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.5.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.4.5.3 Alternate method. Contractors conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the contracting officer, and approval obtained before employment of the alternate method on a government contract.

4.4.5.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C (commercial) as specified (see 6.2).

5.1.1 Unit package.

5.1.1.1 Types I, II and X. One hundred type I or type II applicators or five hundred type X applicators shall be packaged in a bag fabricated of white kraft or other suitable paper or in a box conforming to PPP-B-566 or PPP-B-676. Closure of the bag shall be secured with a strip of tape or by other suitable means. Closure of boxes shall be as specified in the applicable box specification.

5.1.1.2 Type III and type V, class 1. One hundred type III or type V, class 1 applicators shall be packaged in a sealed bag fabricated of suitable material and being capable of maintaining sterility of contents unless damaged or opened. The closure of the bag shall not be capable of being resealed.

5.1.1.3 Type V, class 2. Two applicators shall be packaged in a sealed peel-open type envelope capable of maintaining sterility of contents unless the envelope is damaged or opened. The envelope shall be designed with tabs. The tabs shall be capable of being readily peeled back along the sealed edges of the envelope to expose contents. The envelope shall strip cleanly without tearing or leaving jagged edges of paper extending into the sterile field. The envelope shall not be capable of being resealed. One thousand envelopes, each containing two applicators shall be packaged in a box conforming to PPP-B-566 or PPP-B-676. Closure shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:51Z
Check the source to verify that this is the current version before use.

5.1.1.4 Type VI. One hundred type VI applicators shall be packaged in a bag as specified in 5.1.1.2. Ten bags, each containing one hundred applicators, shall be packaged in a box conforming to PPP-B-566, PPP-B-636, class domestic or PPP-B-676. Closure shall be as specified in the appendix of the applicable box specification.

5.1.1.5 Types IV, VII, VIII and IX. Eight hundred and sixty-four type IV or type VII applicators or five hundred type VIII or type IX applicators shall be packaged in a box conforming to PPP-B-566 or PPP-B-676. Closure shall be as specified in the appendix of the applicable box specification.

5.1.2 Intermediate package. Twenty unit packages containing applicators of one type only of type I thru type IV, type V, class 1, or type IX, shall be packaged in an intermediate package conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.1.3 Level C. Applicators shall be clean dry and packaged in a manner that will afford adequate protection against deterioration and physical damage during shipment from the supply source to the first receiving activity. In addition, sterile items shall be packaged in a manner so as to maintain sterility of contents unless the package is damaged or opened.

5.2 Packing. Packing shall be level A, B or C (commercial) as specified (see 6.2).

5.2.1 Level A. Two intermediate packages containing applicators of one type only of type I thru type IV, type V, class 1, or type IX, or six unit packages of type V, class 2, or twenty-four unit packages of one type only of type VI, type VIII or type X shall be packed in a shipping container conforming to PPP-B-636, class weather-resistant. Closure, waterproofing and banding shall be as specified in the appendix of the box specification.

5.2.2 Level B. The number of unit or intermediate packages as specified (see 6.2) shall be packed in a shipping container conforming to PPP-B-636, class domestic. Closure shall be as specified in the box specification.

5.2.3 Level C. The packaged applicators shall be packed in shipping containers in a manner that will afford adequate protection against damage during shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carriers rules and regulations.

5.3 Marking.

5.3.1 Civil agencies. Marking shall be as specified in Fed. Std. No. 123.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:51Z
Check the source to verify that this is the current version before use.

5.3.2 Military activities.

5.3.2.1 Interior packages. Each unit and intermediate package shall be marked as specified in MIL-STD-129. In addition, marking of sterile items shall include the lot or control number and the method of sterilization. The method of sterilization (G-Gas, R-Radiation, S-Steam, D-Dry heat) shall be shown following the level of packaging and date of packaging. Example: "A 5/78-G" Final sterilization shall be on or after indicated packaging date. Each unit package containing type III and type V, class 1 applicators shall also include the legend "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED" or similar statement. Each envelope containing type V, class 2 applicators shall be marked with the lot or control number and the sterility legend in addition to any commercial markings.

5.3.2.2 Shipping containers. Shipping containers shall be marked as specified in MIL-STD-129. In addition, the marking of sterile items shall include the lot or control number and the method of sterilization.

6. NOTES

6.1 Intended use. The applicators covered by this specification are intended for use in the application of medicaments and for use in dental procedures.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

a. Title, number and date of this specification.
b. Type and class required (1.2).
c. The National stock number (NSN).
d. Selection of applicable levels of packaging and packing (5.1 and 5.2).
e. Quantity of applicators in intermediate and exterior containers.

6.3 This specification is intended to cover the following items appearing in the Federal Supply Catalog:

National Stock Number	Item Identification	Type	Class
6515-00-303-8100	APPLICATOR, DISPOSABLE, WOOD, 1/12 by 6-1/4 inches, 864s	VII	
6515-00-303-8200	APPLICATOR, DISPOSABLE, WOOD, 1/8 by 12 inches, 500s	IX	
6515-00-303-8250	APPLICATOR, DISPOSABLE, WOOD OR PLASTIC, Cotton Tipped End, 1/12 by 6 inches, 100s	V	1
6515-00-905-1473	APPLICATOR, DISPOSABLE, WOOD OR PLASTIC, Cotton Tipped End, 1/12 by 6 inches, 2000s	V	2
6515-01-017-2177	APPLICATOR, DISPOSABLE, Cotton Tipped End, 1/12 by 6 inches, 1000s	VI	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

6.4 The liquid monomer can be obtained as: Lucitone Standard Liquid Catalog No. 684305 from L.D. Caulk Co., Division of Dentsply International, Inc., Milford, Delaware, 19963.

6.5 For military procurements originating from the Defense Personnel Support Center, Philadelphia, PA, certificates of sterility are to be forwarded to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101. For procurements originating from other agencies the certificate of sterility shall be forwarded as indicated by the contracting officer (3.5.1).

MILITARY CUSTODIANS:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DOD - MB

Preparing activity:
DLA-DM

Civil Agency Coordinating Activities:
HEW-FDA
GSA-FSS

Project No. 6515-2097

Orders for this publication are to be placed with General Services Administration, acting as agent for the Superintendent of Documents. See Section 2 of this specification to obtain copies of documents referenced herein. Price 90 cents each.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:51Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

GG-A-616E APPLICATOR, DISPOSABLE

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*

TELEPHONE NO.

DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:51Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

POSTAGE AND FEES PAID



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.