GG-B-00596b (DSA-DM)
April 22, 1970
Interim Revision of
Fed. Spec. GG-B-596a
20 March 1953

INTERIM FEDERAL SPECIFICATION

BOUGIE, EUSTACHIAN, SET OF FIVE

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

The General Services Administration has authorized the use of this interim Federal specification as a valid exception to Federal Specification GG-B-596a.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers Eustachian bougies suitable for dilation of the Eustachian tube.

1.2 Classification.

1.2.1 Type and grade. Eustachian bougies covered by this specification shall be of one type and grade.

1.2.2 Sets and sizes. The bougies shall be furnished in graduated sets of five having the dimensions as shown in 3.3.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

2. APPLICABLE SPECIFICATIONS AND STANDARDS.

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

H-S-951	Synthetic Filament; General Specification for Brushes.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up.

STANDARDS

Federal

FED-STD-123	Marking for Domestic shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D. C. 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston; New York; Atlanta; Chicago; Kansas City, Mo.; Fort Worth; Denver; San Francisco; Auburn, Washington; and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Military

MIL-L-10547 Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

(Copies of Military specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

* 3.1 Material. Eustachian bougies shall be made of solid nylon filament.

3.2 Style and design. Eustachian bougies consisting of graduated sets of five shall be double-ended with tapering shafts which shall be straight and flexible.

* 3.3 Dimensions. Dimensions shall be 8 inches ± 1/4 inch over-all length with maximum thickness at center, tapering gradually toward each end to a minimum diameter at junction with the opposing bulbous ends. Bougies shall have diameters in inches as follows in Table I:

Table I. Bougie Dimensions

Number	End A	Bulb A	Center	End B	Bulb B
	Inches	Inches	Inches	Inches	Inches
1....	0.015	0.021	0.035	0.030	0.033
2....	.018	.024	.035	.035	.035
3....	.030	.037	.037	.030	.037
4....	.035	.044	.050	.035	.044
5....	.035	.044	.055	.050	.053

Tolerance for dimensions in table above is ± 0.002.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

* 3.4 Water immersion. The bougies shall show no appreciable change in physical characteristics when tested as described in 4.2.4.1.

3.5 Melting point. The melting point of the nylon shall be not less than 135° C when tested as described in 4.2.4.2.

3.6 Workmanship. Workmanship shall be first class throughout. Bougies shall be symmetrical in shape and the surface shall be without irregularities or imperfections of any kind.

4. QUALITY ASSURANCE PROVISIONS

* 4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

* 4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in Tables II, III, and IV.

* 4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in Table II.

Table II. Inspection and Acceptable Quality Levels

Type defect	Inspection level	AQL (percent defective)*
Major	S-2	1.0
Minor	S-2	1.5

* If acceptance number is not exceeded, lot shall be accepted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

* 4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified in Table III. The unit of product for inspection purposes shall be one set of bougies. Acceptance number shall be zero for all sample sizes. Where feasible, the same unit of product shall be used for the determination of two or more test characteristics.

Table III. Sampling Criteria for Tests

Characteristic	Requirement	Method	Inspection Level	AQL
Water resistance	3.4	4.2.4.1	S-1	*
Melting point	3.5	4.2.4.2	S-1	*

* Acceptance number shall be zero for all sample sizes.

* 4.2.3 Examination. The bougies shall be examined to determine compliance with the requirements contained in this specification. Examination shall be conducted in accordance with the following classification of defects. Examination shall include, but shall not be limited to the defects listed in Table IV.

Table IV. Sampling Criteria for Tests

Categories	Defects
Major	
101	Surface finish not smooth.
102	Surface not free of burrs.
103	Taper from center to ends not gradual.
104	Shafts not straight.
105	Bougies do not retain original physical characteristics.
106	Dimensions of centers and bulbous ends not as specified.
Minor	
201	Length not as specified.
202	Packaging inadequate.
203	Dimensions of ends not as specified.

4.2.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

4.2.4.1 Water immersion. Bougies shall be immersed in distilled water for a period of 24 hours at room temperature. The bougies shall be examined visually after immersion to determine if changes in physical characteristics have occurred.

4.2.4.2 Melting point. A piece of nylon approximately 1/8 inch long shall be placed in a capillary tube closed at the lower end. The nylon so encased shall be attached to a thermometer at a point corresponding to the thermometer bulb and the assembly immersed to the proper depth in a bath at a temperature of 135° C for 5 minutes. The melting point determination and apparatus used shall be as described in Federal Specification H-S-951 for Class N filaments.

* 5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (6.1).

5.1.1 Unit of issue. One set, consisting of 5 bougies, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table V. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	6 units	576 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be completely wrapped in neutral tissue paper and packaged in an envelope or carton of appropriate size and design. Closure shall be adequate to prevent accidental opening.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-566 or PPP-B-676, except commercial colors will be acceptable, or PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipment shall be in accordance with Fed Std No. 123.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

5.3.2 Military activities.

5.3.2.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number, title and date of this specification.
(b) Applicable levels of packaging and packing.
(c) Federal stock number.

* 6.2 This specification covers the following item appearing in Federal Supply Catalog: FILIFORM SET, EUSTACHIAN, Nylon, Set of 5 FSN 6515-312-0880.

6.3 Supersession data. This specification includes the requirements of Purchase Description No. 5, dated 9 August 1966.

* 6.4 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:

Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1474

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION
00-B-00596b(DSA-DM) FSN:

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:51Z
Check the source to verify that this is the current version before use.