IS IS A REPRINT - ONLY CHANGE IS IN INFORMATION
FOR OBTAINING COPIES OF THIS DOCUMENT

GG-B-596a
20 MARCH 1953

SUPERSEDING
Int. Fed. Spec. GG-B-00596 (Army-Med)
5 August 1952

FEDERAL SPECIFICATION

BOUGIE, EUSTACHIAN, SET OF FIVE

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. — This specification covers Eustachian bougies suitable for dilation of the Eustachian tube.

1.2 Classification.

1.2.1 *Type and grade.* — Eustachian bougies covered by this specification shall be of one type and grade.

1.2.2 *Sets and sizes.*—The bougies shall be furnished in graduated sets of five having the dimensions as shown in 3.3.

2. APPLICABLE SPECIFICATIONS

2.1 Specifications.—The following Federal Specifications of the issues in effect on date of invitation for bids, form a part of this specification:

H-S-951—Synthetic Filaments, General Specification for Brushes.

VV-L-791—Lubricants, Liquid Fuels, and Related Products; Methods of Inspection, Sampling, and Testing.

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications, Standards, and Handbooks and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications, Standards, and Handbooks from established distribution points in their agencies.)

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

cago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C. Prices of additional copies may be obtained from the GSA Regional Offices.)

2.2 Specifications and standards applicable only to individual departments are listed in section 7.

3. REQUIREMENTS

3.1 Material.—Eustachian bougies shall be made of solid nylon filament. The melting point of nylon shall be not less than 135° C. when tested as described in 4.3.2.

3.2 Style and design.—Eustachian bougies consisting of graduated sets of five shall be double-ended with tapering shafts which shall be straight and flexible.

3.3 Dimensions.—Dimensions shall be approximately 8 inches over-all length with maximum thickness at center, tapering gradually toward each end to a minimum diameter at junction with the opposing bulbous ends. Bougies shall have approximate diameters in inches as follows:

Number	End A	Center	End B
	Inches	*Inches*	*Inches*
1	0.015	0.035	0.030
2	.018	.035	.035
3	.030	.037	.030
4	.035	.050	.035
5	.035	.055	.050

203436-53-(210)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

3.4 **Water immersion.**—The bougies shall how no appreciable change in physical characteristics when tested as described in 4.3.1.

3.5 **Workmanship.** — Workmanship shall be first class throughout. Bougies shall be symmetrical in shape and the surface shall be without irregularities or imperfections of any kind.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 **Sampling.**—Unless otherwise specified, at least 2 sets of bougies shall be taken at random from each 100 or fraction thereof delivered.

4.2 **Inspection.**—Samples shall be inspected for material, style and design, dimensions, workmanship, packaging, packing, and marking.

4.3 **Tests.**

4.3.1 *Water immersion.*—Bougies shall be immersed in neutral distilled water at room temperature for a period of 24 hours. At the end of the immersion period, the bougies shall be examined visually for any change in physical characteristics.

4.3.2 *Melting point.* — A piece of nylon, approximately ⅛ inch long, taken from the base shall be placed in a capillary tube closed at the lower end. The melting point of the nylon shall be determined as described in Federal Specification H-S-951. The partial immersion thermometer is described in Federal Specification VV-L-791 under Method No. 950.1.2.

5. PREPARATION FOR DELIVERY

5.1 **Packaging.** — Unless otherwise specified, each set of bougies shall be packaged in accordance with accepted commercial practice.

5.2 **Packing.** — The subject commodity shall be prepared for shipment to permit acceptance by carrier for transportation at the lowest applicable rate, and to afford maximum protection from normal hazards of transportation.

5.3 **Marking.**

5.3.1 *Packages.* — Unless otherwise specified, packages shall be marked with the name and size of the commodity, the quantity contained therein, and the names of the manufacturer and contractor.

5.3.2 *Shipping containers.*—Unless otherwise specified, shipping containers shall be marked with the name and size of the commodity as defined by the contract or order under which the shipment is made, the quantity contained therein, the name of the contractor, the number of the contract or order, the name and address of the consignee, and the gross weight.

6. NOTES

6.1 Purchasers should exercise any desired options offered herein. (See 4.1, 5.1, 5.3, 6.2, and 7.2.6.)

6.2 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.3 This specification covers the following items appearing in the Armed Services Catalog of Medical Matériel:

3-120-880—Bougie, Eustachian, Set of 5.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.

6.4 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the type, etc., normally procured for Federal Government requirements.

6.5 Transportation description.

6.5.1 *Transportation description applicable to these items is:*

Surgical instruments, not otherwise indexed by name.
Carload weight 10,000 pounds subject to Rule 34, Uniform Freight Classification.
Truckload weight 10,000 pounds.

Patent notice.—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

7. DEPARTMENTAL REQUIREMENTS

7.1 The following specifications and standard, of the issues in effect on date of invitation for bids, and special requirements, form a part of this specification for purchases made under this specification by the respective departments.

7.2 Army, Navy, and Air Force.

7.2.1 *Applicable specifications and standard.*

Federal Specifications:

NN-B-601—Boxes; Wood-Cleated-Plywood, for Domestic Shipment.
NN-B-621—Boxes; Wood, Nailed and Lock-Corner.
NN-B-631—Boxes; Wood, Wirebound (for Domestic Shipment).

Military Specifications:

JAN-P-105—Packaging and Packing for Overseas Shipment—Boxes, Wood, Cleated, Plywood.
JAN-P-106—Packaging and Packing for Overseas Shipment—Boxes; Wood, Nailed.
MIL-B-107—Boxes, Wood, Wirebound (Overseas Type).
JAN-P-120—Packaging and Packing for Overseas Shipment — Cartons, Folding, Paperboard.
JAN-P-133—Packaging and Packing for Overseas Shipment—Boxes, Set-Up, Paperboard.
MIL-L-10547—Liners, Case, Waterproof.

Navy Department Specification:

General Specifications for Inspection of Material.[1]

Military Standard:

MIL-STD-129—Marking of Shipments.

(Copies of specifications, standards, and drawings required by contractors in connection with specific procurement functions should be obtained from the procuring agency or as directed by the contracting officer.)

7.2.2 *Packaging.*

7.2.2.1 *Unit packaging.*—Five bougies, as described in 3.3, constitute one unit. Each unit shall be completely wrapped with neutral tissue and packaged in a carton or envelope of appropriate size and design. Carton or envelope shall be adequately secured to prevent accidental opening.

7.2.2.2 *Intermediate packaging.*—Six unit packages (6 units) shall be packaged in a carton of appropriate size conforming to the requirements of Military Specification JAN-P-120 or JAN-P-133. Each carton shall be adequately secured to prevent accidental opening.

[1] Applicable to Navy purchases only.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:51Z
Check the source to verify that this is the current version before use.

7.2.3 *Packing.*

7.2.3.1 *For domestic shipment.* — Ninety-six intermediate cartons (576 units) shall be packed in a shipping container conforming to the requirements of Federal Specification NN-B-601, NN-B-621, or NN-B-631.

7.2.3.2 *For overseas shipment.* — Ninety-six intermediate cartons (576 units) shall be packed in a shipping container designed for type 2 load, and conforming to the requirements of Military Specification JAN-P-105, JAN-P-106, or MIL-B-107. Closure and strapping shall conform to the requirements of the appendix to applicable box specification.

7.2.3.2.1 *Case liner.* — Each box shall be lined with a waterproof case liner conforming to the requirements of Military Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

7.2.3.3 *Packing variation.*—If the number of units in any shipment is not the exact multiple of the number of units specified to be contained in each shipping container, the units not constituting the nearest such multiple shall be packed in a shipping container of suitable size and design and so constructed as to insure safe delivery to specified destination.

7.2.4 *Marking.*

7.2.4.1 *Packages.* — Each unit container and each intermediate carton shall be marked with stock number, nomenclature, unit, quantity, office identifying number, and name of contractor.

7.2.4.2 *Shipping containers.* — Each shipping container shall be marked in accordance with Military Standard MIL-STD-129.

7.2.5 *Inspection procedures (Navy only).* —The general inspection procedures shall be in accordance with Navy Department General Specifications for Inspection of Material.

7.2.6 *Ordering data.*—Procurement documents should specify whether the subject commodity is to be packed for domestic or overseas shipment. (See 7.2.3.)

ARMY INTEREST: M

NAVY INTEREST: Md

Copies of this specification can be purchased for 5 cents each from the General Services Administration, Business Service Center, Region 3, Seventh and D Streets, S. W., Washington 25, D. C. Single copies required for bidding purposes are available without charge at the nearest GSA Regional Office.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:51Z
Check the source to verify that this is the current version before use.