GG-P-00636B (DSA-DM)
15 February 1973
SUPERSEDING
Int. Fed. Spec.
GG-P-00636a (DSA-DM)
17 June 1964 and
Fed. Spec. GG-P-636
9 October 1952

INTERIM FEDERAL SPECIFICATION

PRESS, DENTURE FLASK

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal Agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this Interim Federal Specification as a valid exception to Federal Specification GG-P-636.

1. SCOPE

1.1 This specification covers a denture flask press for holding resin cases under pressure during the curing process.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-C-591	Copper-silicon Alloy, Rod, Wire, Shapes, and Flat Products (Flat Wire, Strip, Sheet, Bar, and Plate).
MMM-A-250	Adhesive Water-resistant (For Sealing Fiberboard Boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

NDARDS

ederal

FED-STD-66	Steel, Chemical Composition and Hardenability.
FED-STD-123	Marking for Domestic Shipment (Civilian Agencies).
FED-STD-151	Metal, Test Methods.

ivities outside the Federal Government may obtain copies of Federal cations and Standards as outlined under General Information in the f Federal Specifications and Standards and at the prices indicated Index. The Index, which includes cumulative monthly supplements as is for sale on a subscription basis by the Superintendent of ts, U.S. Government Printing Office, Washington, D.C., 20402.

gle copies of this specification and other product specifications d by activities outside the Federal Government for bidding purposes ilable without charge at the General Services Administration Regional in Boston, New York, Alanta, Chicago, Kansas City, Mo., Dallas, San Francisco, Los Angeles, Seattle, and Washington, D.C.)

eral Government activities may obtain copies of Federal Specifications ndards and the Index of Federal Specifications and Standards from shed distribution points in their agencies.)

DARDS

litary

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

NGS

ense Personnel Support Center (DPSC)

20125	Flask, Denture, Ejector Type. Lower.
20485	Flask, Denture, Ejector Type, Upper.
20638	Press Denture Flask.

es of specifications, standards and drawings required by suppliers ction with specific procurement functions should be obtained from the g activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

2.2 OTHER PUBLICATIONS. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

B 147 Spec. for High - Strength Yellow Brass (Manganese Bronze) and Leaded High - Strength Yellow Brass (Leaded Manganese Bronze) Sand Castings.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Phila., Pa. 19103).

3. REQUIREMENTS

3.1 Drawings. The drawings forming a part of this specification are specification control drawings. The supplier shall be responsible for preparing his own shop drawings. Where tolerances prescribed could cumulatively result in improper design, the supplier shall provide a product with dimensions within the tolerances prescribed on the drawing to insure proper function, fit, assembly or operation of the item. No deviation from the prescribed dimensions or tolerances is permissible without prior approval of the contracting officer.

3.2 Materials. Materials shall be new and shall conform to the requirements specified herein. Materials not specifically designated in table I shall be of high commercial quality throughout and shall be suitable for the intended purpose.

Table I. Materials

Component	Material	Specification
Frame	Bronze casting	ASTM B 147, Alloy 865
Lifting loop	Bronze wire	-
Pressure screw	Copper silicon alloy	QQ-C-591, Alloy 661
Upper pressure plate	Bronze casting	ASTM B 147, Alloy 865
Lower pressure plate	Bronze casting	ASTM B 147, Alloy 865
Coil spring	Corrosion-resistant steel, 18-8	-
Wrench	Cold rolled steel, B1112, Cadmium plated	FED-STD-66
Holding handle		
Rod	Cold rolled steel, cadmium plated	-
Ferrule	Steel, nickel plated	-
Handle	Hardwood	-

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

Design and construction. The flask press shall satisfactorily
iate two upper or two lower ejector-type denture flasks, in accordance
ense Personnel Support Center (DPSC) Drawing Nos. 20485 (for upper
and 20125 (for lower flask). The flask press shall consist of a
lifting loop, pressure screw, upper pressure plate, lower pressure
:wo coil springs, wrench, and holding handle. The frame base, top
:s shall be a one-piece cast construction and the upper and lower
: plates shall each be a one-piece casting. The pressure screw shall
nently attached to the upper pressure plate. The flask press shall
to Defense Personnel Support Center (DPSC) Drawing No. 20638 for
dimensions and construction.

Spring compression. Each spring shall be capable of sustaining a
ion load such that free length of 1-5/16 ± 1/16 inches shall be
:ed to a height of 27/32 ± 1/32 inch, when tested as specified in 4.4.2.

Finish. Casting surfaces shall be substantially smooth. Castings shall
from blowholes, porosity, hard spots, shrinkage defects, cracks, sand,
.ns, sprues, scale and other defects or other extraneous materials. All other
its shall be smoothly finished. All sharp edges shall be smoothly rounded.

Identification marking. Each unit shall be permanently and legibly
ith the name or registered trademark of the manufacturer.

Workmanship. Workmanship shall be first class throughout. The press
free from defects which detract from its appearance or may impair
iceability.

UALITY ASSURANCE PROVISIONS

Responsibility for inspection. Unless otherwise specified in the
or purchase order, the supplier is responsible for the performance
ispection requirements as specified herein. Except as otherwise
d in the contract or order, the supplier may use his own or any
:ilities suitable for the performance of the inspection requirements
d herein, unless disapproved by the Government. The Government
the right to perform any of the inspections set forth in the specifi-
iere such inspections are deemed necessary to assure supplies and
conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the
or shall be maintained by the contractor and made available to the
it, upon the Government's request, at any time, or from time to time,
ie performance of the contract and for a period of 3 years after
of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the components/materials may be furnished in lieu of actual performance of such testing by the contractor, provided that lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component or material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered are classified as specified in tables II and IV (see 4.2.1 and 4.3) and table III (see 4.2.2).

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in table II. Unit of product for examination purposes shall be one press.

Table II. Sampling for Examination

Type Defect	Inspection Level	AQL (Percent Defective)
Major A	II	1.0
Major B	S-2	4.0
Minor A	II	2.5
Minor B	S-2	6.5

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table III for the press. The acceptance number shall be zero for all sample sizes. The unit of product for test purposes shall be one press. Where feasible, the same sample unit may be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

Table III. Sample criteria for Testing

ent	Characteristic	Require-ment	Test procedure	Inspection level
rts in I	Chemical Analysis	3.2	4.4.1	*
prings	Compression	3.4	4.4.2	S-1

ufficient amount from each lot of raw materials used in the construction item shall be selected for chemical analysis.

Examination. The presses shall be examined to determine compliance with quirements of this specification. Examination shall be conducted in ance with the following classification of defects. Examination shall e, but shall not be limited to, the defects listed in table IV.

Table IV. Classification of Defects

ries	Defects
al	None defined
)	Style and design of press not as specified. Component missing. Component damaged, affecting serviceability. Upper pressure plate does not lower and raise easily.
)	Minimum distance between contact surfaces of plates before exerting pressure not as specified. Minimum distance between sides of frame not as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

Table IV. Classification of Defects (con't)

Categories	Defects (con't)
Minor A	
201	Finish not as specified.
202	Identification marking not as specified.
203	Packaging and packing not as specified.
204	Marking on packaging and pack not as specified.
Minor B	
251	Dimensions (other than 151 and 152) not as specified.

4.4 Test procedures.

4.4.1 Chemical analysis. In the absence of certificates of quality, chemical analysis of the individual components of the press shall be determined in accordance with Method 111 or 112 of FED-STD-151. In the event of conflict, Method 111 shall govern.

4.4.2 Spring compression test. Remove the two stainless steel coil springs by disengaging the two screws from the lower pressure plate assembly. Each spring shall be tested individually by mounting in a suitable apparatus such that a compressive force of 200 pounds is applied. The height of the compressed spring shall be measured.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be Levels A or C as specified (see 6.1).

5.1.1 Unit of issue. One press, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table V shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table V. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	10 units

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

.1 Packing variation permitted. If the required number of units ntire shipment is less than the number of units specified to be ed in an exterior container, such units may be packed in an exterior r of suitable size and design, acceptable to a common carrier, which sure safe delivery to destination.

Level A .

.1 Unit package. Unit package shall be a box of appropriate size gn constructed in accordance with PPP-B-636, type CF, class weather- :. Boxes shall be closed by taping the top and bottom flaps with lde tape conforming to PPP-T-76. Closing of bottom flap with adhesive ıg to MMM-A-250 or metal stitches is permitted.

Level C. Units shall be packaged in standard commercial containers ize and kind commonly used, which will afford the degree of protection for shipment and use of the product for its intended purpose.

'acking. Packing shall be Levels A, B or C as specified (see 6.1).

Level A.

1 Exterior container. Exterior shall be designed for a type 2 constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, . Closure and strapping shall be as specified in the appendix of cable box specification. Special requirements of Table II of shall apply for PPP-B-636 box.

Level B.

1 Exterior container. Exterior container shall be designed for load and constructed in accordance with PPP-B-585, class 1, style 3; , domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. hall be as specified in the appendix of the applicable box specifi- Fiberboard box shall be closed by method II of PPP-B-636. Special nts of Table I of PPP-B-636 shall apply for PPP-B-636 box.

Level C. The subject commodity shall be packed in substantial l containers of the type, size, and kind commonly used for the so constructed as to insure acceptance and safe delivery by common carriers, at the lowest rate, to point of delivery called for in ıct or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T11:52Z
Check the source to verify that this is the currentversion before use.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipment shall be in accordance with Fed Std No. 123.

5.3.2 Military activities.

5.3.2.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 Specification changes. Because of the extensiveness of changes to this document, the margins have not been marked with an asterisk to indicate where changes from the previous issue have been made.

6.3 Federal Stock Number (FSN) coverage. This specification covers the following item appearing in the Federal Supply Catalog:

Stock Number	Item Identification
6520-515-5150	PRESS, DENTURE FLASK

6.4 Limitations. This specification does not cover all types, classes, grades, styles, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover those items which are normally procured to meet Federal Government requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

5B (DSA-DM)

s:
MD
MS
ce - 03

ivities:
MD
MS
ce - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6520-1345

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE: GG-P-00636B (DSA-DM)

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

SE PERSONNEL SUPPORT CENTER
00 South 20th Street
iladelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

BUSINESS
PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:52Z
Check the source to verify that this is the currentversion before use.